

REPORT OVERVIEW

Report number: TR-2026-000019

Overall mathematical result: PASS

Workflow status: Draft review

Generated: 21/09/2026, 01:24:18

Laboratory: samarth labs

Tester: samarth dubey

Test dates: 15/09/2026, 23:21:00 to Not recorded

INSTRUMENT IDENTIFICATION

Manufacturer: NAWI Test Instruments Pvt. Ltd.

Model / type: NTW-80C

Serial number: NTW80C-TEST-001

Accuracy class: Class III

Min: 0.2 kg | Max: 80 kg

e: 0.01 kg | d: 0.005 kg

Power source: AC_MAINS | Nominal voltage: 230 V

TEST CONDITIONS

Temperature: 233 to Not recorded C

Relative humidity: 50 to Not recorded %

Barometric pressure: 1008 to Not recorded hPa

TEST SUMMARY

A.4.2 | Checking of zero | PASS

A.4.3 | Setting to zero before loading | PASS

A.4.4 | Determination of weighing performance | PASS

A.4.5 | Multiple indicating devices | PASS

A.4.6 | Tare | PASS

A.4.7 | Eccentricity | PASS

A.4.8 | Discrimination | PASS

A.4.9 | Sensitivity of non-self-indicating instrument | NOT APPLICABLE

A.4.10 | Repeatability | PASS

A.4.11 | Variation of indication with time | PASS

A.4.12 | Stability of equilibrium | PASS

A.4.13 | Portable weighbridge tests | NOT APPLICABLE

A.5 | Influence Factors | PASS

A.6 | Endurance test | PASS

A.4.4 - WEIGHING PERFORMANCE

Result: PASS

Observation 1 | Load 0.2 kg | Indication 0.2 kg | Corrected error 0.002 kg | MPE 0.005 kg | PASS

Observation 2 | Load 5 kg | Indication 5 kg | Corrected error 0.001 kg | MPE 0.005 kg | PASS
Observation 3 | Load 8 kg | Indication 8 kg | Corrected error 0.003 kg | MPE 0.01 kg | PASS
Observation 4 | Load 20 kg | Indication 20 kg | Corrected error 0 kg | MPE 0.01 kg | PASS
Observation 5 | Load 40 kg | Indication 40 kg | Corrected error 0.002 kg | MPE 0.015 kg | PASS
Observation 6 | Load 80 kg | Indication 80 kg | Corrected error 0.001 kg | MPE 0.015 kg | PASS
Observation 7 | Load 72 kg | Indication 72 kg | Corrected error 0.003 kg | MPE 0.015 kg | PASS
Observation 8 | Load 60 kg | Indication 60 kg | Corrected error 0 kg | MPE 0.015 kg | PASS
Observation 9 | Load 48 kg | Indication 48 kg | Corrected error 0.002 kg | MPE 0.015 kg | PASS
Observation 10 | Load 0.21 kg | Indication 0.21 kg | Corrected error 0.001 kg | MPE 0.005 kg | PASS

A.5 - INFLUENCE FACTORS

Result: PASS

A.5.1 | COMPLETED | PASS

A.5.2 | COMPLETED | PASS

A.5.3 | COMPLETED | PASS

A.5.4 | COMPLETED | PASS

A.6 - ENDURANCE

Mathematical result: PASS

Requirement: 100000 loading applications

Recorded workflow count: 100000 | Synthetic prototype count: 99900

Classification: Prototype workflow; not legal-metrology evidence

Pre-endurance baseline: 0.002 kg

Post-endurance corrected error: 0 kg

Durability error: 0.002 kg | Result: PASS

Abnormal events recorded: 70

Recovery checkpoints recorded: 6

TRACEABILITY

OIML R 76-1:2006 and relevant OIML R 76-2 test-report structure. Values are read from the persisted test record.

EVIDENCE APPENDIX

Evidence 01 | A.4.11.1 - Creep - Setup | 20/09/2026, 15:05:52 | samarth dubey

Evidence 02 | A.4.11.1 - Creep - Setup | 20/09/2026, 15:06:32 | samarth dubey

Evidence 03 | A.4.11.1 - Creep - Setup | 20/09/2026, 15:11:16 | samarth dubey

Evidence 04 | Additional report evidence | 21/09/2026, 00:48:32 | samarth dubey