

CONTROLLED TECHNICAL DOCUMENT

NAWI TEST REPORT

Type evaluation and verification record

DRAFT REPORT

TR-2026-000038

FAIL

DEMONSTRATION / SYNTHETIC DATA

Synthetic observations are identified as such; reported results are the persisted application calculations.

REPORT AND INSTRUMENT IDENTIFICATION

Report number TR-2026-000038	Document status DRAFT REPORT	Workflow status Rejected
Manufacturer Regression Metrology Lab	Applicant Regression Metrology Lab	Model / type designation Prototype Class III Full Suite
Serial number QA-FOUR-III-001	Accuracy class / indication Class III / Self-indicating	Capacity 20 - 30,000 g
e / d 10 / 5 g	Tester Regression QA Tester Four	Reviewer samarth dubey
Applicable standard OIML R 76-1:2006	Revision 01	Issue date 24 Sept 2026, 20:34

APPLICANT AND MANUFACTURER

Field	Recorded information
Applicant contact	Regression QA Tester Four
Email	qa.testers.four@local.test
Telephone	+91 9000000004
Applicant address	Synthetic QA workspace, local prototype
Manufacturer address	Synthetic QA workspace, local prototype

INSTRUMENT TECHNICAL SPECIFICATIONS

Field	Recorded information
Range / interval configuration	Single-Range / Single-Interval
Power source	AC Mains
Nominal voltage	230 V
Operating voltage range	207 - 253 V
Tare configuration	Subtractive - Semi-automatic
Maximum tare effect	10,000 g
Stable equilibrium function	Yes
Multiple indicating devices	Yes
Printing capability	Yes
Data storage capability	Yes

TEST CONDITIONS

Parameter	Recorded start	Recorded end	Unit
Temperature	20	20	deg C
Relative humidity	45	50	%
Atmospheric pressure	1,013	1,013	hPa

Field	Recorded information
Evaluation started	23 Sept 2026, 20:25
Evaluation completed	23 Sept 2026, 23:00

Field	Recorded information
Test setup notes	Synthetic fixture for prototype regression; no physical instrument claim.
Preparation notes	Synthetic fixture for prototype regression; no physical instrument claim.
Environmental notes	SYNTHETIC PROTOTYPE / REGRESSION DATA. Session end conditions are simulated solely to verify the report handoff workflow; no physical laboratory test session occurred.

TEST EQUIPMENT AND CALIBRATION REFERENCES

No test equipment or calibration reference information is recorded.

APPLICABLE TEST SUMMARY

Test ID	Test description	Status	Result	Non-applicability reason
A.4.2	Checking of zero	Completed	PASS	
A.4.3	Setting to zero before loading	Completed	PASS	
A.4.4	Determination of weighing performance	Completed	PASS	
A.4.5	Multiple indicating devices	Completed	PASS	
A.4.6	Tare	Completed	PASS	
A.4.7	Eccentricity	Completed	PASS	
A.4.8	Discrimination	Completed	PASS	
A.4.9	Sensitivity of non-self-indicating instrument	NOT APPLICABLE	NOT APPLICABLE	Not applicable: A.4.9 is the sensitivity test for a non-self-indicating instrument.
A.4.10	Repeatability	Completed	PASS	
A.4.11	Variation of indication with time	Completed	PASS	
A.4.12	Stability of equilibrium	Completed	FAIL	
A.4.13	Portable weighbridge tests	NOT APPLICABLE	NOT APPLICABLE	The instrument is not a portable road-vehicle weighbridge.
A.5	Influence Factors	Completed	PASS	
A.6	Endurance test	Completed	PASS	

DETAILED TEST RESULTS

A.4.2 - CHECKING OF ZERO

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.2	Started 23 Sept 2026, 20:25
Completed 23 Sept 2026, 20:26		

A.4.3 - SETTING TO ZERO BEFORE LOADING

TEST RESULT: PASS

Method Automatic zero-setting / zero-tracking	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.3(b)
Started 23 Sept 2026, 20:26	Completed 23 Sept 2026, 20:26	

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Method Used	A.4.3(b)	Source Phase	A.4.2.3
Zero Reference Established	Yes	Zero Deviation	0
Zero Deviation Unit	g		

A.4.4 - DETERMINATION OF WEIGHING PERFORMANCE

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.4	Started 23 Sept 2026, 20:26
Completed 23 Sept 2026, 20:27		

Weighing observations

No.	Applied load / indication / Delta L	P / Error Ec	MPE	Result
1	20 / 20 / 5 g	20 / 0 g	+/-5 g	PASS
2	3,000 / 3,000 / 5 g	3,000 / 0 g	+/-5 g	PASS
3	5,000 / 5,000 / 5 g	5,000 / 0 g	+/-5 g	PASS
4	7,500 / 7,500 / 5 g	7,500 / 0 g	+/-10 g	PASS

No.	Applied load / indication / Delta L	P / Error Ec	MPE	Result
5	20,000 / 20,000 / 5 g	20,000 / 0 g	+/-10 g	PASS
6	30,000 / 30,000 / 5 g	30,000 / 0 g	+/-15 g	PASS
7	27,000 / 27,000 / 5 g	27,000 / 0 g	+/-15 g	PASS
8	22,500 / 22,500 / 5 g	22,500 / 0 g	+/-15 g	PASS
9	15,000 / 15,000 / 5 g	15,000 / 0 g	+/-10 g	PASS
10	30 / 30 / 5 g	30 / 0 g	+/-5 g	PASS

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Supported	Yes	Observation Unit	g
Zero reference - Unit	g	Zero reference - Zero Indication	0 g
Zero reference - Additional load (Delta L0)	5 g	Zero reference - Calculated E0	0
Zero reference - Recorded	23 Sept 2026, 20:26		

A.4.5 - MULTIPLE INDICATING DEVICES

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.6.3; comparison during Annex A A.4.4	Completed 23 Sept 2026, 20:27
------------------------------------	---	---

Indicating-device comparisons

No.	Load	Direction	Device 1	Device 2	Difference	Limit	Result
1	20 g	INCREASING	20 g	20 g	0 g	5 g	PASS
2	3,000 g	INCREASING	3,000 g	3,000 g	0 g	5 g	PASS
3	5,000 g	INCREASING	5,000 g	5,000 g	0 g	5 g	PASS
4	7,500 g	INCREASING	7,500 g	7,500 g	0 g	10 g	PASS
5	20,000 g	INCREASING	20,000 g	20,000 g	0 g	10 g	PASS
6	30,000 g	INCREASING	30,000 g	30,000 g	0 g	15 g	PASS
7	27,000 g	DECREASING	27,000 g	27,000 g	0 g	15 g	PASS
8	22,500 g	DECREASING	22,500 g	22,500 g	0 g	15 g	PASS
9	15,000 g	DECREASING	15,000 g	15,000 g	0 g	10 g	PASS

No.	Load	Direction	Device 1	Device 2	Difference	Limit	Result
10	30 g	DECREASING	30 g	30 g	0 g	5 g	PASS

A.4.6 - TARE

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.6	Started 23 Sept 2026, 20:27
Completed 23 Sept 2026, 20:29		

A.4.6.2 - Accuracy of tare setting

Rep.	Tare load	L0	I0	Delta L	E0	Limit	Result
1	9,000 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
2	9,000 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
3	9,000 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
4	9,000 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
5	9,000 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS

Authoritative persisted result: PASS. 5 of 5 required repetitions recorded. Last E0: 0 g. Applicable limit: +/-2.5 g.

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Tare Configuration Snapshot - Tare Device Present	Yes	Tare Configuration Snapshot - Tare Type	SUBTRACTIVE
Tare Configuration Snapshot - Maximum Tare Effect - Value	10,000 g	Tare Configuration Snapshot - Maximum Tare Effect - Unit	g
Tare Configuration Snapshot - Tare Operation Mode	Semi-automatic	Tare Configuration Snapshot - Tare Weighing Device Present	No
Tare Configuration Snapshot - Preset Tare Device Present	No		

A.4.7 - ECCENTRICITY

TEST RESULT: PASS

Method Load receptor with not more than four points of support	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.7; calculation A.4.4.3
Started 23 Sept 2026, 20:29	Completed 23 Sept 2026, 20:30	

Position observations

Series / phase	Position Id	Label	Loading Area	Geometry
Quarter 1 - POSITION_1	Position 1	Quarter 1	Approximately one quarter of the load-receptor surface.	quarter
Quarter 2 - POSITION_2	Position 2	Quarter 2	Approximately one quarter of the load-receptor surface.	quarter
Quarter 3 - POSITION_3	Position 3	Quarter 3	Approximately one quarter of the load-receptor surface.	quarter
Quarter 4 - POSITION_4	Position 4	Quarter 4	Approximately one quarter of the load-receptor surface.	quarter

Observations

Series / phase	Load L	Indication I	Additional load (Delta L)	Corrected error	Maximum permissible error	Compliance Result
Quarter 1 - POSITION_1	10,000 g	10,000 g	5 g	0 g	10 g	PASS
Quarter 2 - POSITION_2	10,000 g	10,000 g	5 g	0 g	10 g	PASS
Quarter 3 - POSITION_3	10,000 g	10,000 g	5 g	0 g	10 g	PASS
Quarter 4 - POSITION_4	10,000 g	10,000 g	5 g	0 g	10 g	PASS

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Execution Supported	Yes	Support Point Count	4
Position Count	4	Zero Condition Confirmed	Yes
Test Load - Value	10,000 g	Test Load - Unit	g
Sketch - Type	four-quarter-normal-platform		

A.4.8 - DISCRIMINATION

TEST RESULT: PASS

Method A.4.8.2	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.8.2
Started 23 Sept 2026, 20:30	Completed 23 Sept 2026, 20:33	

Discrimination observations

Load stage	Target	Starting indication	Lower indication / Delta	Upper indication / Delta	Applied / restored load	d checks	Result
Min	20 g	20 g	15 g / 5 g	25 g / 5 g	7 g / 0.5 g	0.5 g / 7 g	PASS
1/2 Max	15,000 g	15,000 g	14,995 g / 5 g	15,005 g / 5 g	7 g / 0.5 g	0.5 g / 7 g	PASS
Max	30,000 g	30,000 g	29,995 g / 5 g	30,005 g / 5 g	7 g / 0.5 g	0.5 g / 7 g	PASS

A.4.9 - SENSITIVITY OF NON-SELF-INDICATING INSTRUMENT

TEST RESULT: NOT APPLICABLE

Field	Recorded information
Status	NOT APPLICABLE
Reason	Not applicable: A.4.9 is the sensitivity test for a non-self-indicating instrument.

A.4.10 - REPEATABILITY

TEST RESULT: PASS

Method Same Load Repeat Weighings	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.6.1 / Annex A A.4.10
Started 23 Sept 2026, 20:34	Completed 23 Sept 2026, 20:34	

Observations

Series / phase	Repetition	Actual Load	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error
Trial 1 / Approximately 80% of Max	1	24,000 g	24,000 g	0 g	5 g	15 g
Trial 2 / Approximately 80% of Max	2	24,000 g	24,000 g	0 g	5 g	15 g
Trial 3 / Approximately 80% of Max	3	24,000 g	24,000 g	0 g	5 g	15 g

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Control Stage	VERIFICATION	Zero Reference E0	0
Observation Unit	g	Procedure Confirmation - Automatic Zero On Confirmed	Yes
Procedure Confirmation - Unloaded Instrument Rest Confirmed	Yes	Procedure Confirmation - Confirmed At	23 Sept 2026, 20:34

A.4.11 - VARIATION OF INDICATION WITH TIME

TEST RESULT: PASS

Method A.4.11.1 + A.4.11.2	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.9.4 / Annex A A.4.11
Started 23 Sept 2026, 20:35	Completed 23 Sept 2026, 21:47	

Variation and zero-return observations

Checkpoint	Elapsed	Load	Indication	Interpolation Delta L / zero deviation	P	Temperature	Result
Creep: T0	0 min	30,000 g	30,000 g	0 g	30,005 g	20 deg C	PASS
Creep: T5	5 min	30,000 g	30,000 g	0 g	30,005 g	20 deg C	PASS
Creep: T15	15 min	30,000 g	30,000 g	0 g	30,005 g	20 deg C	PASS
Creep: T30	30 min	30,000 g	30,000 g	0 g	30,005 g	20 deg C	PASS
Zero return: Post-load zero-return	After 30 min	-	Before 0; after 0 g	0 g	-	-	PASS

Variation calculation summary

Creep result PASS	Creep Delta P (T0 to T30) 0 g	Creep limit 15 g	Creep temperature variation 0 deg C
Zero return result PASS	Zero return zero-return deviation 0 g	Zero return limit 5 g	Zero return zero tracking off Yes
Zero return readings (before / after) 0 / 0 g	Zero return stabilization before: confirmed; after: confirmed		

A.4.12 - STABILITY OF EQUILIBRIUM

TEST RESULT: FAIL

Method Stability Of Equilibrium	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 4.4.2 / Annex A A.4.12
Started 23 Sept 2026, 21:47	Completed 23 Sept 2026, 21:55	

Stability phase summary

Phase	Description	Applicability	Result	Key observation / reason
-------	-------------	---------------	--------	--------------------------

Phase	Description	Applicability	Result	Key observation / reason
A.4.12.1	Stability documentation/ configuration review	APPLICABLE	FAIL	No manufacturer-authored stable-equilibrium documentation was supplied with this synthetic prototype. No claim is made about undocumented internal criteria, adjustable or non-adjustable parameters, parameter security, or worst-case adjustment.
A.4.12.2	Printing / data-storage stability	APPLICABLE	PASS	Test the five-second indication sequence after a disturbance and the print/data-storage command.
A.4.12.3	Zero-setting stability	APPLICABLE	PASS	Check inhibition before stable equilibrium and reuse A.4.2.3 evidence once stable.
A.4.12.4	Tare-balancing stability	APPLICABLE	PASS	Check inhibition before stable equilibrium and reuse A.4.6.2 evidence once stable.
A.4.12.5	Continuous-disturbance inhibition	APPLICABLE	PASS	Record whether every applicable function is inhibited while continuous disturbance remains active.
A.4.12.6	Mobile/vehicle branch	NOT_APPLICABLE	NOT_APPLICABLE	Not applicable - the instrument is configured as stationary.

Observations

Series / phase	Repetition	Additional load (Delta L)	Unit	Disturbance Confirmed	Operation Attempted Immediately	Blocked Before Stable
Trial 1 / A.4.12.3 - Zero-setting stability	1	5 g	g	Yes	Yes	Yes
Trial 2 / A.4.12.3 - Zero-setting stability	2	5 g	g	Yes	Yes	Yes
Trial 3 / A.4.12.3 - Zero-setting stability	3	5 g	g	Yes	Yes	Yes
Trial 4 / A.4.12.3 - Zero-setting stability	4	5 g	g	Yes	Yes	Yes
Trial 5 / A.4.12.3 - Zero-setting stability	5	5 g	g	Yes	Yes	Yes
Trial 1 / A.4.12.4 - Tare-balancing stability	1	5 g	g	Yes	Yes	Yes
Trial 2 / A.4.12.4 - Tare-balancing stability	2	5 g	g	Yes	Yes	Yes
Trial 3 / A.4.12.4 - Tare-balancing stability	3	5 g	g	Yes	Yes	Yes
Trial 4 / A.4.12.4 - Tare-balancing stability	4	5 g	g	Yes	Yes	Yes
Trial 5 / A.4.12.4 - Tare-balancing stability	5	5 g	g	Yes	Yes	Yes

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Setup - Target Load - Value	15,000 g	Setup - Target Load - Unit	g
Setup - Actual Load - Value	15,000 g	Setup - Actual Load - Unit	g

Parameter	Recorded value	Parameter	Recorded value
Setup - Input Unit	g	Setup - Recorded	23 Sept 2026, 21:48
Print Storage - Function Tested	PRINT	Print Storage - Disturbance Confirmed	Yes
Print Storage - Initiated Immediately Confirmed	Yes	Print Storage - Required Repetitions	5
Print Storage - Stability Interval - Value	10 g	Print Storage - Stability Interval - Unit	g
Print Storage - Stability Interval - Kind	1e	Zero Setting - Required Repetitions	5
Tare - Required Repetitions	5	Continuous Disturbance - Continuous Disturbance Confirmed	Yes
Continuous Disturbance - Recorded	23 Sept 2026, 21:55		

A.4.13 - PORTABLE WEIGHBRIDGE TESTS

TEST RESULT: NOT APPLICABLE

Field	Recorded information
Status	NOT APPLICABLE
Reason	The instrument is not a portable road-vehicle weighbridge.

A.5 - INFLUENCE FACTORS

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.5	Started 23 Sept 2026, 21:56
Completed 23 Sept 2026, 22:01		

Influence factor results

Factor	Result	Key value / observation
A.5.1	PASS	Use the manufacturer limiting tilt with the configured level indicator/automatic tilt sensor.
A.5.2	PASS	The instrument uses electric power and is in scope for the warm-up test.
A.5.3	PASS	The specified operating-temperature range is available for the A.5.3 procedure.
A.5.4	PASS	The configured power-source branch determines the A.5.4 voltage limits.

Observations

Series / phase	Label	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
A.5	Not recorded	0	0	5	5	Not recorded
A.5	Not recorded	100	0	5	5	Not recorded
A.5	Not recorded	30,000	0	5	15	Not recorded
A.5	Not recorded	0	0	5	5	Not recorded
A.5	Not recorded	100	0	5	5	Not recorded
A.5	Not recorded	30,000	0	5	15	Not recorded
A.5	Not recorded	15,000 g	0 g	5 g	10 g	g
A.5	Not recorded	15,000 g	0 g	5 g	10 g	g
A.5	Not recorded	15,000 g	0 g	5 g	10 g	g
REFERENCE_START	Reference start	100 g	0 g	5 g	5 g	g
LOWER	LOWER	100 g	0 g	5 g	5 g	g
UPPER	UPPER	15,000 g	0 g	5 g	10 g	g
REFERENCE_END	Reference end	15,000 g	0 g	5 g	10 g	g

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Setup - Environmental Conditions Stable Confirmed	Yes	Setup - Recorded	23 Sept 2026, 21:57
Tilting - Zero Setting Off Confirmed	Yes	Tilting - Recorded	23 Sept 2026, 21:59
Warm Up - Disconnected At	23 Sept 2026, 00:00	Warm Up - Connected At	23 Sept 2026, 08:00
Warm Up - Stabilization At	23 Sept 2026, 08:05	Warm Up - Eight Hour Precondition Confirmed	Yes
Temperature - Recorded	23 Sept 2026, 22:00	Voltage - Recorded	23 Sept 2026, 22:01

A.6 - ENDURANCE TEST

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6	Started 23 Sept 2026, 22:01
Completed 23 Sept 2026, 22:03		

Recovery checkpoints

Applications	Checkpoint time	Actual load	Recorded by	Notes
0	23 Sept 2026, 22:01	Not recorded	Regression QA Tester Four	Session created; software checkpoint only.
10,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
25,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
50,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
75,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
100,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.

Endurance calculation

Field	Recorded information
Required applications	100,000
Recorded applications	100,000
Pre-endurance baseline error	0 g
Post-endurance corrected error	5 g
Durability error	5 g
Abnormal events	0
Cycle state	COMPLETED

Post-endurance indication
15,000 g

REMARKS, DEVIATIONS AND NON-CONFORMITIES

Synthetic fixture for prototype regression; no physical instrument claim.

SYNTHETIC PROTOTYPE / REGRESSION DATA. Session end conditions are simulated solely to verify the report handoff workflow; no physical laboratory test session occurred.

Rejected: A.4.12.1 stability documentation/configuration review is marked FAIL because no manufacturer-authored stable-equilibrium documentation or reference was supplied. The remaining applicable A.4.12 phases and A.5/A.6 calculations are PASS, but the report-level result is FAIL. Please provide genuine supporting manufacturer documentation through the appropriate future workflow; no measurement data were changed.

OVERALL CONCLUSION

Overall evaluation result: FAIL

The persisted evaluation record contains one or more applicable test results that do not satisfy their recorded acceptance criteria.

REVIEW AND APPROVAL

Role	Name	Action / decision	Date
Tester	Regression QA Tester Four	Submitted for review	23 Sept 2026, 22:07
Reviewer	samarth dubey	Rejected	23 Sept 2026, 22:09

Reviewer note

Rejected: A.4.12.1 stability documentation/configuration review is marked FAIL because no manufacturer-authored stable-equilibrium documentation or reference was supplied. The remaining applicable A.4.12 phases and A.5/A.6 calculations are PASS, but the report-level result is FAIL. Please provide genuine supporting manufacturer documentation through the appropriate future workflow; no measurement data were changed.

Review history

Action	Recorded by	Date
Submitted for reviewer assessment	Regression QA Tester Four	23 Sept 2026, 22:07
Rejected by reviewer	samarth dubey	23 Sept 2026, 22:09

TRACEABILITY & REVISION

Technical basis OIML R 76-1:2006	Report format OIML R 76-2 structure	Revision 01
Document status DRAFT	Record basis Persisted tester observations and stored calculation results	Generated 24 Sept 2026, 20:34

EVIDENCE APPENDIX

0 persisted supporting records

No supporting evidence is attached to this report.