

CONTROLLED TECHNICAL DOCUMENT

NAWI TEST REPORT

Type evaluation and verification record

DRAFT REPORT

TR-2026-000038

FAIL

DEMONSTRATION / SYNTHETIC DATA

Synthetic observations are identified as such; reported results are the persisted application calculations.

INSTRUMENT IDENTIFICATION

Field	Recorded information
Manufacturer	Regression Metrology Lab
Applicant	Regression Metrology Lab
Model / type designation	Prototype Class III Full Suite
Serial number	QA-FOUR-III-001
Accuracy class	Class III
Indication type	Self-indicating
Capacity	20 - 30,000 g
Verification interval e	10 g
Scale interval d	5 g

DOCUMENT CONTROL

Field	Recorded information
Report number	TR-2026-000038
Revision	01
Document status	DRAFT REPORT
Issue date	24 Sept 2026, 19:34
Workflow status	Rejected

Field	Recorded information
Applicable standard	OIML R 76-1:2006
Laboratory	Regression Metrology Lab
Tester	Regression QA Tester Four
Reviewer	samarth dubey

APPLICANT AND MANUFACTURER

Field	Recorded information
Applicant	Regression Metrology Lab
Contact person	Regression QA Tester Four
Email	qa.testers.four@local.test
Telephone	+91 9000000004
Address	Synthetic QA workspace, local prototype
Manufacturer	Regression Metrology Lab
Manufacturer address	Synthetic QA workspace, local prototype

INSTRUMENT TECHNICAL SPECIFICATIONS

Field	Recorded information
Instrument type	Prototype Class III Full Suite
Manufacturer	Regression Metrology Lab
Serial number	QA-FOUR-III-001
Accuracy class	Class III
Indication type	Self-indicating
Mass unit	g
Minimum capacity	20 g
Maximum capacity	30,000 g
Verification interval (e)	10 g
Scale interval (d)	5 g
Verification intervals (n)	3,000
Range / interval configuration	Single-Range / Single-Interval
Power source	AC Mains
Nominal voltage	230 V
Operating voltage range	207 - 253 V
Tare configuration	Subtractive - Semi-automatic
Maximum tare effect	10,000 g

Field	Recorded information
Stable equilibrium function	Yes
Multiple indicating devices	Yes
Printing capability	Yes
Data storage capability	Yes

TEST CONDITIONS

Parameter	Recorded start	Recorded end	Unit
Temperature	20	20	deg C
Relative humidity	45	50	%
Atmospheric pressure	1,013	1,013	hPa

Field	Recorded information
Evaluation started	23 Sept 2026, 20:25
Evaluation completed	23 Sept 2026, 23:00
Test setup notes	Synthetic fixture for prototype regression; no physical instrument claim.
Preparation notes	Synthetic fixture for prototype regression; no physical instrument claim.
Environmental notes	SYNTHETIC PROTOTYPE / REGRESSION DATA. Session end conditions are simulated solely to verify the report handoff workflow; no physical laboratory test session occurred.

TEST EQUIPMENT AND CALIBRATION REFERENCES

No test equipment or calibration reference information is recorded.

APPLICABLE TEST SUMMARY

Test ID	Test description	Status	Result	Non-applicability reason
A.4.2	Checking of zero	Completed	PASS	
A.4.3	Setting to zero before loading	Completed	PASS	
A.4.4	Determination of weighing performance	Completed	PASS	
A.4.5	Multiple indicating devices	Completed	PASS	
A.4.6	Tare	Completed	PASS	
A.4.7	Eccentricity	Completed	PASS	
A.4.8	Discrimination	Completed	PASS	

Test ID	Test description	Status	Result	Non-applicability reason
A.4.9	Sensitivity of non-self-indicating instrument	NOT APPLICABLE	NOT APPLICABLE	Not applicable: A.4.9 is the sensitivity test for a non-self-indicating instrument.
A.4.10	Repeatability	Completed	PASS	
A.4.11	Variation of indication with time	Completed	PASS	
A.4.12	Stability of equilibrium	Completed	FAIL	
A.4.13	Portable weighbridge tests	NOT APPLICABLE	NOT APPLICABLE	The instrument is not a portable road-vehicle weighbridge.
A.5	Influence Factors	Completed	PASS	
A.6	Endurance test	Completed	PASS	

DETAILED TEST RESULTS

A.4.2 - CHECKING OF ZERO

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.2-1.0
Technical reference	OIML R 76-1:2006 Annex A A.4.2
Started	23 Sept 2026, 20:25
Completed	23 Sept 2026, 20:26
Support / applicability reason	The instrument has a configured zero-setting function, so the A.4.2 checking-of-zero module applies.

Field	Recorded information
Test procedure version	R76-A4.2-1.0
Workflow status	COMPLETED
Result	PASS

Phases

Code	Name	Applicability	Workflow status	Result	Reason
A.4.2.1	Zero-setting Range	APPLICABLE	COMPLETED	OBSERVED	Use the semi-automatic zero-setting-range procedure.
A.4.2.2	Zero Indicator	APPLICABLE	COMPLETED	OBSERVED	The instrument has a digital zero-indicating device.

Code	Name	Applicability	Workflow status	Result	Reason
A.4.2.3	Accuracy of Zero-setting	APPLICABLE	COMPLETED	PASS	Use the non-automatic or semi-automatic accuracy procedure (A.4.2.3.1).

Observations

Field	Recorded information
Starting Zero State	0
Applied Load	500
Zero Restored	Yes
Observed Positive Range	500
Observed Negative Range	500
Notes	Synthetic prototype observation within the configured zero-setting range.
Unit	g
Input Applied Load	500
Input Observed Positive Range	500
Input Observed Negative Range	500

Observations

Field	Recorded information
Unit	g
Input Unit	g
Starting Indication	0
Increment	0.5
Observed Lower Range	-0.5
Observed Upper Range	0.5
Input Starting Indication	0
Input Observed Lower Range	-0.5
Input Observed Upper Range	0.5

Indication Observations

Sequence	Unit	Value	Input Value	Input Unit	Observed At
1	g	0 g	0 g	g	23 Sept 2026, 20:25

Sequence	Unit	Value	Input Value	Input Unit	Observed At
2	g	0.5 g	0.5 g	g	23 Sept 2026, 20:25
3	g	1 g	1 g	g	23 Sept 2026, 20:25

Calculations

Field	Recorded information
Calculated indication	15,000
Error	0
Corrected error	0
Calculated E0	0
Accuracy Class	III
Verification scale interval (e)	10
M	1,500
Mpe Multiplier	1
Maximum permissible error	10
MPE unit	g
Technical reference	OIML R 76-1:2006 Section 3.5.1 Table 6
Reference edition	R76-1:2006
Range Label	>500e to <=2000e
Compliance Result	PASS
Result	PASS

Observations

Field	Recorded information
Zero Indication I0	0
Additional load (Delta L0)	5
Load L	15,000
Indication I	15,000
Additional load (Delta L)	5
Notes	Synthetic prototype changeover observation; load and indication balanced within MPE.
Unit	g
Input Zero Indication I0	0

Field	Recorded information
Input Delta L0	5
Input Load L	15,000
Input Indication I	15,000
Input Delta L	5

A.4.3 - SETTING TO ZERO BEFORE LOADING

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.3-1.0
Method	Automatic zero-setting / zero-tracking
Technical reference	OIML R 76-1:2006 Annex A A.4.3(b)
Started	23 Sept 2026, 20:26
Completed	23 Sept 2026, 20:26
Test notes	Synthetic prototype: zero reference established at 0 before loading, consistent with A.4.2.3 result.
Support / applicability reason	For semi-automatic or automatic zero-setting, or zero-tracking, the deviation from zero is determined according to A.4.2.3.

Field	Recorded information
Test procedure version	R76-A4.3-1.0
Workflow status	COMPLETED
Result	PASS
Method Used	A.4.3(b)
Method Label	Automatic zero-setting / zero-tracking
Source Phase	A.4.2.3
Zero Reference Established	Yes
Zero Deviation	0
Zero Deviation Unit	g
Operator Notes	Synthetic prototype: zero reference established at 0 before loading, consistent with A.4.2.3 result.

Source Evidence

Field	Recorded information
Result	PASS
Completed	23 Sept 2026, 20:26

Observations

Field	Recorded information
Zero Indication IO	0
Additional load (Delta L0)	5
Load L	15,000
Indication I	15,000
Additional load (Delta L)	5
Notes	Synthetic prototype changeover observation; load and indication balanced within MPE.
Unit	g
Input Zero Indication IO	0
Input Delta L0	5
Input Load L	15,000
Input Indication I	15,000
Input Delta L	5

Calculations

Field	Recorded information
Calculated indication	15,000
Error	0
Corrected error	0
Calculated E0	0
Accuracy Class	III
Verification scale interval (e)	10
M	1,500
Mpe Multiplier	1
Maximum permissible error	10

Field	Recorded information
MPE unit	g
Technical reference	OIML R 76-1:2006 Section 3.5.1 Table 6
Reference edition	R76-1:2006
Range Label	>500e to <=2000e
Compliance Result	PASS
Result	PASS

A.4.4 - DETERMINATION OF WEIGHING PERFORMANCE

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Technical reference	OIML R 76-1:2006 Annex A A.4.4
Started	23 Sept 2026, 20:26
Completed	23 Sept 2026, 20:27
Support / applicability reason	The weighing-performance procedure is available for a configured single-range, single-interval NAWI.

Weighing observations

No.	Applied load / indication / Delta L	P / Error Ec	MPE	Result
1	20 / 20 / 5 g	20 / 0 g	+/-5 g	PASS
2	3,000 / 3,000 / 5 g	3,000 / 0 g	+/-5 g	PASS
3	5,000 / 5,000 / 5 g	5,000 / 0 g	+/-5 g	PASS
4	7,500 / 7,500 / 5 g	7,500 / 0 g	+/-10 g	PASS
5	20,000 / 20,000 / 5 g	20,000 / 0 g	+/-10 g	PASS
6	30,000 / 30,000 / 5 g	30,000 / 0 g	+/-15 g	PASS
7	27,000 / 27,000 / 5 g	27,000 / 0 g	+/-15 g	PASS
8	22,500 / 22,500 / 5 g	22,500 / 0 g	+/-15 g	PASS
9	15,000 / 15,000 / 5 g	15,000 / 0 g	+/-10 g	PASS
10	30 / 30 / 5 g	30 / 0 g	+/-5 g	PASS

Field	Recorded information
Workflow status	COMPLETED
Result	PASS
Reference edition	R76-1:2006
Supported	Yes
Observation Unit	g

Zero reference

Field	Recorded information
Unit	g
Zero Indication	0
Additional load (Delta L0)	5
Input Zero Indication	0
Input Delta L0	5
Calculated E0	0
Recorded	23 Sept 2026, 20:26

A.4.5 - MULTIPLE INDICATING DEVICES

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.5-1.0
Technical reference	OIML R 76-1:2006 Section 3.6.3; comparison during Annex A A.4.4
Completed	23 Sept 2026, 20:27
Support / applicability reason	The instrument has multiple indicating devices.

Field	Recorded information
Test procedure version	R76-A4.5-1.0
Method / reference	OIML R 76-1:2006 Section 3.6.3; comparison during Annex A A.4.4
Workflow status	PASS
Result	PASS

Comparisons

Sequence	Result	Load	Unit	Direction	Difference
1	PASS	20 g	g	INCREASING	0
2	PASS	3,000 g	g	INCREASING	0
3	PASS	5,000 g	g	INCREASING	0
4	PASS	7,500 g	g	INCREASING	0
5	PASS	20,000 g	g	INCREASING	0
6	PASS	30,000 g	g	INCREASING	0
7	PASS	27,000 g	g	DECREASING	0
8	PASS	22,500 g	g	DECREASING	0
9	PASS	15,000 g	g	DECREASING	0
10	PASS	30 g	g	DECREASING	0

Devices

Label	Indication	Unit	Device Id	Type
Device 1	20 g	g	Device 1	DISPLAY
Device 2	20 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	3,000 g	g	Device 1	DISPLAY
Device 2	3,000 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	5,000 g	g	Device 1	DISPLAY
Device 2	5,000 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	7,500 g	g	Device 1	DISPLAY
Device 2	7,500 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	20,000 g	g	Device 1	DISPLAY
Device 2	20,000 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	30,000 g	g	Device 1	DISPLAY
Device 2	30,000 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	27,000 g	g	Device 1	DISPLAY
Device 2	27,000 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	22,500 g	g	Device 1	DISPLAY
Device 2	22,500 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	15,000 g	g	Device 1	DISPLAY
Device 2	15,000 g	g	Device 2	DISPLAY

Devices

Label	Indication	Unit	Device Id	Type
Device 1	30 g	g	Device 1	DISPLAY
Device 2	30 g	g	Device 2	DISPLAY

A.4.6 - TARE

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.6-1.0
Technical reference	OIML R 76-1:2006 Annex A A.4.6
Started	23 Sept 2026, 20:27
Completed	23 Sept 2026, 20:29
Support / applicability reason	A tare device and the characteristics needed for the tare procedures are configured.

Field	Recorded information
Test procedure version	R76-A4.6-1.0
Method / reference	OIML R 76-1:2006 Annex A A.4.6
Workflow status	COMPLETED
Result	PASS

Tare Configuration Snapshot

Field	Recorded information
Tare Device Present	Yes
Tare Type	SUBTRACTIVE
Tare Operation Mode	Semi-automatic
Tare Weighing Device Present	No
Preset Tare Device Present	No

Maximum Tare Effect

Field	Recorded information
Value	10,000
Unit	g

Phases

Code	Name	Applicability	Workflow status	Result	Reason
A.4.6.1	Tare weighing test	APPLICABLE	LOCKED	Not recorded	Perform loading and unloading with the configured tare arrangement.

Code	Name	Applicability	Workflow status	Result	Reason
A.4.6.2	Accuracy of tare setting	APPLICABLE	COMPLETED	PASS	Establish tare-setting accuracy with the indication set to zero by the tare device.
A.4.6.3	Tare weighing device	Not applicable	Not applicable	Not recorded	Not applicable - the instrument has no separate tare-weighing device.

Observations

Sequence	Result	Indication I	Additional load (Delta L)	Corrected error	Maximum permissible error
1	PASS	20 g	5 g	0 g	5 g
2	PASS	5,000 g	5 g	0 g	5 g
3	PASS	10,000 g	5 g	0 g	10 g
4	PASS	20,000 g	5 g	0 g	10 g
5	PASS	25,000 g	5 g	0 g	15 g

Observations

Sequence	Result	Additional load (Delta L)	Unit	Notes	Input Delta L
1	PASS	5 g	g	Synthetic tare-setting repetition, zero reference balanced.	5 g
2	PASS	5 g	g	Synthetic tare-setting repetition 2, zero reference balanced.	5 g
3	PASS	5 g	g	Synthetic tare-setting repetition 3, zero reference balanced.	5 g
4	PASS	5 g	g	Synthetic tare-setting repetition 4, zero reference balanced.	5 g
5	PASS	5 g	g	Synthetic tare-setting repetition 5, zero reference balanced.	5 g

Calculations

Field	Recorded information
Last Error E0	0
Accuracy Limit	2.5
Result	PASS
Source Phase	A.4.2.3

Procedure

Field	Recorded information
Tare Load Fraction	0.3
Recommended Tare Load	9,000
Load L0	100
Load L0 Is Applied	Yes
Accuracy Limit	2.5
Repetitions	5

Completion

Field	Recorded information
Required Repetitions	5
Valid Repetitions	5
Complete	Yes
Result	PASS
Reason	All 5 tare-setting repetitions are recorded.

A.4.7 - ECCENTRICITY

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.7-1.0
Method	Load receptor with not more than four points of support
Technical reference	OIML R 76-1:2006 Annex A A.4.7; calculation A.4.4.3
Started	23 Sept 2026, 20:29
Completed	23 Sept 2026, 20:30
Support / applicability reason	The instrument configuration determines the applicable eccentricity method and loading positions.

Field	Recorded information
Test procedure version	R76-A4.7-1.0
Method / reference	OIML R 76-1:2006 Annex A A.4.7; calculation A.4.4.3
Method	A.4.7.1

Field	Recorded information
Method Label	Load receptor with not more than four points of support
Execution Supported	Yes
Support Point Count	4
Position Count	4
Workflow status	COMPLETED
Result	PASS
Zero Condition Confirmed	Yes

Test Load

Field	Recorded information
Value	10,000
Unit	g

Sketch

Field	Recorded information
Type	four-quarter-normal-platform
Method / reference	OIML R 76-1:2006 Annex A A.4.7.1

Position Ids

Position 1 - Position 2 - Position 3 - Position 4

Positions

Label	Workflow status	Result	Completed	Position Id	Loading Area
Quarter 1	COMPLETED	PASS	23 Sept 2026, 20:30	Position 1	Approximately one quarter of the load-receptor surface.
Quarter 2	COMPLETED	PASS	23 Sept 2026, 20:30	Position 2	Approximately one quarter of the load-receptor surface.
Quarter 3	COMPLETED	PASS	23 Sept 2026, 20:30	Position 3	Approximately one quarter of the load-receptor surface.
Quarter 4	COMPLETED	PASS	23 Sept 2026, 20:30	Position 4	Approximately one quarter of the load-receptor surface.

Observations

Load L	Indication I	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
10,000 g	10,000 g	5 g	0 g	10 g	g

Observations

Load L	Indication I	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
10,000 g	10,000 g	5 g	0 g	10 g	g

Observations

Load L	Indication I	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
10,000 g	10,000 g	5 g	0 g	10 g	g

Observations

Load L	Indication I	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
10,000 g	10,000 g	5 g	0 g	10 g	g

A.4.8 - DISCRIMINATION

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.8-1.0
Method	A.4.8.2
Technical reference	OIML R 76-1:2006 Annex A A.4.8.2
Started	23 Sept 2026, 20:30
Completed	23 Sept 2026, 20:33
Support / applicability reason	The digital discrimination procedure is applicable for this configured instrument.

Field	Recorded information
Test procedure version	R76-A4.8-1.0
Method / reference	OIML R 76-1:2006 Annex A A.4.8.2
Method	A.4.8.2
Workflow status	COMPLETED

Field	Recorded information
Result	PASS

Stages

Label	Workflow status	Result	Completed	Order	Recommended Increment Count
Min	COMPLETED	PASS	23 Sept 2026, 20:33	1	10
½ Max	COMPLETED	PASS	23 Sept 2026, 20:33	2	10
Max	COMPLETED	PASS	23 Sept 2026, 20:33	3	10

Target Load

Field	Recorded information
Value	20
Unit	g

One Tenth D

Field	Recorded information
Value	0.5
Unit	g

One Point Four D

Field	Recorded information
Value	7
Unit	g

Observation

Field	Recorded information
Unit	g
Input Base Load	20
Input Starting Indication	20
Input Removed Additional Load	5
Input Lower Indication	15
Input Restored Increment	0.5
Input Applied Additional Load	7

Field	Recorded information
Input Upper Indication	25
Base Load	20
Starting Indication	20
Additional Increment Count	10
Removed Additional Load	5
Lower Indication	15
Restored Increment	0.5
Applied Additional Load	7
Upper Indication	25
One Tenth D	0.5
One Point Four D	7
Expected Lower Indication	15
Expected Upper Indication	25
Actual Lower Difference	5
Actual Upper Difference	5
Lower Pass	Yes
Upper Pass	Yes
Result	PASS
Technical reference	OIML R 76-1:2006 Annex A A.4.8.2
Reference edition	R76-A4.8-1.0
Notes	Synthetic prototype: starting additions 10 x 0.5 g; observed downward transition after removing 5 g, then restored 0.5 g and added 7 g for the upper transition.
Recorded	23 Sept 2026, 20:33

Target Load

Field	Recorded information
Value	15,000
Unit	g

One Tenth D

Field	Recorded information
-------	----------------------

Field	Recorded information
Value	0.5
Unit	g

One Point Four D

Field	Recorded information
Value	7
Unit	g

Observation

Field	Recorded information
Unit	g
Input Base Load	15,000
Input Starting Indication	15,000
Input Removed Additional Load	5
Input Lower Indication	14,995
Input Restored Increment	0.5
Input Applied Additional Load	7
Input Upper Indication	15,005
Base Load	15,000
Starting Indication	15,000
Additional Increment Count	10
Removed Additional Load	5
Lower Indication	14,995
Restored Increment	0.5
Applied Additional Load	7
Upper Indication	15,005
One Tenth D	0.5
One Point Four D	7
Expected Lower Indication	14,995
Expected Upper Indication	15,005
Actual Lower Difference	5

Field	Recorded information
Actual Upper Difference	5
Lower Pass	Yes
Upper Pass	Yes
Result	PASS
Technical reference	OIML R 76-1:2006 Annex A A.4.8.2
Reference edition	R76-A4.8-1.0
Notes	Synthetic prototype observation at half maximum load; actual displayed transitions recorded at one scale interval.
Recorded	23 Sept 2026, 20:33

Target Load

Field	Recorded information
Value	30,000
Unit	g

One Tenth D

Field	Recorded information
Value	0.5
Unit	g

One Point Four D

Field	Recorded information
Value	7
Unit	g

Observation

Field	Recorded information
Unit	g
Input Base Load	30,000
Input Starting Indication	30,000
Input Removed Additional Load	5
Input Lower Indication	29,995

Field	Recorded information
Input Restored Increment	0.5
Input Applied Additional Load	7
Input Upper Indication	30,005
Base Load	30,000
Starting Indication	30,000
Additional Increment Count	10
Removed Additional Load	5
Lower Indication	29,995
Restored Increment	0.5
Applied Additional Load	7
Upper Indication	30,005
One Tenth D	0.5
One Point Four D	7
Expected Lower Indication	29,995
Expected Upper Indication	30,005
Actual Lower Difference	5
Actual Upper Difference	5
Lower Pass	Yes
Upper Pass	Yes
Result	PASS
Technical reference	OIML R 76-1:2006 Annex A A.4.8.2
Reference edition	R76-A4.8-1.0
Notes	Synthetic prototype observation at maximum load; actual displayed transitions recorded at one scale interval.
Recorded	23 Sept 2026, 20:33

A.4.9 - SENSITIVITY OF NON-SELF-INDICATING INSTRUMENT

TEST RESULT: NOT APPLICABLE

Field	Recorded information
Status	NOT APPLICABLE

Field	Recorded information
Reason	Not applicable: A.4.9 is the sensitivity test for a non-self-indicating instrument.

A.4.10 - REPEATABILITY

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.10-1.0
Method	Same Load Repeat Weighings
Technical reference	OIML R 76-1:2006 Section 3.6.1 / Annex A A.4.10
Started	23 Sept 2026, 20:34
Completed	23 Sept 2026, 20:34
Support / applicability reason	The repeatability procedure is included for the configured NAWI test route.

Field	Recorded information
Test procedure version	R76-A4.10-1.0
Method / reference	OIML R 76-1:2006 Section 3.6.1 / Annex A A.4.10
Control Stage	VERIFICATION
Method	Same Load Repeat Weighings
Workflow status	COMPLETED
Result	PASS
Zero Reference E0	0
Observation Unit	g

Procedure Confirmation

Field	Recorded information
Automatic Zero On Confirmed	Yes
Unloaded Instrument Rest Confirmed	Yes
Confirmed At	23 Sept 2026, 20:34
Tester Name Snapshot	Regression QA Tester Four

Series

Label	Workflow status	Result	Completed	Series Id	Order
Approximately 80% of Max	COMPLETED	PASS	23 Sept 2026, 20:34	Series 1	1

Target Load

Field	Recorded information
Value	24,000
Unit	g

Observations

Indication	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit	Notes
24,000 g	0 g	5 g	15 g	g	Synthetic prototype/regression observation; same 24,000 g load and stable indication used for repeatability simulation.
24,000 g	0 g	5 g	15 g	g	Synthetic prototype/regression reading 2; stable 24,000 g indication.
24,000 g	0 g	5 g	15 g	g	Synthetic prototype/regression reading 3; stable 24,000 g indication completes the repeatability series.

Summary

Field	Recorded information
Target Mpe Rule	OIML R 76-1:2006 Section 3.5.1 Table 6
Minimum Result	24,005
Maximum Result	24,005
Repeatability Range	0
Absolute Mpe	15
Result	PASS

Target Mpe

Field	Recorded information
Value	15
Unit	g

Maximum permissible error

Field	Recorded information
Value	15
Unit	g

A.4.11 - VARIATION OF INDICATION WITH TIME

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.11-1.0
Method	A.4.11.1 + A.4.11.2
Technical reference	OIML R 76-1:2006 Section 3.9.4 / Annex A A.4.11
Started	23 Sept 2026, 20:35
Completed	23 Sept 2026, 21:47
Support / applicability reason	The creep and zero-return procedures apply to accuracy Classes II, III and IIII.

Field	Recorded information
Test procedure version	R76-A4.11-1.0
Method / reference	OIML R 76-1:2006 Section 3.9.4 / Annex A A.4.11
Method	A.4.11.1 + A.4.11.2
Workflow status	COMPLETED
Result	PASS
Observation Unit	g

Plan

Field	Recorded information
Method	A.4.11.1 + A.4.11.2
Execution Supported	Yes
Method / reference	OIML R 76-1:2006 Section 3.9.4 / Annex A A.4.11

Recommended Load

Field	Recorded information
-------	----------------------

Field	Recorded information
Value	30,000
Unit	g

Creep Checkpoints

Checkpoint	Minutes
T0	0
T5	5
T15	15
T30	30
T60	60
T120	120
T180	180
T240	240

Zero Return Limit

Field	Recorded information
Value	5
Unit	g
Rule	0.5e for a single-range, single-interval instrument.

Procedure Confirmation

Field	Recorded information
Normal Oscillation Confirmed	Yes
Confirmed At	23 Sept 2026, 20:35
Tester Name Snapshot	Regression QA Tester Four

Environmental Readings

20 - 20 - 20 - 20

Creep

Field	Recorded information
Workflow status	COMPLETED

Field	Recorded information
Result	PASS
Started	23 Sept 2026, 20:35
Completed	23 Sept 2026, 21:16

Observation

Field	Recorded information
Actual Load	30,000
Input Actual Load	30,000
Unit	g
Indication	30,000
Indication Unit	g
Initial Delta L	0
Initial P	30,005
P0	30,005
P5	30,005
P15	30,005
P30	30,005
Maximum permissible error	15
MPE unit	g
Mpe Rule Reference	OIML R 76-1:2006 Section 3.5.1 Table 6
Mpe Rule Set Id	R76-1:2006:TABLE-6:INITIAL-VERIFICATION
Mpe Rule Version	R76-1:2006
Source Clause	OIML R 76-1:2006 Section 3.9.4 / Annex A A.4.11
Delta30	0
Delta15 30	0
Early Termination Limit	5
Stability Limit	2
Temperature Variation	0 deg C
Temperature Condition	SATISFIED
Early Termination Allowed	Yes

Field	Recorded information
Required Checkpoint	T30
Result	PASS
Recorded	23 Sept 2026, 21:16
Tester Name Snapshot	Regression QA Tester Four
Notes	Synthetic prototype/regression data; stable near-capacity indication, laboratory conditions held at 20 deg C.

Checkpoints

Indication	Additional load (Delta L)	Unit	Checkpoint	Minutes	P
30,000 g	0 g	g	T0	0	30,005
30,000 g	0 g	g	T5	5	30,005
30,000 g	0 g	g	T15	15	30,005
30,000 g	0 g	g	T30	30	30,005

Zero Return

Field	Recorded information
Workflow status	COMPLETED
Result	PASS
Started	23 Sept 2026, 21:16
Completed	23 Sept 2026, 21:47

Observation

Field	Recorded information
Actual Load	30,000
Input Actual Load	30,000
Unit	g
Zero Before	0
Zero After	0
Input Zero After	0
Zero After Unit	g
Zero Return Deviation	0

Field	Recorded information
Allowed Limit	5
Zero Tracking Off Confirmed	Yes
Stabilized Before	Yes
Stabilized After	Yes
Result	PASS
Recorded	23 Sept 2026, 21:47
Tester Name Snapshot	Regression QA Tester Four

A.4.12 - STABILITY OF EQUILIBRIUM

TEST RESULT: FAIL

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A4.12-1.0
Method	Stability Of Equilibrium
Technical reference	OIML R 76-1:2006 Section 4.4.2 / Annex A A.4.12
Started	23 Sept 2026, 21:47
Completed	23 Sept 2026, 21:55
Support / applicability reason	The configured stable-equilibrium function is evaluated using the applicable branches. Applicable stable-equilibrium functions: printing, data storage, zero-setting, tare.

Field	Recorded information
Test procedure version	R76-A4.12-1.0
Method / reference	OIML R 76-1:2006 Section 4.4.2 / Annex A A.4.12
Method	Stability Of Equilibrium
Workflow status	COMPLETED
Result	FAIL

Plan

Field	Recorded information
Differentiated Scale Divisions	No
Documentation Required	Yes

Field	Recorded information
Method / reference	OIML R 76-1:2006 Section 4.4.2 / Annex A A.4.12

Target Load

Field	Recorded information
Value	15,000
Unit	g

Print Storage Interval

Field	Recorded information
Value	10
Unit	g
Kind	1e

Zero Tare Accuracy Limit

Field	Recorded information
Value	2.5
Unit	g
Kind	0.25e

Applicable Operations

PRINT - STORE - ZERO - TARE

Phases

Field	Recorded information
Documentation	Yes
Print Storage	Yes
Zero Setting	Yes
Tare	Yes
Continuous Disturbance	Yes
Mobile	No

Phases

Code	Name	Applicability	Workflow status	Result	Reason
A.4.12.1	Stability documentation/ configuration review	APPLICABLE	COMPLETED	FAIL	Review the manufacturer documentation for the stable-equilibrium principle, criteria, parameters, and security.
A.4.12.2	Printing / data-storage stability	APPLICABLE	COMPLETED	PASS	Test the five-second indication sequence after a disturbance and the print/data-storage command.
A.4.12.3	Zero-setting stability	APPLICABLE	COMPLETED	PASS	Check inhibition before stable equilibrium and reuse A.4.2.3 evidence once stable.
A.4.12.4	Tare-balancing stability	APPLICABLE	COMPLETED	PASS	Check inhibition before stable equilibrium and reuse A.4.6.2 evidence once stable.
A.4.12.5	Continuous-disturbance inhibition	APPLICABLE	COMPLETED	PASS	Record whether every applicable function is inhibited while continuous disturbance remains active.
A.4.12.6	Mobile/vehicle branch	Not applicable	Not applicable	Not applicable	Not applicable - the instrument is configured as stationary.

Dependency

Field	Recorded information
Code	A.4.2.3

Observations

Result	Additional load (Delta L)	Unit	Repetition	Disturbance Confirmed	Operation Attempted Immediately
PASS	5 g	g	1	Yes	Yes
PASS	5 g	g	2	Yes	Yes
PASS	5 g	g	3	Yes	Yes
PASS	5 g	g	4	Yes	Yes
PASS	5 g	g	5	Yes	Yes

Dependency

Field	Recorded information
Code	A.4.6.2

Observations

Result	Additional load (Delta L)	Unit	Repetition	Disturbance Confirmed	Operation Attempted Immediately
PASS	5 g	g	1	Yes	Yes

Result	Additional load (Delta L)	Unit	Repetition	Disturbance Confirmed	Operation Attempted Immediately
PASS	5 g	g	2	Yes	Yes
PASS	5 g	g	3	Yes	Yes
PASS	5 g	g	4	Yes	Yes
PASS	5 g	g	5	Yes	Yes

Documentation

Field	Recorded information
Manufacturer Documentation Details	No manufacturer-authored stable-equilibrium documentation was supplied with this synthetic prototype. No claim is made about undocumented internal criteria, adjustable or non-adjustable parameters, parameter security, or worst-case adjustment.
Documentation Available	No
Basic Principle Documented	No
Criteria Documented	No
Adjustable Parameters Documented	No
Non Adjustable Parameters Documented	No
Parameter Security Documented	No
Worst Case Adjustment Identified	No
Manufacturer Documentation Reference	Not supplied - synthetic QA profile
Reviewed At	23 Sept 2026, 21:48
Tester Name Snapshot	Regression QA Tester Four

Setup

Field	Recorded information
Input Actual Load	15,000
Input Unit	g
Recorded	23 Sept 2026, 21:48
Tester Name Snapshot	Regression QA Tester Four

Target Load

Field	Recorded information
Value	15,000

Field	Recorded information
Unit	g

Actual Load

Field	Recorded information
Value	15,000
Unit	g

Print Storage

Field	Recorded information
Function Tested	PRINT
Disturbance Confirmed	Yes
Initiated Immediately Confirmed	Yes
Workflow status	COMPLETED
Tester Name Snapshot	Regression QA Tester Four
Required Repetitions	5
Result	PASS

Repetitions

Result	Repetition	Function Tested	Input Printed Value	Printed Value Unit	Observed Value Unit
PASS	1	PRINT	15,000	g	g
PASS	2	PRINT	15,000	g	g
PASS	3	PRINT	15,000	g	g
PASS	4	PRINT	15,000	g	g
PASS	5	PRINT	15,000	g	g

Printed Value

Field	Recorded information
Value	15,000
Unit	g

Observed Values

Unit	Value
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g

Input Observed Values

15,000 - 15,000 - 15,000 - 15,000 - 15,000

Evaluation

Field	Recorded information
Adjacent Value Count	1
Printed Value Included	Yes
Interval	10
Criterion Satisfied	Yes
Result	PASS
Minimum Value	15,000
Maximum Value	15,000

Printed Value

Field	Recorded information
Value	15,000
Unit	g

Observed Values

Unit	Value
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g

Input Observed Values

15,000 - 15,000 - 15,000 - 15,000 - 15,000

Evaluation

Field	Recorded information
Adjacent Value Count	1
Printed Value Included	Yes
Interval	10
Criterion Satisfied	Yes
Result	PASS
Minimum Value	15,000
Maximum Value	15,000

Printed Value

Field	Recorded information
Value	15,000
Unit	g

Observed Values

Unit	Value
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g

Input Observed Values

15,000 - 15,000 - 15,000 - 15,000 - 15,000

Evaluation

Field	Recorded information
Adjacent Value Count	1
Printed Value Included	Yes
Interval	10

Field	Recorded information
Criterion Satisfied	Yes
Result	PASS
Minimum Value	15,000
Maximum Value	15,000

Printed Value

Field	Recorded information
Value	15,000
Unit	g

Observed Values

Unit	Value
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g

Input Observed Values

15,000 - 15,000 - 15,000 - 15,000 - 15,000

Evaluation

Field	Recorded information
Adjacent Value Count	1
Printed Value Included	Yes
Interval	10
Criterion Satisfied	Yes
Result	PASS
Minimum Value	15,000
Maximum Value	15,000

Printed Value

Field	Recorded information
Value	15,000
Unit	g

Observed Values

Unit	Value
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g
g	15,000 g

Input Observed Values

15,000 - 15,000 - 15,000 - 15,000 - 15,000

Evaluation

Field	Recorded information
Adjacent Value Count	1
Printed Value Included	Yes
Interval	10
Criterion Satisfied	Yes
Result	PASS
Minimum Value	15,000
Maximum Value	15,000

Stability Interval

Field	Recorded information
Value	10
Unit	g
Kind	1e

Zero Setting

Field	Recorded information
-------	----------------------

Field	Recorded information
Required Repetitions	5
Result	PASS

Repetitions

Result	Additional load (Delta L)	Unit	Repetition	Disturbance Confirmed	Operation Attempted Immediately
PASS	5 g	g	1	Yes	Yes
PASS	5 g	g	2	Yes	Yes
PASS	5 g	g	3	Yes	Yes
PASS	5 g	g	4	Yes	Yes
PASS	5 g	g	5	Yes	Yes

Tare

Field	Recorded information
Required Repetitions	5
Result	PASS

Repetitions

Result	Additional load (Delta L)	Unit	Repetition	Disturbance Confirmed	Operation Attempted Immediately
PASS	5 g	g	1	Yes	Yes
PASS	5 g	g	2	Yes	Yes
PASS	5 g	g	3	Yes	Yes
PASS	5 g	g	4	Yes	Yes
PASS	5 g	g	5	Yes	Yes

Continuous Disturbance

Field	Recorded information
Continuous Disturbance Confirmed	Yes
Result	PASS
Recorded	23 Sept 2026, 21:55
Tester Name Snapshot	Regression QA Tester Four

Operations

Operation	Observed
PRINT	BLOCKED
STORE	BLOCKED
ZERO	BLOCKED
TARE	BLOCKED

A.4.13 - PORTABLE WEIGHBRIDGE TESTS

TEST RESULT: NOT APPLICABLE

Field	Recorded information
Status	NOT APPLICABLE
Reason	The instrument is not a portable road-vehicle weighbridge.

A.5 - INFLUENCE FACTORS

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A5-1.0
Technical reference	OIML R 76-1:2006 Annex A A.5
Started	23 Sept 2026, 21:56
Completed	23 Sept 2026, 22:01
Support / applicability reason	The generated A.5 branch set is derived from the persisted instrument configuration.

Field	Recorded information
Test procedure version	R76-A5-1.0
Method / reference	OIML R 76-1:2006 Annex A A.5
Workflow status	COMPLETED
Result	PASS

Plan

Field	Recorded information
Method / reference	OIML R 76-1:2006 Annex A A.5

Target Load

Field	Recorded information
Value	15,000
Unit	g
Label	0.50 x Max

Voltage

Field	Recorded information
Supported	Yes
Branch	A.5.4.1
Reference	230
Lower	175.95
Upper	278.3
Formula	$0.85 U_{min}$ to $1.10 U_{max}$
Three Phase	No

Temperature

Field	Recorded information
Supported	Yes
Reference Temperature	20 deg C
Two Hour Stabilization Minutes	120
Maximum Change Rate Celsius Per Minute	1
High Temperature Absolute Humidity Limit GPer M3	20 deg C

Targets

Label	Value
Reference temperature	20
Specified high temperature	40
Specified low temperature	-10

Label	Value
5 deg C	5
Reference temperature	20

Warm Up Checkpoints

5 - 15 - 30

Tilt

Field	Recorded information
Workflow status	APPLICABLE
Reason	Use the manufacturer limiting tilt with the configured level indicator/automatic tilt sensor.
Method	A.5.1.1
Execution Supported	Yes
Tilt Limit	2

Directions

LONGITUDINAL - TRANSVERSE

Test Loads

Label	Unit	Value
10e	g	100 g
Between 1/2 Max and Max	g	15,000 g

Phases

Code	Name	Applicability	Workflow status	Result	Reason
A.5.1	Tilting	APPLICABLE	COMPLETED	PASS	Use the manufacturer limiting tilt with the configured level indicator/automatic tilt sensor.
A.5.2	Warm-up time	APPLICABLE	COMPLETED	PASS	The instrument uses electric power and is in scope for the warm-up test.
A.5.3	Temperature tests	APPLICABLE	COMPLETED	PASS	The specified operating-temperature range is available for the A.5.3 procedure.
A.5.4	Voltage variations	APPLICABLE	COMPLETED	PASS	The configured power-source branch determines the A.5.4 voltage limits.

Setup

Field	Recorded information
Environmental Conditions Stable Confirmed	Yes
Notes	Synthetic Prototype / Regression Data. Stable ambient conditions are represented for this software workflow test; not a physical laboratory observation.
Recorded	23 Sept 2026, 21:57
Tester Name Snapshot	Regression QA Tester Four

Tilting

Field	Recorded information
Zero Setting Off Confirmed	Yes
Result	PASS
Recorded	23 Sept 2026, 21:59
Tester Name Snapshot	Regression QA Tester Four

Observations

Load	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error	Calculated indication
0	0	0	5	5	5
100	100	0	5	5	105
30,000	30,000	0	5	15	30,005
0	0	0	5	5	5
100	100	0	5	5	105
30,000	30,000	0	5	15	30,005

Raw

Field	Recorded information
Direction	LONGITUDINAL
Position	No Load
Required Tilt	2
Actual Tilt	2
Load	0
Indication	0
Additional load (Delta L)	0

Field	Recorded information
Zero Error	0
Unit	g

Raw

Field	Recorded information
Direction	LONGITUDINAL
Position	Low Load
Required Tilt	2
Actual Tilt	2
Load	100
Indication	100
Additional load (Delta L)	0
Zero Error	0
Unit	g

Raw

Field	Recorded information
Direction	LONGITUDINAL
Position	Max Load
Required Tilt	2
Actual Tilt	2
Load	30,000
Indication	30,000
Additional load (Delta L)	0
Zero Error	0
Unit	g

Raw

Field	Recorded information
Direction	TRANSVERSE
Position	No Load

Field	Recorded information
Required Tilt	2
Actual Tilt	2
Load	0
Indication	0
Additional load (Delta L)	0
Zero Error	0
Unit	g

Raw

Field	Recorded information
Direction	TRANSVERSE
Position	Low Load
Required Tilt	2
Actual Tilt	2
Load	100
Indication	100
Additional load (Delta L)	0
Zero Error	0
Unit	g

Raw

Field	Recorded information
Direction	TRANSVERSE
Position	Max Load
Required Tilt	2
Actual Tilt	2
Load	30,000
Indication	30,000
Additional load (Delta L)	0
Zero Error	0
Unit	g

Warm Up

Field	Recorded information
Disconnected At	23 Sept 2026, 00:00
Connected At	23 Sept 2026, 08:00
Stabilization At	23 Sept 2026, 08:05
Eight Hour Precondition Confirmed	Yes
Result	PASS

Observations

Result	Load	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error
PASS	15,000 g	15,000 g	0 g	5 g	10 g
PASS	15,000 g	15,000 g	0 g	5 g	10 g
PASS	15,000 g	15,000 g	0 g	5 g	10 g

Temperature

Field	Recorded information
Result	PASS
Recorded	23 Sept 2026, 22:00

Points

Load	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
15,000 g	15,000 g	0 g	5 g	10 g	g
15,000 g	15,000 g	0 g	5 g	10 g	g
15,000 g	15,000 g	0 g	5 g	10 g	g
15,000 g	15,000 g	0 g	5 g	10 g	g
15,000 g	15,000 g	0 g	5 g	10 g	g

Voltage

Field	Recorded information
Result	PASS
Recorded	23 Sept 2026, 22:01

Observations

Label	Load	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error
Reference Start	100 g	100 g	0 g	5 g	5 g
LOWER	100 g	100 g	0 g	5 g	5 g
UPPER	15,000 g	15,000 g	0 g	5 g	10 g
Reference End	15,000 g	15,000 g	0 g	5 g	10 g

A.6 - ENDURANCE TEST

TEST RESULT: PASS

Field	Recorded information
Workflow status	Completed
Applicability	APPLICABLE
Test version	R76-A6-1.0
Technical reference	OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6
Started	23 Sept 2026, 22:01
Completed	23 Sept 2026, 22:03
Support / applicability reason	The instrument is Class II, III or IIII with Max not exceeding 100 kg.

Recovery checkpoints

Applications	Checkpoint time	Actual load	Recorded by	Notes
0	23 Sept 2026, 22:01	Not recorded	Regression QA Tester Four	Session created; software checkpoint only.
10,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
25,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
50,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
75,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.
100,000	23 Sept 2026, 22:02	Not recorded g	Regression QA Tester Four	Software recovery checkpoint; not an additional OIML interval.

Endurance calculation

Field	Recorded information
-------	----------------------

Field	Recorded information
Required applications	100,000
Recorded applications	100,000
Synthetic applications	100,000
Pre-endurance baseline error	0 g
Post-endurance corrected error	5 g
Durability error	5 g
Abnormal events	0
Cycle state	COMPLETED

Field	Recorded information
Test procedure version	R76-A6-1.0
Method / reference	OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6
Workflow status	COMPLETED
Result	PASS
Required loading applications	100,000
Recorded loading applications	100,000
Synthetic loading applications	100,000
Cycle State	COMPLETED
Resumed At	23 Sept 2026, 22:01

Applicability

Field	Recorded information
Code	A.6
Name	Endurance test
Order	14
Route	A.6
Method / reference	OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6
Workflow status	APPLICABLE
Reason	The instrument is Class II, III or IIII with Max not exceeding 100 kg.
Method	ENDURANCE
Method Label	Endurance after all other tests

Field	Recorded information
Execution Supported	Yes
Clause	A.6
Accuracy Class	III
Maximum capacity	30,000
Unit	g

Target Load

Field	Recorded information
Value	15,000
Unit	g
Label	Approximately 50% of Max

Checkpoints

Timestamp	Notes	Cycle Number	Operator
23 Sept 2026, 22:01	Session created; software checkpoint only.	0	Regression QA Tester Four
23 Sept 2026, 22:02	Software recovery checkpoint; not an additional OIML interval.	10,000	Regression QA Tester Four
23 Sept 2026, 22:02	Software recovery checkpoint; not an additional OIML interval.	25,000	Regression QA Tester Four
23 Sept 2026, 22:02	Software recovery checkpoint; not an additional OIML interval.	50,000	Regression QA Tester Four
23 Sept 2026, 22:02	Software recovery checkpoint; not an additional OIML interval.	75,000	Regression QA Tester Four
23 Sept 2026, 22:02	Software recovery checkpoint; not an additional OIML interval.	100,000	Regression QA Tester Four

Actual Load

Field	Recorded information
Value	15,000
Unit	g
Input Value	15,000
Input Unit	g

Actual Load

Field	Recorded information
Value	15,000
Unit	g
Input Value	15,000
Input Unit	g

Actual Load

Field	Recorded information
Value	15,000
Unit	g
Input Value	15,000
Input Unit	g

Actual Load

Field	Recorded information
Value	15,000
Unit	g
Input Value	15,000
Input Unit	g

Actual Load

Field	Recorded information
Value	15,000
Unit	g
Input Value	15,000
Input Unit	g

Phases

Code	Name	Workflow status	Result	Completed	Method / reference
A.6.1	Pre-endurance weighing	COMPLETED	PASS	23 Sept 2026, 22:01	OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6
A.6.2	Endurance loading applications	COMPLETED	PASS	23 Sept 2026, 22:02	OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6

Code	Name	Workflow status	Result	Completed	Method / reference
A.6.3	Post-endurance weighing	COMPLETED	PASS	23 Sept 2026, 22:03	OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6
A.6.4	Durability assessment	COMPLETED	PASS	23 Sept 2026, 22:03	OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6

Pre Weighing

Field	Recorded information
Source Test Id	6ab3e8aa3d6633777927cf6f
Source Sequence	9
Load	15,000
Unit	g
Calculated indication	15,000
Error	0
Corrected error	0
Maximum permissible error	10
Recorded	23 Sept 2026, 22:01
Tester Name Snapshot	Regression QA Tester Four
Method / reference	OIML R 76-1:2006 A.4.4.1 / A.4.4.3

Actual Load

Field	Recorded information
Value	15,000
Unit	g
Input Value	15,000
Input Unit	g

Loading Conditions

Field	Recorded information
Actual Load	15,000
Unit	g
Loading Method	MANUAL

Field	Recorded information
Loading Frequency	Synthetic batch increments in supported prototype workflow; not physical endurance loading.
Loading Behavior	Synthetic regression record only. No physical loading cycle was performed; this verifies the batch workflow, count integrity, persistence, and checkpoint behavior.
Loaded Equilibrium Confirmed	Yes
Unloaded Equilibrium Confirmed	Yes
Applied Force Within Normal Operation	Yes
Normal Conditions Confirmed	Yes
Supporting Evidence	SYNTHETIC PROTOTYPE / REGRESSION DATA. Conditions and procedure confirmations below are simulated for software workflow validation; not laboratory evidence.
Normalized Unit	g
Recorded	23 Sept 2026, 22:01
Tester Name Snapshot	Regression QA Tester Four

Durability Assessment

Field	Recorded information
Supported	Yes
Durability error	5
Maximum permissible error	10
MPE unit	g
Result	PASS
Technical reference	OIML R 76-1:2006 Section 3.9.4.3
Mpe Reference	OIML R 76-1:2006 Section 3.5.1 Table 6

Post Weighing

Field	Recorded information
Load	15,000
Indication	15,000
Additional load (Delta L)	0
Zero Error	0
Unit	g
Notes	SYNTHETIC PROTOTYPE / REGRESSION DATA. Matched post-endurance software-workflow observation; no physical endurance loading or weighing was performed.

Field	Recorded information
Calculated indication	15,005
Error	5
Corrected error	5
P	15,005
E	5
E0	0
Ec	5
Recorded	23 Sept 2026, 22:03
Tester Name Snapshot	Regression QA Tester Four
Method / reference	OIML R 76-1:2006 A.4.4.1 / A.4.4.3

Maximum permissible error

Field	Recorded information
Supported	Yes
Accuracy Class	III
Load	15,000
Verification scale interval (e)	10
M	1,500
Mpe Multiplier	1
Maximum permissible error	10
Maximum permissible error	10
MPE unit	g
Display Mpe	+/-10 g
Technical reference	OIML R 76-1:2006 Section 3.5.1 Table 6
Reference edition	R76-1:2006
Range Label	>500e to <=2000e
Load Type	GROSS

REMARKS, DEVIATIONS AND NON-CONFORMITIES

Synthetic fixture for prototype regression; no physical instrument claim.

Synthetic fixture for prototype regression; no physical instrument claim.

SYNTHETIC PROTOTYPE / REGRESSION DATA. Session end conditions are simulated solely to verify the report handoff workflow; no physical laboratory test session occurred.

Rejected: A.4.12.1 stability documentation/configuration review is marked FAIL because no manufacturer-authored stable-equilibrium documentation or reference was supplied. The remaining applicable A.4.12 phases and A.5/A.6 calculations are PASS, but the report-level result is FAIL. Please provide genuine supporting manufacturer documentation through the appropriate future workflow; no measurement data were changed.

OVERALL CONCLUSION

Overall evaluation result: FAIL

The persisted evaluation record contains one or more applicable test results that do not satisfy their recorded acceptance criteria.

REVIEW AND APPROVAL

Role	Name	Action / decision	Date
Tester	Regression QA Tester Four	Submitted for review	23 Sept 2026, 22:07
Reviewer	samarth dubey	Rejected	23 Sept 2026, 22:09

Reviewer note

Rejected: A.4.12.1 stability documentation/configuration review is marked FAIL because no manufacturer-authored stable-equilibrium documentation or reference was supplied. The remaining applicable A.4.12 phases and A.5/A.6 calculations are PASS, but the report-level result is FAIL. Please provide genuine supporting manufacturer documentation through the appropriate future workflow; no measurement data were changed.

Review history

Action	Recorded by	Date
Submitted for reviewer assessment	Regression QA Tester Four	23 Sept 2026, 22:07
Rejected by reviewer	samarth dubey	23 Sept 2026, 22:09

TRACEABILITY AND REVISION

Field	Recorded information
Applicable standard	OIML R 76-1:2006
Report structure reference	OIML R 76-2, where applicable to the stored report structure
Data source	Persisted test report and test records
Revision	01
Issue / generation date	24 Sept 2026, 19:34