

CONTROLLED TECHNICAL DOCUMENT

NAWI TEST REPORT

Type evaluation and verification record

FINAL REPORT

TR-2026-000042

PASS

DEMONSTRATION / SYNTHETIC DATA

Synthetic observations are identified as such; reported results are the persisted application calculations.

REPORT AND INSTRUMENT IDENTIFICATION		
Report number TR-2026-000042	Document status FINAL REPORT	Workflow status Approved / completed
Manufacturer gym	Applicant meril	Model / type designation gym
Serial number asdasd	Accuracy class / indication Class II / Self-indicating	Capacity 200 - 1,000 g
e / d 10 / 5 g	Tester samarth dubey	Reviewer samarth dubey
Applicable standard OIML R 76-1:2006	Revision 01	Issue date 24 Sept 2026, 20:33

APPLICANT AND MANUFACTURER

Field	Recorded information
Applicant contact	mrp
Email	meril@gmail.com
Telephone	+91 4234234234
Applicant address	dasdasd
Manufacturer address	dasdasd

INSTRUMENT TECHNICAL SPECIFICATIONS

Field	Recorded information
Range / interval configuration	Single-Range / Single-Interval
Power source	AC Mains
Nominal voltage	230 V
Software version	eweqw
Load cell information	qweqw
Tare configuration	Subtractive - Semi-automatic
Maximum tare effect	500 g
Stable equilibrium function	No
Multiple indicating devices	Yes
Printing capability	No
Data storage capability	No

TEST CONDITIONS

Parameter	Recorded start	Recorded end	Unit
Temperature	34	20	deg C
Relative humidity	34	50	%
Atmospheric pressure	41	1,013.25	hPa

Field	Recorded information
Evaluation started	24 Sept 2026, 00:04

Field	Recorded information
Evaluation completed	24 Sept 2026, 18:04
Environmental notes	Synthetic Prototype / Regression Data - simulated end-of-session conditions for workflow verification; not a physical laboratory observation.

TEST EQUIPMENT AND CALIBRATION REFERENCES

No test equipment or calibration reference information is recorded.

APPLICABLE TEST SUMMARY

Test ID	Test description	Status	Result	Non-applicability reason
A.4.2	Checking of zero	Completed	PASS	
A.4.3	Setting to zero before loading	Completed	PASS	
A.4.4	Determination of weighing performance	Completed	PASS	
A.4.5	Multiple indicating devices	Completed	PASS	
A.4.6	Tare	Completed	PASS	
A.4.7	Eccentricity	Completed	PASS	
A.4.8	Discrimination	Completed	PASS	
A.4.9	Sensitivity of non-self-indicating instrument	NOT APPLICABLE	NOT APPLICABLE	Not applicable: A.4.9 is the sensitivity test for a non-self-indicating instrument.
A.4.10	Repeatability	Completed	PASS	
A.4.11	Variation of indication with time	Completed	PASS	
A.4.12	Stability of equilibrium	NOT APPLICABLE	NOT APPLICABLE	The instrument does not have a configured stable-equilibrium function.
A.4.13	Portable weighbridge tests	NOT APPLICABLE	NOT APPLICABLE	The instrument is not a portable road-vehicle weighbridge.
A.5	Influence Factors	Completed	PASS	
A.6	Endurance test	Completed	PASS	

DETAILED TEST RESULTS

A.4.2 - CHECKING OF ZERO

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.2	Started 24 Sept 2026, 00:04
Completed 24 Sept 2026, 03:16		

A.4.3 - SETTING TO ZERO BEFORE LOADING

TEST RESULT: PASS

Method Automatic zero-setting / zero-tracking	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.3(b)
Started 24 Sept 2026, 03:16	Completed 24 Sept 2026, 03:16	

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Method Used	A.4.3(b)	Source Phase	A.4.2.3
Zero Reference Established	Yes	Zero Deviation	0
Zero Deviation Unit	g	Execution Mode	Synthetic Simulation

A.4.4 - DETERMINATION OF WEIGHING PERFORMANCE

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.4	Started 24 Sept 2026, 03:23
Completed 24 Sept 2026, 03:24		

Weighing observations

No.	Applied load / indication / Delta L	P / Error Ec	MPE	Result
1	200 / 200 / 5 g	200 / 0 g	+/-5 g	PASS
2	250 / 250 / 5 g	250 / 0 g	+/-5 g	PASS
3	300 / 300 / 5 g	300 / 0 g	+/-5 g	PASS
4	500 / 500 / 5 g	500 / 0 g	+/-5 g	PASS

No.	Applied load / indication / Delta L	P / Error Ec	MPE	Result
5	600 / 600 / 5 g	600 / 0 g	+/-5 g	PASS
6	1,000 / 1,000 / 5 g	1,000 / 0 g	+/-5 g	PASS
7	950 / 950 / 5 g	950 / 0 g	+/-5 g	PASS
8	900 / 900 / 5 g	900 / 0 g	+/-5 g	PASS
9	750 / 750 / 5 g	750 / 0 g	+/-5 g	PASS
10	210 / 210 / 5 g	210 / 0 g	+/-5 g	PASS

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Zero reference - Unit	g	Zero reference - Zero Indication	0 g
Zero reference - Additional load (Delta L0)	5 g	Zero reference - Calculated E0	0
Zero reference - Recorded	24 Sept 2026, 03:23	Supported	Yes
Observation Unit	g		

A.4.5 - MULTIPLE INDICATING DEVICES

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.6.3; comparison during Annex A A.4.4	Completed 24 Sept 2026, 17:03
------------------------------------	---	---

Indicating-device comparisons

No.	Load	Direction	Device 1	Device 2	Difference	Limit	Result
1	200 g	INCREASING	200 g	200 g	0 g	5 g	PASS
2	250 g	INCREASING	250 g	250 g	0 g	5 g	PASS
3	300 g	INCREASING	300 g	300 g	0 g	5 g	PASS
4	500 g	INCREASING	500 g	500 g	0 g	5 g	PASS
5	600 g	INCREASING	600 g	600 g	0 g	5 g	PASS
6	1,000 g	INCREASING	1,000 g	1,000 g	0 g	5 g	PASS
7	950 g	DECREASING	950 g	950 g	0 g	5 g	PASS
8	900 g	DECREASING	900 g	900 g	0 g	5 g	PASS
9	750 g	DECREASING	750 g	750 g	0 g	5 g	PASS

No.	Load	Direction	Device 1	Device 2	Difference	Limit	Result
10	210 g	DECREASING	210 g	210 g	0 g	5 g	PASS

A.4.6 - TARE

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.6	Started 24 Sept 2026, 17:21
Completed 24 Sept 2026, 17:23		

A.4.6.2 - Accuracy of tare setting

Rep.	Tare load	L0	I0	Delta L	E0	Limit	Result
1	300 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
2	300 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
3	300 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
4	300 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS
5	300 g	100 g	100 g	5 g	0 g	+/-2.5 g	PASS

Authoritative persisted result: PASS. 5 of 5 required repetitions recorded. Last E0: 0 g. Applicable limit: +/-2.5 g.

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Tare Configuration Snapshot - Tare Device Present	Yes	Tare Configuration Snapshot - Tare Type	SUBTRACTIVE
Tare Configuration Snapshot - Maximum Tare Effect - Value	500 g	Tare Configuration Snapshot - Maximum Tare Effect - Unit	g
Tare Configuration Snapshot - Tare Operation Mode	Semi-automatic	Tare Configuration Snapshot - Tare Weighing Device Present	No
Tare Configuration Snapshot - Preset Tare Device Present	No		

A.4.7 - ECCENTRICITY

TEST RESULT: PASS

Method Mobile instrument - applicable four-quarter platform positions	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.7; calculation A.4.4.3
Started 24 Sept 2026, 17:26	Completed 24 Sept 2026, 17:27	

Position observations

Series / phase	Position Id	Label	Loading Area	Geometry
Quarter 1 - POSITION_1	Position 1	Quarter 1	Approximately one quarter of the load-receptor surface.	quarter
Quarter 2 - POSITION_2	Position 2	Quarter 2	Approximately one quarter of the load-receptor surface.	quarter
Quarter 3 - POSITION_3	Position 3	Quarter 3	Approximately one quarter of the load-receptor surface.	quarter
Quarter 4 - POSITION_4	Position 4	Quarter 4	Approximately one quarter of the load-receptor surface.	quarter

Observations

Series / phase	Load L	Indication I	Additional load (Delta L)	Corrected error	Maximum permissible error	Compliance Result
Quarter 1 - POSITION_1	330 g	330 g	5 g	0 g	5 g	PASS
Quarter 2 - POSITION_2	330 g	330 g	5 g	0 g	5 g	PASS
Quarter 3 - POSITION_3	330 g	330 g	5 g	0 g	5 g	PASS
Quarter 4 - POSITION_4	330 g	330 g	5 g	0 g	5 g	PASS

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Execution Supported	Yes	Support Point Count	4
Position Count	4	Zero Condition Confirmed	Yes
Test Load - Value	330 g	Test Load - Unit	g
Sketch - Type	four-quarter-normal-platform		

A.4.8 - DISCRIMINATION

TEST RESULT: PASS

Method A.4.8.2	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.4.8.2
Started 24 Sept 2026, 17:27	Completed 24 Sept 2026, 17:28	

Discrimination observations

Load stage	Target	Starting indication	Lower indication / Delta	Upper indication / Delta	Applied / restored load	d checks	Result
Min	200 g	200 g	195 g / 5 g	205 g / 5 g	7 g / 0.5 g	0.5 g / 7 g	PASS
1/2 Max	500 g	500 g	495 g / 5 g	505 g / 5 g	7 g / 0.5 g	0.5 g / 7 g	PASS
Max	1,000 g	1,000 g	995 g / 5 g	1,005 g / 5 g	7 g / 0.5 g	0.5 g / 7 g	PASS

A.4.9 - SENSITIVITY OF NON-SELF-INDICATING INSTRUMENT

TEST RESULT: NOT APPLICABLE

Field	Recorded information
Status	NOT APPLICABLE
Reason	Not applicable: A.4.9 is the sensitivity test for a non-self-indicating instrument.

A.4.10 - REPEATABILITY

TEST RESULT: PASS

Method Same Load Repeat Weighings	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.6.1 / Annex A A.4.10
Started 24 Sept 2026, 17:28	Completed 24 Sept 2026, 17:29	

Observations

Series / phase	Repetition	Actual Load	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error
Trial 1 / Approximately 80% of Max	1	800 g	800 g	5 g	0 g	5 g
Trial 2 / Approximately 80% of Max	2	800 g	800 g	5 g	0 g	5 g
Trial 3 / Approximately 80% of Max	3	800 g	800 g	5 g	0 g	5 g
Trial 4 / Approximately 80% of Max	4	800 g	800 g	5 g	0 g	5 g
Trial 5 / Approximately 80% of Max	5	800 g	800 g	5 g	0 g	5 g
Trial 6 / Approximately 80% of Max	6	800 g	800 g	5 g	0 g	5 g

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Control Stage	VERIFICATION	Zero Reference E0	0
Observation Unit	g	Procedure Confirmation - Automatic Zero On Confirmed	Yes
Procedure Confirmation - Unloaded Instrument Rest Confirmed	Yes	Procedure Confirmation - Confirmed At	24 Sept 2026, 17:28

A.4.11 - VARIATION OF INDICATION WITH TIME

TEST RESULT: PASS

Method A.4.11.1 + A.4.11.2	Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.9.4 / Annex A A.4.11
Started 24 Sept 2026, 17:29	Completed 24 Sept 2026, 18:31	

Variation and zero-return observations

Checkpoint	Elapsed	Load	Indication	Interpolation Delta L / zero deviation	P	Temperature	Result
Creep: T0	0 min	950 g	950 g	5 g	950 g	20 deg C	PASS
Creep: T5	5 min	950 g	950 g	5 g	950 g	20 deg C	PASS
Creep: T15	15 min	950 g	950 g	5 g	950 g	20 deg C	PASS
Creep: T30	30 min	950 g	950 g	5 g	950 g	20 deg C	PASS
Zero return: Post-load zero-return	After 30 min	-	Before 0; after 0 g	0 g	-	-	PASS

Variation calculation summary

Creep result PASS	Creep Delta P (T0 to T30) 0 g	Creep limit 5 g	Creep temperature variation 0 deg C
Zero return result PASS	Zero return zero-return deviation 0 g	Zero return limit 5 g	Zero return zero tracking off Yes
Zero return readings (before / after) 0 / 0 g	Zero return stabilization before: confirmed; after: confirmed		

A.4.12 - STABILITY OF EQUILIBRIUM

TEST RESULT: NOT APPLICABLE

Field	Recorded information
-------	----------------------

Field	Recorded information
Status	NOT APPLICABLE
Reason	The instrument does not have a configured stable-equilibrium function.

A.4.13 - PORTABLE WEIGHBRIDGE TESTS

TEST RESULT: NOT APPLICABLE

Field	Recorded information
Status	NOT APPLICABLE
Reason	The instrument is not a portable road-vehicle weighbridge.

A.5 - INFLUENCE FACTORS

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Annex A A.5	Started 24 Sept 2026, 18:31
Completed 24 Sept 2026, 18:35		

Influence factor results

Factor	Result	Key value / observation
A.5.1	PASS	Use the 50/1000 tilt procedure because no level indicator or automatic tilt sensor is configured.
A.5.2	PASS	The instrument uses electric power and is in scope for the warm-up test.
A.5.3	PASS	The specified operating-temperature range is available for the A.5.3 procedure.
A.5.4	PASS	The configured power-source branch determines the A.5.4 voltage limits.

Observations

Series / phase	Label	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
A.5	Not recorded	0	0	5	5	Not recorded
A.5	Not recorded	195	0	0	5	Not recorded
A.5	Not recorded	975	0	0	5	Not recorded
A.5	Not recorded	0	0	5	5	Not recorded
A.5	Not recorded	195	0	0	5	Not recorded
A.5	Not recorded	975	0	0	5	Not recorded

Series / phase	Label	Indication	Additional load (Delta L)	Corrected error	Maximum permissible error	Unit
A.5	Not recorded	945 g	0 g	0 g	5 g	g
A.5	Not recorded	945 g	0 g	0 g	5 g	g
A.5	Not recorded	945 g	0 g	0 g	5 g	g
REFERENCE_START	Reference start	95 g	0 g	0 g	5 g	g
LOWER	LOWER	95 g	0 g	0 g	5 g	g
UPPER	UPPER	95 g	0 g	0 g	5 g	g
REFERENCE_END	Reference end	95 g	0 g	0 g	5 g	g
REFERENCE_START	Reference start	695 g	0 g	0 g	5 g	g
LOWER	LOWER	695 g	0 g	0 g	5 g	g
UPPER	UPPER	695 g	0 g	0 g	5 g	g
REFERENCE_END	Reference end	695 g	0 g	0 g	5 g	g

Technical observations and calculations

Parameter	Recorded value	Parameter	Recorded value
Setup - Environmental Conditions Stable Confirmed	Yes	Setup - Recorded	24 Sept 2026, 18:31
Tilting - Zero Setting Off Confirmed	Yes	Tilting - Recorded	24 Sept 2026, 18:33
Warm Up - Disconnected At	22 Sept 2026, 08:00	Warm Up - Connected At	22 Sept 2026, 16:00
Warm Up - Stabilization At	22 Sept 2026, 16:00	Warm Up - Eight Hour Precondition Confirmed	Yes
Temperature - Recorded	24 Sept 2026, 18:34	Voltage - Recorded	24 Sept 2026, 18:35

A.6 - ENDURANCE TEST

TEST RESULT: PASS

Applicability APPLICABLE	Technical reference OIML R 76-1:2006 Section 3.9.4.3 / Annex A A.6	Started 24 Sept 2026, 18:35
Completed 24 Sept 2026, 18:38		

Recovery checkpoints

Applications	Checkpoint time	Actual load	Recorded by	Notes
0	24 Sept 2026, 18:35	Not recorded	samarth dubey	Session created; software checkpoint only.

Applications	Checkpoint time	Actual load	Recorded by	Notes
10,000	24 Sept 2026, 18:36	Not recorded g	samarth dubey	Software recovery checkpoint; not an additional OIML interval.
25,000	24 Sept 2026, 18:36	Not recorded g	samarth dubey	Software recovery checkpoint; not an additional OIML interval.
50,000	24 Sept 2026, 18:36	Not recorded g	samarth dubey	Software recovery checkpoint; not an additional OIML interval.
75,000	24 Sept 2026, 18:36	Not recorded g	samarth dubey	Software recovery checkpoint; not an additional OIML interval.
100,000	24 Sept 2026, 18:37	Not recorded g	samarth dubey	Software recovery checkpoint; not an additional OIML interval.

Endurance calculation

Field	Recorded information
Required applications	100,000
Recorded applications	100,000
Synthetic applications	99,990
Pre-endurance baseline error	0 g
Post-endurance corrected error	0 g
Durability error	0 g
Abnormal events	0
Cycle state	COMPLETED

Post-endurance indication
495 g

REMARKS, DEVIATIONS AND NON-CONFORMITIES

Synthetic Prototype / Regression Data - simulated end-of-session conditions for workflow verification; not a physical laboratory observation.

Reviewed synthetic regression record: all applicable tests and persisted calculations PASS.

OVERALL CONCLUSION

Overall evaluation result: PASS

The persisted evaluation record indicates that all applicable completed tests satisfy their recorded acceptance criteria.

REVIEW AND APPROVAL

Role	Name	Action / decision	Date
------	------	-------------------	------

Role	Name	Action / decision	Date
Tester	samarth dubey	Submitted for review	24 Sept 2026, 18:40
Reviewer	samarth dubey	Approved	24 Sept 2026, 18:42

Reviewer note

Reviewed synthetic regression record: all applicable tests and persisted calculations PASS.

Review history

Action	Recorded by	Date
Submitted for reviewer assessment	samarth dubey	24 Sept 2026, 18:40
Approved by reviewer	samarth dubey	24 Sept 2026, 18:42

TRACEABILITY & REVISION

Technical basis OIML R 76-1:2006	Report format OIML R 76-2 structure	Revision 01
Document status FINAL	Record basis Persisted tester observations and stored calculation results	Generated 24 Sept 2026, 20:33

EVIDENCE APPENDIX

1 persisted supporting record

Evidence 01 - Instrument - Nameplate

Field	Recorded information
Linked test	Report-level evidence
Description	download (2).png
Captured	24 Sept 2026, 00:03
Captured by	samarth dubey
Evidence type	nameplate

