

Section II: Schedule of requirements

This section details the requirements that the goods and services under this tender must meet. It is divided in four sub-sections:

Section II.1. Lists of goods, classification level and requirements.

Section II.2. Certification and documentation requirements.

Section II.3. Related services requirements.

Section II.4. Delivery schedule.

The goods listed in Section II.1. have associated certification and documentation requirements described in Section II.2., related service requirements described in Section II.3. and Delivery schedule as described in Section II.4..

Each section indicates the mandatory documentation that BIDDERS must include with the offers and the documentation to be submitted together with their offer.

Section II.1. Lists of goods and classification level and requirements

Table 1. List of laboratory equipment to be installed; and classification level of complexity.

Lot	Description	Quantities	Level			
			0	1	2	3
LOT 1	Blood gas analyzer	2				x
	Automated hematology analyzer	5				x
	Automated coagulation analyzer	2				x
	Automated biochemistry analyzer	5				x
	Automated immunoassay Analyzers	1				x
	Automated ELISA analyzer	7				x
	Semi-automated coagulation analyzer	5			x	
LOT 2	Gel-Based Automated Electrophoresis System	1				x
	Real-Time PCR machine	2				x
	McFarland densitometer	12			x	
	Laboratory flame spectrophotometer	5			x	
	UV and Visible Spectrophotometer	3			x	
	Horizontal gel electrophoresis	2			X	
	Transilluminator	1			X	

1. General requirements applicable to all the goods to be offered

The equipment offered by the BIDDER, as well as each of its components and accessories, must comply with the following general requirements:

- 1.1. The goods offered must meet or exceed the technical specifications set out in:
 - Annex (Form) D1. Technical Bid Form for the goods described in table 1(Lot 1)
 - Annex (Form) D2. Technical Bid Form for the goods described in table 1 (Lot 2)
- 1.2. All BIDDERS must submit the authorisations, certifications and supporting documentation stated in Section II.2 for all the offered goods, to demonstrate compliance with the minimum technical requirements of UNOPS detailed .
- 1.3. When the BIDDER offers goods with technical specifications that differ from the minimum requirements due to advanced or superior features, the BIDDER must provide a technical report based on technical and/or scientific evidence describing these features so that they can be taken into account in the evaluation.

- 1.4.** Equipment that uses electrical energy must be able to operate at the voltage available in Madagascar (plug type E+F (also known as Type E/F or EF), 230VAC $\pm$ 10%, 50H. In addition, the BIDDER must include in their offer the protective devices required for the equipment in accordance with the manufacturer's recommendations,
- 1.5.** The BIDDER must offer new (unused), state of the art, non-prototype, non-reconditioned, non-testing equipment that is not in the process of being discontinued or becoming obsolete. The equipment must incorporate all recent improvements in design, software and materials. The BIDDER must ensure that the date of manufacture for the supplied equipment, along with all accessories and integral components, does not exceed twenty-four (24) months prior to the contract signature date.
- 1.6.** Offers for equipment classified as surplus, remnant, discontinued or otherwise outdated or designated as 'Export only' will not be accepted. Additionally, equipment models not approved for use in their country of origin will not be accepted. Offers for equipment subject to current medical alerts issued by national or international regulatory bodies—such as those recognized by the International Medical Device Regulators Forum (IMDRF)—will likewise be rejected.
- 1.7.** When applicable, all consumables, reagents, disposables, and similar items supplied with every piece of equipment needed to ensure the optimal functioning of the equipment, must have a minimum shelf life or expiration date of at least one (1) year from the date of commissioning to the installation site or final destination facility.
- 1.8.** In the event that the BIDDER includes in the offer supplies with a useful life of less than one (1) year as specified, the BIDDER must submit, at the time of offer submission, a Commitment Letter undertaking to replace the products as necessary. Supporting documentation—such as a product brochure or written confirmation from the manufacturer—must be provided. A minimum useful life of eight (8) months will be accepted, provided such commitment is formally declared.
- 1.9.** The BIDDER must include in the offer, the required goods, including all components and accessories considered indispensable for the operation and optimal functioning of the goods, even if they are not detailed in the minimum technical specifications described in the Comparative Tables.
- 1.10.** The BIDDER must include in its offer all necessary pre-installation work required for the proper setup and functioning of the offered equipment, in accordance with the manufacturer's recommendations and the conditions set out in Section II.3 Related service requirements (if applicable).
- 1.11.** Details of the technical specifications and minimum technical requirements for each product are given in the Comparative Tables, it is the BIDDER's responsibility to provide clear evidence of compliance with the specified technical requirements. The UNOPS evaluation team reserves the right to consider a non-responsive if explicit and verifiable evidence of compliance with any required specification is not provided.
- 1.12.** For each technical requirement listed in the forms in the Comparative Tables, the BIDDER must include in the column named "Comments/Technical Data" the name/link or reference to the document (catalogue, technical data sheet, user/service manual or other

technical information attached), indicating the specific page where compliance with each line of the technical requirement can be verified.

- 1.13. By submitting their offer, each BIDDER authorizes UNOPS to verify the authenticity of any documentation submitted—whether issued by regulatory authorities, conformity assessment bodies, independent organizations, or clients—directly with the respective issuing entities.
- 1.14. In addition, the BIDDER shall complete in the Comparative Tables, Section A, the details relating to:
 - General information on the goods (manufacturer, model and country of origin);
 - Delivery requirements per unit (estimated delivery time to the point specified in Section II.4. Delivery schedule).
- 1.15. For all the equipment requiring consumables, solutions, reagents, kits, the BIDDER shall provide the complete list of accessories, reagents, calibrators, quality control solutions consumables etc. available and prices. The Supplier shall maintain their availability and prices for at least all the basic warranty duration. Please note that the technical specifications in the Comparative Tables also indicate the mandatory consumables and starting kits of reagents that must be included in the offer. See the Section II.2, 3.5 and Comparative Tables.
- 1.16. Bidders should be aware that, for all open or combined open/closed platform laboratory analysers using generic or brand-specific reagent kits, open channels for generic kits must be available for use from the day the equipment is put into operation.

2. **Specific requirements applicable to all the goods to be offered**

- 2.1. Specific requirements applicable to all the goods to be offered are available in the annexed technical bid forms as Comparative Data Tables.

Section II.2. Authorizations, certification and supporting documentation requirements

This section details the requirements and documentation that the BIDDER, whether manufacturer, representative or distributor, must comply with; the authorisations, certifications and documentation that must be submitted with the offer for each of the offered goods. Compliance with the requirements set out in this Section II.2. applies to all the goods listed in Table 1 of Section II.1 unless differently specified.

Documentation to be provided as a part of the offer

1. Requirements applicable to the manufacturer of the goods being offered

- 1.1. **Manufacturer's authorisation:** If the BIDDER is not the manufacturer of the offered goods, the BIDDER must include with the offer an **authorisation letter from the manufacturer, granting the BIDDER the right to sell, distribute, install and** provide warranty and extended warranty coverage and services for the offered products.

2. Requirements to be met by the local technical support

- 2.1. **Definition of locally based technical services company.** For the offered equipment, it is mandatory that the BIDDER be supported by a locally based technical service company in Madagascar, or in the sub-region (the countries geographically located in the Indian Ocean area and Southern Africa, specifically Mauritius, Réunion Island, and South Africa). This technical service company will be jointly responsible with the BIDDER to provide warranty and extended warranty coverage and related services as specified in **Section II.3. Related services requirements.**
- 2.2. **Market presence.** BIDDERS must have been operational for a minimum of five(5) years,. The BIDDER shall include documentation as part of the submission to demonstrate compliance with this requirement. This documentation shall include at least two (2) copies of service contracts. UNOPS reserves the right to verify the authenticity of the documents or contact the institutions with which the bidder claims to have contracts.
- 2.3. **Technical support experience and capacity.** Local technical support company must have been operational for a minimum of three (3) years. The BIDDER shall include, with the offer, detailed information on the available local technical support:
 - 2.3.1. **Related services** (installation, training, preventive and corrective maintenance, etc.) performed by themselves or local technical support companies.
The BIDDER Include contact information for the institutions they have been serving.
 - 2.3.2. **Technical capacity, experience and qualifications of technical personnel:** the BIDDER must include the list of technical personnel who will carry out all the activities covered by the related services (assembly, installation commissioning, training and technical service), detailing experience in years and related technical or professional training. Minimum of two (2) years of experience in the repair and maintenance of biomedical equipment are required for level 3.

- 2.3.2.1.** For the equipment classified as level 3 (see Table 1, Section II.1), in addition to the list of technical personnel indicating experience and technical or professional training, the BIDDER shall provide at contract signature the technical service training certificates, application training certificate, issued by the manufacturer of the goods offered or supporting documentation of at least two (2) years of experience in the application, repair and maintenance of similar goods.

2.3.2.2.

3. Technical documentation to be submitted for each of the offered lot

- 3.1. Product catalogue, brochure or technical data sheet:** describing the main characteristics and technical data of the product, in the original and in the latest version prepared and published by the manufacturer, in English or French. If the catalogue, brochure, or technical data sheet submitted by the BIDDER includes multiple products, the BIDDER shall clearly identify the specific model or code number of the product being offered. Additionally, the BIDDER must indicate the exact page(s) where the corresponding model/code, as well as any related images, diagrams, schematics, and specifications of the offered product, can be found. This requirement applies to each specification referenced in the proposal.
- 3.2. Details of superior or equivalent characteristics:** In the case where the technical specifications of the offered goods are equivalent or superior to those required in the technical requirements, the BIDDER must provide details of the superior or equivalent characteristics.
- 3.3. Catalog of parts, accessories, spare parts, consumables:** The catalog must clearly indicate the reference codes or part numbers of the products offered such as: accessories, consumables and additional parts.
- 3.4. List of consumables, reagents and kits:** For all the equipment requiring consumables, solutions, reagents, kits, the BIDDER shall provide the complete list of available accessories, reagents, calibrators, quality control solutions, consumables etc. and prices . The Supplier shall maintain their availability and prices for at least all the basic warranty duration. Please note that the technical specifications in the Comparative Tables also indicate the mandatory consumables and starting kits of reagents that must be included in the offer. See the Section II.2, 3.5 and Comparative Tables.

4. Certification requirements related to products, for all the devices:

- 4.1. Environmental management system:** the BIDDER must prove that the manufacturer it represents complies with one of the following options:
The manufacturer has an environmental management system certified under the ISO 14001 standard or equivalent for the factories in which the offered equipment is produced. The above must be verified by one of the following options:
- The factory has an Environmental Management System certified under a standard similar to the reference standard ISO 14001, BIDDER must provide documentation

that confirms that the Environmental Management System implemented by the manufacturer meets or exceeds the reference standard. This documentation may include scientific or technical laboratory test results, technical statements/reports signed by an engineer, or other similar documentation.

- In the event that the manufacturer is not certified, but has an Environmental Management System, the BIDDER must provide documentation from the manufacturer that confirms this fact, identifying how this System aligns with the Sustainable Development Goals established by the General Assembly of the United Nations.
- In the event that the manufacturer does not have an Environmental Management System but has an Action Plan on environmental sustainability, the BIDDER must present documentation from the manufacturer detailing the methodology to increase the organization's sustainability measures through the presentation of a written summary of said methodology.

5. Certification requirements related to IVD and MD products, except for level 0 devices:

5.1. Compliance with a quality management system (QMS) certified ISO 13485:2016 or ISO 9001 or equivalent for the manufacture of the goods offered, in force at the time of the offer. Equivalent or harmonized quality management systems (QMS) with respect to ISO 13485:2016 or ISO 9001 of local regulatory bodies of founding countries of the GHTF (United States, Canada, Japan, Australia, European Union) are valid. The above-mentioned certificate must be issued by CAB (Conformity Assessment Body), by notified bodies or by bodies accredited by the indicated regulatory authorities, must be current and must indicate at least the following:

- Quality standard.
- Name of the certifying body.
- Country of Issue.
- Registration/certificate number.
- Date of issue of the certificate.
- Certificate expiration date.

The document presented must be current, with its respective annexes where the scope of the production line related to the nature of the equipment offered is identified.

5.2. Marketing authorization or recognition from regulatory agencies: the lots offered must meet specific certification or marketing authorization requirements depending on the risk classification of the equipment. These authorizations or certificates must be issued by the regulatory authorities or authorized competent bodies of one of the following countries or regions:

- **European Union:** Regulation (EU) Regulation (EU) 2017/746 In Vitro Diagnostic Medical Devices in the European Union or (EU) 2017/745 Medical Devices in the European Union . Declarations of conformity for class 1 devices.
- United States of America: FDA authorization for marketing in the United States of America

The marketing authorizations must be valid at the time of submission of the offer and be applicable to the brands and models of equipment offered (these must be indicated in the certifications and/or annexes). If they expire within the next six months, The SUPPLIER must present a letter “commitment to deliver the new certificate before the expiration of the current certificate.”

5.3. Specific certifications, if indicated as part of the technical requirements (**Comparative Tables**).

6. Warranty:

6.1. The BIDDER shall specify the mechanism to be used to respond to requests for technical assistance, complaints, repairs or modifications during the warranty period and shall provide full contact details including telephone number, address, name and title of the contact person.

7. Specifications for packaging and storage of the product: The BIDDER shall attach to the offer the packaging specifications, including dimensions and weight, if the equipment contains parts or materials classified as hazardous or requiring special handling (e.g. batteries). In addition, the BIDDER shall indicate the size requirements and/or recommended environmental conditions for temporary storage of the goods.

8. Declaration of acceptance of all the related services requirement

8.1. Declaration of the BIDDER that in case of award all the related services will be provided according to related services requirement.

8.2. Annex I: Declaration of commitment to comply with Warranty conditions

Documentation to be provided as a part of the commissioning of the goods.

- **Local technical support:**

- During the reception of the equipment in the final destination, the information related to the locally based technical service company with a detailed list of the qualified personnel detailed in the offer shall be provided to the beneficiary institution.
- During the warranty and extended warranty periods, no individual other than those expressly listed and presented in the offer by the BIDDER shall perform any servicing of the Equipment, unless the SUPPLIER has provided prior written justification to the beneficiary institution demonstrating that the proposed personnel possess the minimum qualifications and competencies required to carry out such servicing.
- If personnel not listed in the submitted offer are recorded as responsible for servicing a piece of equipment that subsequently fails, the incident may still qualify as a valid warranty claim—regardless of whether the failure is due to a manufacturing defect.

- **Installation requirements:** the SUPPLIER shall provide installation requirements for the laboratory equipment in table 1 of Section II.1 (where applicable), including requirements relating to space, electricity, connectivity, water, drainage, special equipment (furniture, workstations, etc) and infrastructure (if applicable). All requirements needed to install and

optimally operate the equipment shall be included. As the equipment will be installed in containerized laboratories where space is limited, the SUPPLIER shall provide layouts of the equipment to account for its dimensions and specific requirements in English or French.

- **User manual** in English or French. The User manual must contain at least the following information
 - Product Overview.
 - Modules, parts, illustrated and described, including accessories.
 - Functions and operation.
 - Safety instructions.
 - Cleaning and disinfection instructions.
 - Technical specifications.
- **Technical or service manual** in English or French. The technical or service manual must contain at least the following information
 - Product Overview.
 - Modules, parts, illustrated and described, including accessories.
 - Indication of functions and operation.
 - Safety instructions.
 - Cleaning and disinfection instructions.
 - Technical specifications.
 - All technical diagram and necessary access, password, service key and software licences during all the warranty coverage
- **Warranty coverage.** The SUPPLIER and its designated Local technical support shall provide the beneficiary institution with an official Warranty Coverage document, duly signed by a legal representative. This document must fully comply with the coverage requirements specified in **Section II.3. Related services requirements.**

Section II.3. Related services requirements

Related services are understood to be those services, activities or aspects that must be considered prior to and during the acquisition of a medical device to guarantee the correct delivery, installation, commissioning, safe, adequate and sustainable use of the medical equipment.

Table 2 lists the minimum services that the awarded BIDDER, hereafter referred to as the SUPPLIER, shall perform based on the classification level of the laboratory equipment, as indicated in Tables 1 of Section II.1.

The documentation required in numbers 2 to 9 of Table 2 shall not be included with the offer and will be required in the following steps after signature of the contract

Table 2. Related services and documentation requirements.

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
1	Basic Warranty				
	The SUPPLIER shall issue and deliver to the beneficiary institution a warranty valid for one (1) years , covering all parts and service labor related to manufacturing defects. This i	X	X	X	X
1.2	The SUPPLIER shall acknowledge and include in the issued warranty that its validity begins on the date of official receipt of the goods by the beneficiary institution, after testing and commissioning at the location specified in the contract.	x	x	x	x
1.3	The SUPPLIER shall acknowledge and include in the issued warranty the details of the local technical support available in Madagascar, clearly stating that the local technical support entity shares joint and full responsibility for fulfilling all warranty obligations.	x	x	x	x
1.4	The SUPPLIER shall acknowledge and include in the issued warranty that all labor required to diagnose technical issues related to the equipment shall be provided at no additional cost throughout the entire warranty period, regardless of whether the issue is ultimately found to be a manufacturing defect.	x	x	x	x
1.5	The SUPPLIER shall perform all activities related to maintenance at the site where the equipment is installed, at no additional cost to the beneficiary institution. Exceptionally the device can be transported to the Supplier location at Supplier's risk, cost and organization providing that this process is not delaying the diagnostic and repair.	x	x	x	x
1.6	The SUPPLIER shall acknowledge and include in the issued warranty that coverage applies to all delivered items, including accessories, and encompasses all manufacturing defects and faults. The only exception shall be cases of misuse, duly demonstrated by the SUPPLIER.	x	x	x	x
1.7	The SUPPLIER shall implement the support mechanism specified in the offer to respond to requests for technical assistance, complaints, repairs, or modifications during the warranty period. The SUPPLIER shall also provide complete contact information, including telephone number, physical	x	x	x	x

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
	address, and the name and title of the designated contact person.				
1.8	The SUPPLIER shall acknowledge and include in the issued warranty all costs associated with warranty works such as: • All transport, insurance and packaging costs necessary for the transport of equipment/parts/components required for the works related to the warranty. Return, replacement of defective items or parts. • All travel expenses, insurance and associated labour costs required to carry out maintenance/repair work under the warranty. • Any cost of equipment transportation or technician travel expenses shall be at SUPPLIER's charge. These services shall be provided throughout the entire warranty period.	x	x	x	x
1.9	The SUPPLIER shall acknowledge and include in the issued warranty the condition that in the event of an equipment exchange under warranty due to verifiable manufacturing defects, the warranty period for the replacement equipment shall be equal to the original warranty period and shall commence upon delivery and acceptance of the replacement equipment.	x	x	x	x
1.10	The SUPPLIER shall acknowledge and include in the issued warranty that all warranty claims must be resolved within fifteen (15) calendar days from the date of notification.	x	x	x	x
1.11	The SUPPLIER shall provide a clear list of the consumables, reagents and kits required for each product, along with their unit prices. This list will be used by the Government as reference for planning and procurement purposes during the basic warranty period.		x	x	x
1.12	The SUPPLIER must ensure that, from the date on which all open or open/closed platform laboratory analysers are operational, the open channels to allow the use of generic kits of reagents are made available or activated.		x	x	x

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
1.13	Regarding consumables, reagents and kits. If the consumables/kits/reagents to be supplied with the equipment have a sufficient shelf life, but projected usage indicates they may expire before being fully utilized by the MoH, the SUPPLIER shall submit a sworn statement of commitment to replace or subsequently deliver the necessary items to fulfill the original requirements.			x	x
2	Monitoring of production progress				
2.1	The SUPPLIER shall, within a period not exceeding ten (10) calendar days after signing the contract for goods, a manufacturing schedule signed by the manufacturer, indicating the theoretical time required for the manufacturer to have the products in its warehouse ready for delivery.	x	x	x	x
2.2	The SUPPLIER shall inform UNOPS of any delays in the manufacturing process.	x	x	x	x
3	On-site factory inspection				
3.1	The SUPPLIER shall permit and fully cooperate with factory inspections. UNOPS reserves the right to conduct on-site inspections at the SUPPLIER's manufacturing facilities, either directly or through an appointed third party, for the purpose of verifying the manufacturing processes, quality control measures, and the materials used in the production of the goods under contract.	x	x	x	x
4	Delivery				
4.1	The SUPPLIER shall be responsible for the delivery and installation of the equipment at the designated beneficiary institution, in accordance with the timelines, specifications, and conditions set forth in the Contract.		x	x	x
4.2	The SUPPLIER shall, within a period not exceeding ten (10) calendar days from the date of signature of the contract for goods, submit a <u>work plan</u> indicating the schedule for delivery of the goods to the final location(s) agreed with UNOPS. This should include pre- installation and installation requirements for the beneficiary sites.	x	x	x	x
4.3	The SUPPLIER shall ensure that all goods are adequately packaged to guarantee their protection and integrity during	x	x	x	x

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
	loading, handling, transportation, and unloading. The packaging must be suitable for the mode of transport and the nature of the goods, and shall comply with any applicable safety and regulatory requirements.				
4.4	The SUPPLIER shall supply and affix to the equipment a label showing the serial number and identification of the relevant project in accordance with UNOPS requirements.	x	x	x	x
4.5	The SUPPLIER shall transport the equipment inside the beneficiary institution to the installation site/room.	x	x	x	x
4.7	Preliminary inspection and acceptance of the goods				
4.8	The SUPPLIER shall permit and fully cooperate with the preliminary inspection and acceptance process conducted by UNOPS. UNOPS reserves the right to inspect the goods at the manufacturer's warehouse prior to delivery, including verification of quantities against the packing list and assessment of the physical condition and integrity of the packages. Any evidence of damage identified during the inspection shall be documented, photographed, and reported to the SUPPLIER by UNOPS. The SUPPLIER shall be responsible for resolving all such issues at no additional cost.	x	x	x	x
4.9	A representative of the SUPPLIER shall take part in the inspection of the packages prior to shipment to Madagascar.	x	x	x	x
5	Installation				
5.1	The SUPPLIER shall open the packages, assemble and install it according to the installation requirements in the installation place as agreed on with UNOPS		x	x	x
5.2	The SUPPLIER shall clean up the site of any packaging/shipping material after installation.		x	x	x
5.3	The SUPPLIER shall make available all the consumables, measurement and calibration instruments required during commissioning operations.		x	x	x
6	Testing				
6.1	The SUPPLIER shall perform on-site testing, calibration and commissioning of the equipment and its accessories with certification of conformity to standards. After installation, testing and calibration completion, the equipment shall be operational and ready to use.		x	x	x

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
6.2	The SUPPLIER shall perform and include electrical safety testing for all Medical Devices and In Vitro Diagnostic (IVD) Devices supplied.		X	X	X
6.3	The SUPPLIER shall verify and confirm that the equipment is fully functional and free from any physical damage and/or defects prior to proceeding with the commissioning process.	X	X	X	X
6.4	The SUPPLIER shall provide all supplies, accessories, and consumables required for acceptance testing, and shall bear full responsibility for their provision. These items must correspond exactly to those listed in the submitted offer.	X	X	X	X
6.5	The SUPPLIER, upon completion of the final installation and testing of the equipment, shall provide the beneficiary institution with a complete set of commissioning forms. These forms shall include the full record of all tests performed, the corresponding results obtained, and any observations or remarks relevant to the commissioning process.		X	X	X
7	Training				
7.1	The SUPPLIER must provide trainings on the use/operation of the equipment and all its accessories to the end-users according to the following scheme: - End-user training duration: Certifying, at least 2 days. - The SUPPLIER shall train the users for the use and daily maintenance of all the equipment and software. - The training course shall be conducted at the location where the equipment is delivered and installed. - The trainers shall be certified experts belonging to the manufacturer and/or the representatives in the country of the SUPPLIER. The training shall be conducted in French and translated into Malagasy, or vice versa, as necessary to ensure full understanding by all participants. - The training course for users shall focus at least on the following topics: - General equipment functions in the offered configuration, alarm signals and error signals			X	X

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
	showing all the possible equipment functionalities; 2. Calibrations, daily cleaning and maintenance operations in order to assure the longest equipment life; 3. Correct equipment utilization and related possible risks for users; 4. Description of all settings, parameters; - The training course essential topics shall be video/audio recorded and an on-line tutorial will be made available for the users. - The users successfully trained will receive a certificate from the manufacturer or authorized distributor.				
	The SUPPLIER must provide trainings on the maintenance of the equipment and all its accessories to the maintenance staff according to the following scheme: Maintenance personnel training: at least 3 days. - The Supplier shall train maintenance technicians made available by the beneficiary on the most frequent problems that could occur during equipment utilization and that are under the maintenance technicians' competencies. - The training course for maintenance technicians shall be held in a single location where the equipment is delivered and installed or in another training center in which the same equipment is available. - The trainers shall be qualified experts belonging to the manufacturer and/or by qualified experts certified by the manufacturer. The training shall be conducted in French and translated into Malagasy, or vice versa, as necessary to ensure full understanding by all participants. - The training course for maintenance technicians shall focus on at least the following topics: 5. Basic Users training 6. Scope of equipment, machine parameters, operating environment 7. Equipment principles of operation			X	X

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
	8. Study of Circuits, block diagrams, circuit diagrams, electronic board, hydraulic circuit, pneumatic circuit 9. Device units, boards and electronics testing 10. Common fault / Failures & Solutions 11. Troubleshooting 12. Routine testing & Preventive maintenance (planning of operations: frequency, tools required, overhaul kit and consumables or supplies). 13. QC norms and standards - Demo - Hand On - Practical session 14. Preventive and corrective maintenance (with fault simulation and correction): case studies. 15. Quality control (control procedures, measuring equipment, standards, etc.). 16. Parts replacements: Disassembly and assembly of machine units. 17. Practical description of machine parameters (operating mode /programs). 18. Describe the machine's technical parameters in practice (technical mode). - The training course shall be video/audio recorded and an on-line tutorial will be made available for the trainees. - The trainee will receive all the software keys that are needed for equipment maintenance.				
7.2	The SUPPLIER must provide trainings on the maintenance of the equipment and all its accessories to the maintenance staff according to the following scheme: - Maintenance personnel training: Certifying, at least 3 days. - The Supplier shall train maintenance technicians made available by the beneficiary on the most frequent problems that could occur during equipment utilization and that are under the maintenance technicians' competencies. - The training course for maintenance technicians shall				x

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
	be held in a single location where the equipment is delivered and installed or in another training center in which the same equipment is available. - The trainers shall be qualified experts belonging to the manufacturer and/or by qualified experts certified by the manufacturer. The training shall be conducted in French and translated into Malagasy, or vice versa, as necessary to ensure full understanding by all participants. - The training course for maintenance technicians shall focus on at least the following topics: - Theoretical session 19. Basic Users training 20. Scope of equipment, machine parameters, operating environment 21. Equipment principles of operation 22. Study of Circuits, block diagrams, circuit diagrams, electronic board, hydraulic circuit, pneumatic circuit 23. Device units, boards and electronics testing 24. Common fault / Failures & Solutions 25. Troubleshooting 26. Routine testing & Preventive maintenance (planning of operations: frequency, tools required, overhaul kit and consumables or supplies). 27. QC norms and standards - Demo - Hand On - Practical session 28. Preventive and corrective maintenance (with fault simulation and correction): case studies. 29. Quality control (control procedures, measuring equipment, standards, etc.). 30. Parts replacements: Disassembly and assembly of machine units. 31. Practical description of machine parameters (operating mode /programs). 32. Describe the machine's technical parameters in practice (technical mode). - The training course shall be video/audio recorded and an on-line tutorial will be made available for the				

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
	trainees. - The trainee will receive all the software keys that are needed for equipment maintenance. The technicians successfully trained will receive a certificate from the manufacturer or authorized distributor.				
7.3	The SUPPLIER shall provide both a written manual and a video tutorial in French, containing clear instructions for end users on the safe use, cleaning, and daily maintenance of the supplied equipment. These materials must be delivered at the time of installation and remain accessible to the end users thereafter.	x	x		
8	Formal/final reception of goods				
8.1	The SUPPLIER shall use its best efforts to ensure that the goods are delivered, installed, and prepared in a manner that enables their formal receipt by the beneficiary institution without delay. The formal acceptance of the goods shall be conducted by the beneficiary institution after UNOPS has verified that all items specified in the Terms of Reference, technical specifications, and contract have been delivered, properly assembled, and that the required training has been completed. A receipt report shall be signed by all parties involved—namely, the SUPPLIER, the beneficiary institution, and UNOPS—to confirm acceptance.	x	x	x	x
8.2	The SUPPLIER shall accept the condition that, in the event any item is found to be defective or if any parts or accessories are missing, final acceptance of the goods shall be suspended until the missing items are delivered or the defective item(s) are replaced with items in perfect condition. The SUPPLIER shall bear all costs associated with the shipment of the missing items to the final destination/place of installation, as well as the cost of returning any defective items to the factory and shipping the replacements from the factory to the final destination.	x	x	x	x
9	Extended Warranty				
9.1	The SUPPLIER shall provide an extended warranty that includes the following: During the one (1) years of the Basic Warranty period, the warranty shall cover additionally the calibration,			x	x

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
	preventive and corrective maintenance, and technical support. An additional one(1) year following the expiration of the Basic Warranty, during which the SUPPLIER shall continue to provide calibration, preventive and corrective maintenance, and technical support under the same terms and conditions. This results in a total warranty period of two (2) years from the date of acceptance by the end-user after testing and commissioning at the location specified in the contract.				
	An additional two (2) years following the expiration of the Basic Warranty, during which the SUPPLIER shall continue to provide calibration, preventive and corrective maintenance, and technical support under the same terms and conditions. This results in a total warranty period of three (3) years from the date of acceptance by the end-user after testing and commissioning at the location specified in the contract.				X
9.2	The SUPPLIER shall include calibration, preventive and corrective maintenance, and technical support, to all provided equipment and systems according to the manufacturers' recommendations.			x	x
9.3	The SUPPLIER shall perform all activities related to calibration, preventive and corrective maintenance, and technical support at the site where the equipment is installed, at no additional cost to the beneficiary institution. This obligation includes, but is not limited to, the provision of calibration kits, preventive maintenance kits, quality control kits, labor, transportation, and accommodation expenses. These services shall be provided throughout the entire warranty period, including both the Basic and Extended Warranty periods.			x	x
9.4	The SUPPLIER shall provide refresher training for end users and maintenance personnel during each preventive maintenance and calibration visit.			x	x

No.	Service details	Applicable criterium ⁽¹⁾			
		0	1	2	3
9.5	The SUPPLIER shall update to the latest version of software released by the manufacturers, including all licences without any additional cost.			×	×
9.6	The SUPPLIER shall provide the beneficiary institution with a report detailing the technical procedures carried out and the results obtained following the corrective and preventive maintenance activities.			×	×
9.7	The SUPPLIER shall provide continuous remote monitoring, technical support, and primary maintenance services for all installed systems and equipment.				×
9.8	The SUPPLIER shall guarantee a minimum uptime of 95% for each piece of equipment over the course of a calendar year during the warranty period (i.e., at least 347 days of operational availability out of 365), excluding interruptions caused by scheduled maintenance or external factors beyond the system's control. If the total downtime exceeds the allowed 5% [equivalent to eighteen (18) days per calendar year], the SUPPLIER shall extend the warranty period by ten (10) calendar days for each day of downtime beyond the permitted threshold.				×
9.9	The SUPPLIER shall extend the total warranty period proportionally for each additional day of delay, if during warranty the equipment malfunction cannot be resolved remotely and on-site intervention occurs more than five (5) business days after the initial report of the issue.				×
9.10	The SUPPLIER shall provide the beneficiary institution with loan equipment, free of charge, if the malfunctioning equipment is not repaired within thirty (30) calendar days, for use until the original equipment is fully restored.				×

⁽¹⁾ Application of each criterium according to the classification levels of Table 1.

Section II.4. Delivery schedule

The BIDDER shall provide the manufacturing and delivery schedule as per indicated below and the estimated Delivery time of the equipment to the indicated location in Madagascar.

UNOPS Requirements	
Delivery place & Incoterms rules	Incoterms: DPU (Delivered at Place Unloaded), Incoterms 2020. Delivery time: The delivery period shall not exceed twelve (12) weeks from the date of contract signature. Delivery location: Level 0 : Antananarivo Level 1-3 : Refer to the delivery documents International BIDDERS must include in the quoted price, all costs associated with exportation procedures, where applicable. This includes, but is not limited to, freight forwarding, offloading of goods, delivery to the final destination, and any other related expenses.
	UNOPS Right to vary requirements At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed $\pm 20\%$, without any change in the unit prices or other terms and conditions of the ITB