Learning a Decoder-Free Depth Encoder for Fixed-Focus Fluorescence Microscopy

Anonymous CVPR submission

Abstract

Optical encoders learned jointly with a reconstruction network are optimised for that network, motivating designs whose gains extend across reconstruction methods. We demonstrate that optimising depth discrimination yields an optical encoder that improves recovery across the tested reconstruction methods. An angle-scanned plane wave passes through a fixed, defocused phase plate, generating depth-dependent excitation patterns while the specimen and detection focus remain stationary. The plate is trained through the optical forward model to minimise weighted squared correlations between detector-integrated depth signatures, without specimen examples or a reconstruction solver. Evaluation uses total variation regularisation and plug-and-play reconstruction with either pretrained or neuron-trained denoisers. At eight-fold measurement compression and equal total excitation dose, the learned plate improves reconstruction of simulated mouse pyramidal neuron volumes relative to its random initialisation. Mean local depth error decreases from 0.26 *to* $0.15\,\mu\mathrm{m}$ *with total variation, from* 0.49 *to* $0.22\,\mu\mathrm{m}$ *with pretrained denoisers, and from* 0.20 *to* $0.14\,\mu\mathrm{m}$ *with neuron-trained denoisers. The gains persist with background and read noise and at reduced dose. The proposed objective also yields* 12–$22\,\%$ *lower depth error than plates trained with worst-case coherence and Fisher-information objectives under the same budget. Together, these results demonstrate improved depth recovery across the tested reconstruction methods without coupling optical training to a decoder.*

1. Introduction

Computational imaging enables three-dimensional reconstruction from two-dimensional measurements, but recovery depends on how well those measurements distinguish structures at different depths. When different axial arrangements produce similar observations, reconstruction relies more heavily on prior assumptions [4, 10]. In fluorescence microscopy, neuronal branches that overlap in projection may lie at different depths, making this distinction important for recovering their spatial organisation. We investigate how learning the illumination can improve depth discrimination at the measurement stage.




Figure 1. **One learned plate, three decoders.** Mean local depth error of 60 mouse pyramidal neuron volumes from six cells at eight-fold compression and equal excitation dose; each decoder reconstructs the same measurements of each encoder. The plate is trained without a decoder and lowers the depth error of every tested decoder (right: change against the random plate it started from, cell-bootstrap 95 % interval, crops improved). The in-focus DMD stays near the error of a flat estimate with every decoder (DMD and reference lines: the 12 prespecified volumes).

We address this problem by encoding depth in the excitation illumination. A plane wave is scanned in angle through a fixed, defocused phase plate, producing structured illumination patterns that vary with depth (Fig. 2). As the incident angle changes, depth-dependent lateral shifts cause structures at different axial positions to experience different sequences of excitation intensities. A fixed-focus camera records the resulting fluorescence images for computational volume recovery, without axial specimen scanning or detection refocusing.

The effectiveness of this encoding depends on the phase plate. Recent approaches learn optical encoders jointly with reconstruction networks [12, 23, 29], optimising the measurements for a particular decoder. This couples the optical design to a reconstruction method that may change after fabrication. Building on decoder-free optical design [25, 28, 33], we instead optimise the plate through the op-

tical forward model to improve depth discrimination. We minimise weighted squared correlations between detector-integrated measurement signatures of voxels at different depths but the same lateral position. Training requires neither specimen examples nor a reconstruction solver, and we evaluate the resulting encoder across different reconstruction methods.

We evaluate the learned illumination using simulated measurements of mouse pyramidal neuron volumes. Reconstruction uses total variation regularisation and plug-and-play methods with pretrained and neuron-trained denoisers. Across all three methods, the learned phase plate improves mean depth accuracy relative to its random initialisation at the same measurement budget and excitation dose (Fig. 1). We also examine robustness to background and read noise and reduced excitation dose, and compare the training objective with alternative decoder-independent criteria. Our contributions are threefold:

1. **Fixed-focus depth encoding.** We develop an angle-scanned illumination system with a fixed, defocused phase plate and a voxel-integrated vectorial forward model for three-dimensional fluorescence microscopy.
2. **Decoder-free optical optimisation.** We learn the phase plate by minimising weighted squared correlations between detector-integrated depth signatures, without specimen examples or a reconstruction solver.
3. **Evaluation across reconstruction methods.** We demonstrate improved depth recovery with total variation and plug-and-play reconstruction, compare against alternative decoder-free objectives, and assess robustness to noise and reduced excitation dose.

2. Related work

Depth encoding in fluorescence microscopy. Light-field microscopy, DiffuserCam and Miniscope3D encode volumetric information in the detection response [3, 6, 33]. Excitation coding offers a complementary route. Yeh *et al*. [34] reconstruct three-dimensional fluorescence using translated speckle illumination and axial sample scanning, while Pascucci *et al*. [24] exploit axial speckle decorrelation to recover volumes from a two-dimensional scan. Yilmaz *et al*. [35] vary the incident angle through a fixed scattering layer for two-dimensional recovery, and Jiang *et al*. [14] use tilted pattern projection to produce depth-dependent lateral shifts. Our acquisition combines these mechanisms: the incident angle is scanned through a fixed, defocused phase plate, with a stationary specimen and fixed camera focus, and the plate is learned.

Learning optical encoders. End-to-end methods optimise optics jointly with a reconstruction network, for extended depth of field [29], monocular depth [7, 32], 3D localisation [23], coded illumination [17] and compressive fluorescence microscopy [12]. The learned optics are then matched to one decoder. Decoder-free design instead scores the measurements: Shechtman *et al*. [28] maximise Fisher information, Miniscope3D [33] penalises the coherence of depth-dependent point-spread functions, building on coherence minimisation in sensing design [9], and IDEAL [20, 25] optimises an estimate of mutual information. Our objective belongs to this family but is built for an excitation encoder: it decorrelates the detector-integrated signatures that a voxel leaves across scan angles at different depths, and requires no specimen examples. We compare it with coherence and Fisher-information objectives trained under the same budget, and report its effect under several decoders rather than one.

Priors, hallucination and guarantees. Plug-and-play methods insert learned denoisers into iterative solvers [15, 30, 37], and diffusion models now provide generative priors for inverse problems [8, 16, 38]. A prior can produce structure that the forward model does not determine [4, 10, 13], and posterior spread from a learned model is not a calibrated error bound [22]. Distribution-free methods calibrate reconstruction errors on held-out data [1]; conformal risk control [2] bounds the expected value of a monotone loss. We use it to calibrate depth placement, and use set-membership bounds over measurement-consistent volumes [11, 21] to examine what the data determine without the learned prior.

Forward modelling. Incoherent structured-illumination models combine excitation intensity with the emission response [34], and wave-optics models account for diffraction in volumetric imaging [6]. We integrate fine-grid optical responses over reconstruction voxels, with a vectorial high-NA relay [19, 26]. Our reconstruction uses established components, adapting multi-slice fusion [18] to three-dimensional ADMM; we do not propose a new reconstruction algorithm.

3. Angle-scanned phase-encoded 3D imaging

Acquisition. We vary the illumination direction across two angular coordinates while keeping the phase plate fixed (Fig. 2). For transverse position $\mathbf{r} = (x, y)$, the incident plane wave is $u_t^{\mathrm{in}}(\mathbf{r}) = a_t e^{\mathrm{i}k_{\mathrm{ex}}\mathbf{s}_t^\top \mathbf{r}}$, with $\mathbf{s}_t = \cos\theta_t(\cos\alpha_t, \sin\alpha_t)$, where α_t is the azimuth and θ_t parametrises the direction in the vertical plane containing that azimuth. The corresponding forward direction is $(\mathbf{s}_t, \sin\theta_t)$, and $k_{\mathrm{ex}} = 2\pi/\lambda_{\mathrm{ex}}$ for the simulated refractive index of one. Thus, the two coordinates control the orientation and signed magnitude of the transverse phase gradient. The scan contains 256 distinct directions with $0.5 \le \|\mathbf{s}_t\| \le \sqrt{3}/2$. Each incident field passes through the phase plate, $\tau(\mathbf{r}) = e^{\mathrm{i}\phi(\mathbf{r})}$, and propagates a distance D to the intermediate illumination image plane by angular-spectrum propaga-




Figure 2. **Fixed-focus acquisition and reconstruction.** Left: a tilted plane wave illuminates a fixed phase plate placed 6.6 μm (specimen-space units) before an intermediate illumination image plane; an NA-0.8 vectorial relay maps the field into the specimen; a camera at a fixed detection focus records one image per illumination angle. Only the input angle changes: plate, specimen and focus are stationary, and nothing is scanned axially. Blue: excitation; orange: fluorescence. Right: one of the three decoders of Sec. 4, multi-slice plug-and-play recovery, alternates a conjugate-gradient data-consistency step with a 2D denoiser applied to the xy, yz and zx slices of a shared volume and an ℓ_1 shrinkage, fused by ADMM.

tion. The NA-0.8 objective pupil filters the modulated field before a vectorial Richards–Wolf relay [19, 26] produces the specimen excitation $I_t = \sum_{c\in\{x,y,z\}} |E_{t,c}|^2$, assuming x-polarised illumination. Coordinates and distances are expressed in specimen-equivalent units. The camera records one fluorescence image per direction at a fixed detection focus. Acquisition is sequential, not single-shot.

Why the angle encodes depth. Propagation converts the plate's phase structure into spatially varying excitation. Scanning the incident angle shifts this pattern differently at different depths, giving each location a depth-dependent illumination sequence. An ideal thin-screen model illustrates the mechanism: without pupil clipping, a change Δs of the transverse direction cosine displaces the pattern at propagation distance d from the plate by $d\,\Delta s/\sqrt{1-\bar{s}^2}$ (paraxially $d\,\Delta s$), where $\bar{s}$ is the mean direction cosine and $d = D$ at the conjugate plane of focus. Different depths therefore translate at different rates, and plate defocus places the shear pivot outside the specimen. This relation provides intuition; the high-NA vectorial model includes the finite pupil and departures from pure translation, and Sec. 6.1 measures the actual shear.

Image formation. Let $X_0(\mathbf{r}, z)$ denote the specimen's fluorescence response per unit excitation. Under linear, unsaturated fluorescence, the emission source is proportional to $X_0 I_t$. For laterally shift-invariant detection, the noiseless camera-plane irradiance is

$$g_t(\mathbf{r}) = \int_\Omega h_{\mathrm{em}}(\mathbf{r}-\mathbf{r}', z_{\mathrm{focal}}-z')X_0(\mathbf{r}', z')I_t(\mathbf{r}', z')\,\mathrm{d}^2\mathbf{r}'\,\mathrm{d}z', \tag{1}$$

where h_{em} is the emission intensity point-spread function and Ω is the specimen volume. Pixel p, occupying region D_p, records the normalised sample $\hat{Y}^0_{t,p} = |D_p|^{-1}\int_{D_p} g_t(\mathbf{r})\,\mathrm{d}^2\mathbf{r}$. Fluorescence contributions therefore add incoherently before pixel averaging. Multiplication by the excitation pattern mixes specimen spatial frequencies before detection blur, so the camera need not resolve the illumination pattern itself. Depth discrimination depends on how these measurements vary across illumination angles.

Voxel-integrated operator. To capture excitation variations within reconstruction voxels, we evaluate the optics on a finer grid and integrate the measurement responses over each voxel. Let B be the fine-grid measurement operator, including emission propagation and pixel averaging, and let U replicate each reconstruction coefficient over its fine-grid cells. Then $\mathsf{A} = \mathsf{B}\mathsf{U}$ preserves the fine-grid response for every specimen represented in this piecewise-constant basis.

Sub-voxel specimen structure remains unrepresented. For a fine-grid specimen $\mathbf{x}_f$, its coarse representation $\mathbf{x}_{\mathrm{gt}} = \mathsf{R}\mathbf{x}_f$, obtained by voxel averaging, satisfies

$$\mathbf{y} = \mathsf{A}\mathbf{x}_{\mathrm{gt}} + \underbrace{\mathsf{B}(\mathbf{x}_f - \mathsf{U}\mathsf{R}\mathbf{x}_f)}_{\boldsymbol{\delta}_{\mathrm{disc}}} + \boldsymbol{\varepsilon}. \tag{2}$$

Thus, integration retains within-voxel optical variation, while $\boldsymbol{\delta}_{\mathrm{disc}}$ captures specimen approximation error.

Linear, incoherent fluorescence enables efficient assembly:

$$\mathsf{A}_{(t,p),j} = \sum_{i\in\mathcal{P}_j} I_t(\mathbf{r}_i, z_i)\, g_{p,i}, \tag{3}$$

where $\mathcal{P}_j$ contains voxel j's fine-grid cells, $(\mathbf{r}_i, z_i)$ is the location of cell i, and $g_{p,i}$ is its emission footprint at pixel

p, including its integration weight. These footprints are computed once and reused across angles. After precomputation, direct assembly costs $O(TPF)$, compared with $O(TNF \log F)$ for constructing columns through separate FFT-based forward simulations, with T patterns, P detector samples, F fine-grid cells and N reconstruction voxels.

Learning the plate. We optimise the plate phase ϕ to reduce correlations between measurement signatures at different depths, with the aim of improving depth discrimination. For a voxel at lateral position $\mathbf{r}$ and depth z, let $\mathbf{a}_{\mathbf{r},z}$ denote its centred, unit-norm operator column (Eq. (3)), collecting detector responses across illumination angles. We minimise

$$\mathcal{L}(\phi) = \mathbb{E}_{\mathbf{r}} \sum_{z \neq z'} w_{zz'} \left(\mathbf{a}_{\mathbf{r},z}^{\top} \mathbf{a}_{\mathbf{r},z'}\right)^2, \tag{4}$$

where pairs of planes mirrored about the detection focus, $(z, 14 - z)$ on the 16-plane grid, receive weight two because they are hardest to separate, and other pairs receive weight one. Squaring penalises correlations of either sign; normalisation prevents uniform amplification of the signatures from reducing the loss.

Training differentiates through the excitation, vectorial relay and detection model, without specimen examples or a reconstruction solver. The objective is a surrogate for depth discrimination, not a guarantee of stable recovery; whether it helps is tested with the decoders of Sec. 4, none of which is used in training. Training details are given in the supplement.

4. Decoders and depth intervals

The plate is trained without any decoder. We then decode its measurements with three reconstruction methods of different kinds. None of them is trained on measurements, so the same prior serves every encoder; only scalar weights are selected per encoder and compression, on validation volumes. We call them TV, pretrained PnP and learned PnP.

TV. Box-constrained least squares with isotropic 3D total variation [27], solved by ADMM [5]: a non-learned regularity prior.

Multi-slice plug-and-play. We combine least-squares data fitting with multi-slice denoising and sparsity through plug-and-play ADMM [18, 30]. For orientation $o \in \{xy, yz, zx\}$, P_o rearranges the shared volume into slices, which a 2D denoiser $\mathcal{D}$ processes independently, giving the volume denoiser $\mathcal{D}_o = \mathsf{P}_o^{-1} \mathcal{D} \mathsf{P}_o$. With branches $\mathcal{J}$, each iteration solves

$$\mathbf{x}^{k+1} = \arg\min_{\mathbf{x}} \tfrac{1}{2} \|\mathsf{A}\mathbf{x} - \mathbf{y}\|_2^2 + \sum_{j \in \mathcal{J}} \tfrac{\rho_j}{2} \|\mathbf{x} - \mathbf{v}_j^k + \mathbf{d}_j^k\|_2^2 \tag{5}$$

by conjugate gradients, followed by $\mathbf{v}_j^{k+1} = \mathcal{D}_j(\mathbf{x}^{k+1} + \mathbf{d}_j^k)$ for each denoising branch, soft thresholding for the sparsity branch, and dual updates $\mathbf{d}_j^{k+1} = \mathbf{d}_j^k + \mathbf{x}^{k+1} - \mathbf{v}_j^{k+1}$. *Pretrained PnP* uses three slice branches with a DnCNN [36] trained on natural images, plus the sparsity branch. *Learned PnP* replaces that DnCNN by a 2D network trained on neuron volumes in the three slice branches and adds a fourth branch with a 3D network trained on the same volumes [15, 37]; both are blind Gaussian denoisers. Iteration counts and the selection grid are fixed (supplement).

Conformal depth intervals. For each lateral column c of a reconstruction with signal, let $\hat{z}_c$ be its axial intensity centroid and z_c the reference one. Around each column we place the interval $\hat{z}_c \pm \lambda$, and define the loss $L(\lambda)$ of a volume as the signal-weighted fraction of its columns whose reference centroid falls outside. L is non-increasing in λ and bounded by one, so conformal risk control [2] applies: with mean loss $\hat{R}_n(\lambda)$ over n calibration volumes reconstructed exactly like the test volumes,

$$\hat{\lambda} = \inf \left\{ \lambda : \tfrac{n}{n+1} \hat{R}_n(\lambda) + \tfrac{1}{n+1} \leq \alpha \right\} \tag{6}$$

gives $\mathbb{E}[L(\hat{\lambda})] \leq \alpha$ for a new volume, provided the calibration volumes and the new one are exchangeable. This controls the expected fraction of misplaced signal per volume; it is not a coverage statement for every column, for every specimen, or for connectivity.

Bounds without the learned prior. The data together with intensity bounds define a measurement-consistency set $\mathcal{C}(\mathbf{y}) = \{\mathbf{x} \in [0,1]^N : \ell_i \leq (\mathsf{A}\mathbf{x})_i \leq h_i\}$, where the per-row intervals contain the noise-free measurement with stated joint probability under the noise model. Per-voxel bounds $\min_{\mathcal{C}} x_v$ and $\max_{\mathcal{C}} x_v$ (two linear programs per voxel) show how far any consistent volume can move. As intuition for why the plate objective should narrow them, consider two voxels with unit-norm columns of correlation ρ in an ℓ_2 ball of radius ε: the range of either voxel is $\pm\varepsilon/\sqrt{1-\rho^2}$, from the Schur complement of the 2×2 Gram matrix. This simplified model ignores the coupling of many voxels, the box and the difference between the centred columns of the loss and the raw columns of A; we measure the bounds directly on a reduced field where exact optimisation is feasible.

5. Experimental setup

All experiments are simulations with the optics of Sec. 3: 532 nm excitation and detection, excitation NA 0.8, detection NA 0.1, a phase plate with 0.32 μm features at $D = 6.6\,\mu$m, 256 directions, 0.08 μm optical and 0.32 μm analysis grids, 0.64 μm reconstruction voxels, and linear fluorescence in a homogeneous non-scattering medium.

Volumes and compression. The reconstruction grid is $16 \times 48 \times 48$ ($N = 36{,}864$, $30.7 \times 30.7 \times 10.2\,\mu\mathrm{m}^3$) with 12×12 camera samples per image, so $M = 144k$ for k directions and the compression is $C = N/M \in \{1, 4, 8\}$ for $k = 256, 64, 32$ (five random direction subsets for $C > 1$). Reference volumes are two-photon stacks of mouse pyramidal neurons from the public DEEP^2 data [31], min–max normalised and resampled to the analysis grid. The test set holds two prespecified crops from each of six cells of three animals, chosen by a structure score; for the decoder comparison (Sec. 6.2) we added eight crops per cell drawn at random from the eligible ones, after all weights had been selected (60 crops). Other analyses use the 12 prespecified crops. The neuron-trained denoisers are trained on 459 crops of a fourth animal; crops of a held-out stack of that animal, simulated like the test crops, calibrate the conformal intervals (35 non-overlapping crops; all 240 overlapping crops as a sensitivity check). Scalar weights of every decoder are selected per encoder and compression on two vascular crops of a separate stack, on grids that extend automatically at their edges. No test volume is used for training, tuning or calibration.

Encoders and photon budget. The random plate (three realisations), a defocused binary amplitude mask at the plate's position, the same scan without a plate, and an in-focus DMD with 256 random binary patterns share relay, pupil, detector and assembly. Their excitation is simulated on an optical grid larger than the field, so beam edges lie outside it, and normalised to the same deposited dose. Counts are $\mathrm{Poisson}(\kappa_0\, 256\, y_i/k)$: a fixed total budget shared by the k frames, with κ_0 set per specimen so that the random plate reaches 15 dB at $k = 256$ (1000 photons per pixel per frame at the mean signal). The learned plate starts from random realisation 0 and changes only inside the specimen field; it is trained with two seeds. On one RTX 4500 Ada GPU, 1000 training steps take 20 min, assembling the plate's operator 6 min, and one reconstruction at $C = 8$ 4–6 s (PnP) or 17 s (TV).

Metrics and statistics. Besides PSNR we report foreground SSIM and the *local depth error*: the mean absolute difference between the true and reconstructed axial centroids of each lateral column, weighted by the true column brightness over columns above 10 % of the brightest. Two references calibrate it: a depth-blind oracle with the exact lateral structure and each column spread over the crop's mean axial profile scores $1.32\,\mu\mathrm{m}$, and a flat estimate, uniform in depth, $2.15\,\mu\mathrm{m}$. Crops of one cell are not independent, so intervals resample cells (six cells, two crops each, 10^4 resamples) and encoders are paired on the same objects, direction subsets, dose and noise protocol; decoders of one encoder share its measurements.

	random plate		learned plate		depth change vs random		DMD
decoder	PSNR	depth	PSNR	depth	seed 0 [95 % CI]	seed 1	depth†
TV	36.70	0.259	38.29	0.150	−42 % [−46, −38]	−43 %	2.31
pretrained PnP	35.72	0.490	37.56	0.224	−54 % [−57, −51]	−55 %	2.17
learned PnP	38.01	0.197	39.60	0.139	−30 % [−32, −27]	−31 %	2.25

Table 1. **The learned plate under three decoders** ($C = 8$, mean over 60 test crops from six cells × 5 measurement draws, equal dose; PSNR in dB, local depth error in μm). Plates are paired on the same objects, direction subsets, dose and noise protocol; depth change with cell-bootstrap 95 % interval. †DMD on the 12 prespecified crops (flat estimate $2.15\,\mu\mathrm{m}$).

Reduced field. Exact bounds over $\mathcal{C}(\mathbf{y})$ use dense linear programs, so they run on a reduced field of the same optics ($4 \times 12 \times 12$ voxels, 3×3 pixels, $k = 32$) with Bonferroni intervals at $\alpha = 0.05$ and Gaussian noise at 15 dB, on 24 synthetic phantoms with three noise seeds and on 48 windows cut from the test neurons (supplement).

6. Results

6.1. Defocus under the angle scan encodes depth

We first test the untrained mechanism. Figure 3 compares the random plate with three encoders that share its relay, detector, dose, volumes, decoders and tuning protocol. A robust line fit of the cross-correlation shift per depth gives shear slopes of 0.74–0.94 of the thin-screen prediction, with the pivot near the plate's conjugate plane. Per-pixel excitation sequences decorrelate within about $1\,\mu\mathrm{m}$ of depth, and point emitters are assigned a wrong depth in 1.5–1.8 % of cases (three plates) against 4.4–5.3 % (three defocused masks), 25 % (DMD, half at the mirror depth) and 89 % without the plate. The in-focus DMD stays at the depth error of a flat estimate (2.17 against $2.15\,\mu\mathrm{m}$) while the plate reaches $0.36\,\mu\mathrm{m}$ (pretrained PnP, $C = 1$). The defocused amplitude mask comes close to the plate: over three realisations of each, the plate leads in PSNR in all nine pairings ($+0.4$ to $+1.2$ dB at $C = 8$) but not in depth error. Depth is therefore encoded by defocus under the angle scan, and the random plate is already a strong baseline; the question is how much learning adds.

6.2. The learned plate under three decoders

Figure 1 and Tab. 1 decode the learned plate's measurements with each method of Sec. 4. On the 12 prespecified crops, the learned plate lowers the depth error at $C = 8$ by 37 % with TV, 47 % with pretrained PnP and 26 % with learned PnP (supplement). Because these crops were chosen by a structure score, we then drew 48 more crops at random from the same six cells and reconstructed them with the weights already selected. On these new crops the reductions are 43, 56 and 31 %, and depth improves on every crop under every decoder. Over all 60 crops the reductions are




Figure 3. **Defocus under the angle scan encodes depth; in-focus patterns and a plain scan do not** (full-size field, equal deposited excitation dose, overfilled illumination aperture; phase plate: the random plate). (a) One excitation pattern per encoder, xz section through the specimen depth ($\pm 5\,\mu$m, focus dotted), each tile scaled to its own 99.5th percentile. (b) Lateral shift of the plate's excitation between two scan angles at every depth, from the cross-correlation peak, for plate defocus $D = 3.3$, 6.6 and $13.2\,\mu$m; lines: thin-screen prediction of Sec. 3. (c) Point emitters assigned a wrong depth by maximum likelihood with unknown brightness at $k = 32$ (dots: three plate and three mask realisations; hatched: mirror depth). (d, e) Reconstructions of the 12 prespecified neuron crops from 6 cells: local depth error and foreground SSIM versus compression, pretrained PnP (solid, no plate dashed; plate and mask at $C = 4, 8$: mean and range over three realisations) and TV (dashed); the DMD's depth error of $2.17\,\mu$m is off scale. No plate: simulated aperture four times the field in (a, d, e), twice in (c). All values are tabulated in the supplement.

42, 54 and 30 % with intervals that exclude zero, PSNR rises by 1.6–1.8 dB, and a second training seed gives 31–55 %. On the prespecified crops the gain holds at every compression we tested: -25 to -37 % with TV, -45 to -48 % with pretrained PnP and -26 to -32 % with learned PnP at $C = 1$, 4 and 8 (supplement).

Fewer frames. On the prespecified crops and at the same total dose, the learned plate with 32 frames is better than the random plate with all 256 frames under pretrained PnP (-35 % [-42, -28]) and has a lower mean under TV (0.153 against 0.187 μm). Under learned PnP the means are similar (0.164 against 0.161 μm), but the interval [-17, $+15$] % is too wide to establish equivalence.

Closer performance across decoders. Over the 60 crops, the mean depth errors of TV and the two PnP decoders span 0.20–0.49 μm with the random plate and 0.14–0.22 μm with the learned plate. This is closer average performance, not a measure of agreement between the reconstructed volumes, which we have not tested.

6.3. Choice of plate objective

Table 2 trains the plate with three alternative decoder-free objectives under the same start, sampling and 1000 steps: a smooth worst-case coherence of the depth columns, following Miniscope3D [33]; a Poisson Fisher-information objective on depth with unknown brightness, following [28]; and our loss without centring. All improve on the random plate. Coherence and Fisher information keep 50–63 % of our depth gain over it: paired on the same measurements, our objective lowers their depth error by 12 and 14 % at $C = 8$ (intervals exclude zero; 11 and 10 of 12 crops) and by 19 and 14 % at $C = 1$. With a second training seed and under all three decoders, ours wins all 12 prespecified comparisons at $C = 8$ (12–22 %, intervals exclude zero; supplement). Centring made no difference. This is a controlled comparison of our implementations of these objectives, not a ranking of the objectives in general or of the imaging systems they come from; none was tuned beyond the shared budget.

	$C = 1$		$C = 8$		ours vs this, $C = 8$
plate objective	PSNR	depth	PSNR	depth	depth [95 % CI]
none (random plate)	39.18	0.161	37.11	0.221	-26 % [-35, -18]
worst-case coherence	39.95	0.135	38.20	0.186	-12 % [-16, -9]
depth Fisher information	40.15	0.128	38.15	0.190	-14 % [-19, -10]
depth decorrelation, uncentred	40.86	0.108	38.85	0.164	0 % [-1, $+1$]
depth decorrelation (ours)	40.82	0.109	38.83	0.164	–

Table 2. **Plate objectives** with the same start, sampling, step budget and seed, decoded with learned PnP (12 prespecified test crops; depth in μm). Each plate has its own validation tuning. Last column: depth change of our plate against each row, paired by crop, with cell-bootstrap interval. Second seed and other decoders: supplement.

6.4. Calibrated depth intervals

Conformal intervals. With $\alpha = 0.1$, the half-width $\hat{\lambda}$ of Eq. (6) is 0.47 μm with the learned plate and 0.65 μm with the random plate, 27 % narrower (second seed: 24.8 %; all 240 overlapping crops: 26 %). The theorem guarantees that the expected signal-weighted fraction of misplaced columns is at most 10 % for volumes exchangeable with the calibra-

	conformal depth interval ($\alpha = 0.1$, $C = 8$)			bound width
plate	$\pm\hat{\lambda}$ (μm)	vs random [95 % CI]	test miscov.	reduced field
random	0.646	–	0.060	1
learned, seed 0	0.474	−26.7 % [−30.1, −22.9]	0.075	0.79
learned, seed 1	0.486	−24.8 % [−29.6, −21.6]	0.060	0.79

Table 3. **Depth intervals and bounds.** Conformal half-width of learned-PnP reconstructions, calibrated on 35 non-overlapping held-out crops (calibration-bootstrap 95 % interval), and the realised miscoverage on the 12 prespecified test crops × 5 draws (signal-weighted, target $\alpha = 0.1$). Bound width: mean per-voxel width over $\mathcal{C}(\mathbf{y})$ on the reduced field (15 dB, 72 queries), relative to the random plate.

tion set. Our calibration crops come from one held-out stack of the animal used to train the priors, and the test cells come from three other animals, so exchangeability is an assumption rather than a design property. Empirically the intervals held on the test cells (realised miscoverage 0.06–0.075), which supports, but does not prove, that they transfer across animals.

Bounds without the learned prior. On the reduced field, using only the measurements, the intensity box and the noise model, the per-voxel bounds over $\mathcal{C}(\mathbf{y})$ are 21 % narrower with the learned plate, close to the $0.80\times$ suggested by the two-column intuition of Sec. 4 for the operator's drop in mean depth correlation from 0.868 to 0.761. Over 12.5–25 dB the narrowing equals 1.9 dB of SNR. On 24 sparse synthetic phantoms with three noise draws each, connections certified by the measurements alone rise from 3 to 12 of 48 connected instances at 15 dB; the gain comes from four phantoms and none is lost, but the noise draws of one phantom are not independent, so we do not attach a per-instance significance. On 48 windows cut from the test neurons, no voxel is determined by any plate at 15–45 dB: at eight-fold compression dense neuronal tissue is not determined by the measurements alone. Statements about neuronal structure therefore need a calibrated prior, and we limit ours to depth placement.

6.5. An in-focus encoder leaves decoders without depth

With the in-focus DMD every decoder stays at 2.17–2.37 μm, the error of a flat estimate (Fig. 1 and Tab. 1); on the 48 random crops learned PnP gives 1.81 μm, above the depth-blind oracle (1.42) and near flat (1.93 μm). A second, independent DMD pattern set, run with TV and both PnP decoders, reproduces them to within 0.06 dB and 0.07 μm. Splitting each column's error into its mean over depth (lateral error) and the zero-mean remainder (depth error) shows where the prior's gain goes. At $C = 8$ learned PnP shrinks the random plate's lateral error to $0.48\times$ of pre-trained PnP's and its depth error to $0.71\times$ (median over the prespecified crops; lower on 12 of 12), but for the DMD it shrinks the lateral error to $0.77\times$ and leaves the depth error at $1.04\times$, which is 99 % of the DMD's total. Figure 4 shows the consequence: the DMD's dendrites look sharper in xy, and its xz section remains without depth. A prior could in principle guess depth from population statistics, but these measurements give it nothing to confirm the guess, and none of the tested priors recovered it.

6.6. Calibration and tolerance

A learned plate is useful only if its gain survives fabrication and alignment. With the operator rebuilt from the fabricated plate, rounding the phase to four levels or adding 0.5 rad RMS height error to an 8-level plate keeps a depth reduction of 25.5 and 25.2 %, against 25.9 % for the nominal plate (learned PnP, $C = 8$). A plate 0.5 μm closer or farther than designed keeps 27.8 and 26.4 % after recalibration, and scan-pointing errors up to 0.01 in direction cosine keep 18–22 % without recalibration. Uncalibrated plate-distance and lateral plate errors shift the excitation code almost rigidly (correlation ≥ 0.97 after the offset); they cost both plates 0.15–0.41 μm and remove the learned plate's advantage, so they must be registered, *e.g.* with one bead. Simulating the optics at the reconstruction spacing instead of integrating over voxels raises the forward discrepancy from 1.07 to 25.7 % (supplement).

7. Discussion and limitations

Training without a decoder. The plate is fabricated once, while decoders are replaced as priors improve. A plate optimised end-to-end with one network is matched to that network, and its benefit to another decoder must be tested anew. Our plate was trained without a decoder and improved depth under a regulariser and two plug-and-play decoders. Three methods, two of which share an algorithm, do not establish transfer to every future decoder; they show that the gain is not tied to one. We did not compare with end-to-end design, which asks for the best pair of optics and decoder rather than optics that serve several; our comparison is with decoder-free objectives under the same budget (Tab. 2).

What the gain is and is not. The learned plate improves depth placement, and conformal calibration turns the improvement into narrower intervals with an expected-risk guarantee under exchangeability. It does not make neuronal topology identifiable: at eight-fold compression no plate determines dense neuronal windows from the measurements alone (Sec. 6.4). In supplementary experiments, conformal connectivity calls on these reconstructions missed their nominal error rate, and query-adaptive direction selection did not meet its prespecified criterion. Connections drawn from such reconstructions remain the prior's choice.




Figure 4. **Same decoder, different encoders** ($C = 8$, prespecified display crop with the largest reference depth spread, draw 0, identical display). Top: lateral maximum-intensity projection, colour coding the depth of the brightest voxel; bottom: xz projection; below: PSNR and local depth error of this crop. Learned PnP sharpens the DMD's lateral structure but not its depth. On this crop the learned plate improves on the random plate only slightly (0.46 against $0.49\ \mu$m); the cohort mean is in Tab. 1. Learned-plate column: reconstruction from the conformal study's rerun, which reproduces the recorded row to 0.02 dB. Bar: $5\ \mu$m.

Limitations. All results are simulations, without hardware validation, of linear fluorescence in a homogeneous non-scattering medium with an ideal relay. Background with read noise, or a tenth of the dose, keeps a 27–55 % gain (supplement); scattering and aberrations will reduce every encoder's depth information. The test set has 60 crops from six cells of three animals, which are not independent; bootstrap intervals over six cells are approximate, although the learned plate lowers the depth error on 59–60 of 60 crops under TV and both PnP decoders. The conformal study and the objective comparison use the 12 prespecified crops. The plate is trained with two seeds and the priors on one animal. The conformal calibration uses one held-out stack of that animal; a different tissue or preparation would need its own calibration. Plate distance and lateral position must be registered before use.

8. Conclusion

A phase plate scanned in angle before an intermediate image plane gives each depth its own excitation sequence, and a specimen-free objective on the resulting measurement signatures improves that encoding without a decoder in the loop. In simulation, the learned plate lowered the depth error of TV and two plug-and-play decoders by 30–54 % and narrowed calibrated depth intervals by 27 %, while an in-focus encoder left every tested decoder without depth. Learned optics can be trained for what the measurements record and then tested across decoders, rather than being matched to one.

LLM Usage Statement

LLMs were used for grammar and fact checking.

References

[1] Anastasios N. Angelopoulos, Amit Pal Kohli, Stephen Bates, Michael Jordan, Jitendra Malik, Thayer Alshaabi, Srigokul Upadhyayula, and Yaniv Romano. Image-to-image regression with distribution-free uncertainty quantification and applications in imaging. In *Proceedings of the 39th International Conference on Machine Learning (ICML), PMLR 162*, pages 717–730, 2022. 2

[2] Anastasios N. Angelopoulos, Stephen Bates, Adam Fisch, Lihua Lei, and Tal Schuster. Conformal risk control. In *International Conference on Learning Representations (ICLR)*, 2024. 2, 4

[3] Nick Antipa, Grace Kuo, Reinhard Heckel, Ben Mildenhall, Emrah Bostan, Ren Ng, and Laura Waller. DiffuserCam: lensless single-exposure 3D imaging. *Optica*, 5(1):1–9, 2018. 2

[4] Sayantan Bhadra, Varun A. Kelkar, Frank J. Brooks, and Mark A. Anastasio. On hallucinations in tomographic image reconstruction. *IEEE Transactions on Medical Imaging*, 40 (11):3249–3260, 2021. 1, 2

[5] Stephen Boyd, Neal Parikh, Eric Chu, Borja Peleato, and Jonathan Eckstein. Distributed optimization and statistical learning via the alternating direction method of multipliers. *Foundations and Trends in Machine Learning*, 3(1):1–122, 2011. 4

[6] Michael Broxton, Logan Grosenick, Samuel Yang, Noy Cohen, Aaron Andalman, Karl Deisseroth, and Marc Levoy. Wave optics theory and 3-D deconvolution for the light field microscope. *Optics Express*, 21(21):25418–25439, 2013. 2

[7] Julie Chang and Gordon Wetzstein. Deep optics for monocular depth estimation and 3D object detection. In *Proceedings of the IEEE/CVF International Conference on Computer Vision (ICCV)*, pages 10193–10202, 2019. 2

[8] Hyungjin Chung, Jeongsol Kim, Michael T. Mccann, Marc L. Klasky, and Jong Chul Ye. Diffusion posterior sampling for

general noisy inverse problems. In *International Conference on Learning Representations (ICLR)*, 2023. 2

[9] Michael Elad. Optimized projections for compressed sensing. *IEEE Trans. Signal Process.*, 55(12):5695–5702, 2007. 2

[10] Nina M. Gottschling, Vegard Antun, Anders C. Hansen, and Ben Adcock. The troublesome kernel: On hallucinations, no free lunches, and the accuracy-stability tradeoff in inverse problems. *SIAM Review*, 67(1):73–104, 2025. 1, 2

[11] Justin P. Haldar. On ambiguity in linear inverse problems: Entrywise bounds on nearly data-consistent solutions and entrywise condition numbers. *IEEE Transactions on Signal Processing*, 71:1083–1092, 2023. 2

[12] Udith Haputhanthri, Andrew Seeber, and Dushan Wadduwage. Differentiable microscopy for content and task aware compressive fluorescence imaging, 2022. 1, 2

[13] David Iagaru, Nina M. Gottschling, Anders C. Hansen, and Josselin Garnier. On hallucinations in inverse problems: Fundamental limits and provable assessment methods. arXiv:2605.13146, 2026. 2

[14] Shaowei Jiang, Jun Liao, Zichao Bian, Pengming Song, Garrett Soler, Kazunori Hoshino, and Guoan Zheng. Axially shifted pattern illumination for macroscale turbidity suppression and virtual volumetric confocal imaging without axial scanning. *Opt. Lett.*, 44(4):811, 2019. 2

[15] Ulugbek S. Kamilov, Charles A. Bouman, Gregery T. Buzzard, and Brendt Wohlberg. Plug-and-play methods for integrating physical and learned models in computational imaging: Theory, algorithms, and applications. *IEEE Signal Processing Magazine*, 40(1):85–97, 2023. 2, 4

[16] Tero Karras, Miika Aittala, Timo Aila, and Samuli Laine. Elucidating the design space of diffusion-based generative models. In *Advances in Neural Information Processing Systems (NeurIPS)*, 2022. 2

[17] Michael R. Kellman, Emrah Bostan, Nicole A. Repina, and Laura Waller. Physics-based learned design: Optimized coded-illumination for quantitative phase imaging. *IEEE Transactions on Computational Imaging*, 5(3):344–353, 2019. 2

[18] Soumendu Majee, Thilo Balke, Craig A. J. Kemp, Gregery T. Buzzard, and Charles A. Bouman. Multi-slice fusion for sparse-view and limited-angle 4D CT reconstruction. *IEEE Transactions on Computational Imaging*, 7:448–462, 2021. 2, 4

[19] A. Marian, F. Charrière, T. Colomb, F. Montfort, J. Kühn, P. Marquet, and C. Depeursinge. On the complex three-dimensional amplitude point spread function of lenses and microscope objectives: theoretical aspects, simulations and measurements by digital holography. *Journal of Microscopy*, 225(2):156–169, 2007. 2, 3

[20] Eric Markley, Henry Pinkard, Leyla Kabuli, Nalini Singh, and Laura Waller. Computationally efficient information-driven optical design with interchanging optimization, 2025. 2

[21] M. Milanese and A. Vicino. Optimal estimation theory for dynamic systems with set membership uncertainty. *Automatica*, 27(6):997–1009, 1991. 2

[22] Dominik Narnhofer, Andreas Habring, Martin Holler, and Thomas Pock. Posterior-variance-based error quantification for inverse problems in imaging. *SIAM Journal on Imaging Sciences*, 17(1):301–333, 2024. 2

[23] Elias Nehme, Daniel Freedman, Racheli Gordon, Boris Ferdman, Lucien E. Weiss, Onit Alalouf, Tal Naor, Reut Orange, Tomer Michaeli, and Yoav Shechtman. DeepSTORM3D: dense 3D localization microscopy and PSF design by deep learning. *Nature Methods*, 17:734–740, 2020. 1, 2

[24] M. Pascucci, S. Ganesan, A. Tripathi, O. Katz, V. Emiliani, and M. Guillon. Compressive three-dimensional super-resolution microscopy with speckle-saturated fluorescence excitation. *Nature Communications*, 10:1327, 2019. 2

[25] Henry Pinkard, Leyla Kabuli, Eric Markley, Tiffany Chien, Jiantao Jiao, and Laura Waller. Information-driven design of imaging systems. In *Advances in Neural Information Processing Systems*, 2025. 1, 2

[26] B. Richards and E. Wolf. Electromagnetic diffraction in optical systems, II. structure of the image field in an aplanatic system. *Proceedings of the Royal Society of London. Series A*, 253(1274):358–379, 1959. 2, 3

[27] Leonid I. Rudin, Stanley Osher, and Emad Fatemi. Nonlinear total variation based noise removal algorithms. *Physica D: Nonlinear Phenomena*, 60(1–4):259–268, 1992. 4

[28] Yoav Shechtman, Steffen J. Sahl, Adam S. Backer, and W. E. Moerner. Optimal point spread function design for 3D imaging. *Physical Review Letters*, 113(13):133902, 2014. 1, 2, 6

[29] Vincent Sitzmann, Steven Diamond, Yifan Peng, Xiong Dun, Stephen Boyd, Wolfgang Heidrich, Felix Heide, and Gordon Wetzstein. End-to-end optimization of optics and image processing for achromatic extended depth of field and super-resolution imaging. *ACM Transactions on Graphics*, 37(4):114, 2018. 1, 2

[30] Singanallur V. Venkatakrishnan, Charles A. Bouman, and Brendt Wohlberg. Plug-and-play priors for model based reconstruction. In *IEEE Global Conference on Signal and Information Processing (GlobalSIP)*, pages 945–948, 2013. 2, 4

[31] Navodini Wijethilake, Mithunjha Anandakumar, Cheng Zheng, Peter T. C. So, Murat Yildirim, and Dushan N. Wadduwage. DEEP-squared: deep learning powered De-scattering with Excitation Patterning. *Light: Science & Applications*, 12:228, 2023. 5

[32] Yicheng Wu, Vivek Boominathan, Huaijin Chen, Aswin Sankaranarayanan, and Ashok Veeraraghavan. PhaseCam3D – learning phase masks for passive single view depth estimation. In *IEEE International Conference on Computational Photography (ICCP)*, 2019. 2

[33] Kyrollos Yanny, Nick Antipa, William Liberti, Sam Dehaeck, Kristina Monakhova, Fanglin Linda Liu, Konlin Shen, Ren Ng, and Laura Waller. Miniscope3D: optimized single-shot miniature 3D fluorescence microscopy. *Light: Science & Applications*, 9:171, 2020. 1, 2, 6

[34] Li-Hao Yeh, Shwetadwip Chowdhury, Nicole A. Repina, and Laura Waller. Speckle-structured illumination for 3D phase and fluorescence computational microscopy. *Biomedical Optics Express*, 10(7):3635–3653, 2019. 2

[35] Hasan Yilmaz, Elbert G. van Putten, Jacopo Bertolotti, Ad Lagendijk, Willem L. Vos, and Allard P. Mosk. Speckle

correlation resolution enhancement of wide-field fluorescence imaging. *Optica*, 2(5):424, 2015. 2

[36] Kai Zhang, Wangmeng Zuo, Yunjin Chen, Deyu Meng, and Lei Zhang. Beyond a Gaussian denoiser: Residual learning of deep CNN for image denoising. *IEEE Transactions on Image Processing*, 26(7):3142–3155, 2017. 4

[37] Kai Zhang, Yawei Li, Wangmeng Zuo, Lei Zhang, Luc Van Gool, and Radu Timofte. Plug-and-play image restoration with deep denoiser prior. *IEEE Transactions on Pattern Analysis and Machine Intelligence*, 44(10):6360–6376, 2022. 2, 4

[38] Yuanzhi Zhu, Kai Zhang, Jingyun Liang, Jiezhang Cao, Bihan Wen, Radu Timofte, and Luc Van Gool. Denoising diffusion models for plug-and-play image restoration. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition Workshops (CVPRW)*, pages 1219–1229, 2023. 2