

Microfluidic Devices in Healthcare: Challenges and Solutions

Name: Souvik Sarkar , Roll no. 22ME10083

Introduction

Microfluidics is an interdisciplinary branch of study, including biology, Engineering, Medical Sciences, Genomics, Physics, chemistry, etc. Over the last two decades, it has evolved from an academic research topic into a powerful platform for point-of-care diagnostics, organ-on-chip systems, drug testing, single-cell analysis, etc. The field was significantly influenced by foundational work from various Researchers, whose reviews outlined the scientific and practical potential of microfluidic systems in chemistry and biology. In this short report I have shared the Current progress of the healthcare devices, areas where it is lagging and how can we overcome this hurdles

Motivation: Though Microfluidics create a huge impact and grows much faster but it is not able to make as large progress as Semiconductor Industry in commercialization. Microfluidic systems are particularly attractive in healthcare because they allow automation, miniaturization, and integration of complex laboratory processes into compact chip-based platforms. While the academic literature is flooded with sophisticated microfluidic devices, there remains a significant transitional gap between research labs and Actual Medical Diagnostics. Most devices are not approved by an organization such as FDA (US). So what are the Clinical Gold standard and Benchmark we are not considering and how to overcome this limitations is main motive of this Report

Progress in Healthcare Applications

Lets look into the current significant progress in creating a huge impact in the Microfluidic Devices in Healthcare Industry. Below are some examples

Significant progress has been observed in: i. Single-cell genomics and proteomics , ii. Cancer circulating tumor cell (CTC) capture, iii. Rapid infectious disease diagnostics, iv. Organ-on-chip drug testing systems, v. Smartphone-integrated biosensing

Gaps in Common Microfluidic Devices

In this section, the working principles and limitations (what issues we are facing with those devices) of most valuable devices have been jot down. This will help to understand

most of the correlation gap between Research Proposal vs Commercialization

Devices(Types)	Principals	Limitations
Capillary Blood Glucose Sensor (BGM)	i.Capillary(Plasma and platelet rich) suction draws blood into a channel where a hydrophilic separation layer filters out RBCs ii.Enzymatic reaction (Glucose Oxidase), triggering a potential difference. The resulting Faradic Current	i. Fluctuation of 0.4-0.6 Hematocrit(HCT) to Venous Blood creates Huge Concentration ii. Squeezing of finger dilutes samples with Interstitial Fluid which causes reduces the Protein leads to systematic underestimation
Continuous Glucose Monitoring(CGM)	Uses Similar Concepts with Micro-needles utilizing a sensor-transmitter system to monitor glucose levels in real-time, operates at the subcutaneous layer below the epidermis	ISF glucose concentration lags 5-20 minutes behind the arterial blood, Unwanted protein molecule creates Bio Fouling which reduce the sensitivity
Hemoglobin Test (Hb)	Capillary Blood into paper strip using Lucas-Washburn equation, after enzymic reaction changes color (Capillary Wicking)	Higher RBC content makes blood "thicker" (higher viscosity), clogging paper pores and causing unpredictable capillary rates
Abbott i-STAT (cTnI) Roche CARDIAC (cobas)	Electrochemical ELISA: Capillary blood enters a cartridge where plasma is separated. An antibody-antigen reaction occurs on an Electrochemical Array, detecting marker levels via Electrochemical Oxidation Fluorescence Immunoassay: Uses a multiplexed approach for Troponin T, Myoglobin, and CK-MB. Markers are labeled with fluorescent tags and measured via optical sensors in a controlled flow	Due to very low concentration White Blood Cells and Platelets shows non-specific binding causes > 50% false signals. Also the HCTs inside the samples creates clogging in channels Due to improper mixing with reagents Chromatic Differentiation often fails, also background Noises affect the process

Pregnancy Test (HCG Test)	Urine moves via capillary action Lateral Flow Immunoassay , conjugate pad containing gold-labeled antibodies for Human Chorionic Gonadotropin	Extremely high hormone levels (Most of the sample evaporated) can saturate antibodies, leading to false negatives
Lab-on-Chip (Integrated)	i. Microfluidic channels are fabricated directly onto or within standard Printed Circuit Board(PCB) using Soft Lithography (PDMS) ii. Built-in heating elements, electrochemical electrodes, and signal processing units to be in direct contact with the fluid.	i. Standard materials (Glass, Silica) can exhibit surface roughness shows biocompatibility issues, difficulty in sealing ii. Expensive Manufacturing and Surface engineering Process for materials such as Polydimethylsiloxane(PDMS)
CTC-Chip (Cancer Diagnostics)	To isolate Circulating Tumor Cells (CTCs) using an array of micro-posts coated with antibodies (e.g., anti-EpCAM) to capture rare tumor cells while allowing billions of RBCs and WBCs to flow through.	Cancer cells are fragile; Noticeable deformation takes place due to high High Shear Stress , generally "White blood cell clogging" reduces the purity of the isolated sample.
Microfluidic ELISA	i. Operates at $Re \ll 1$ absence of turbulence ensures that molecular diffusion is the primary driver for binding ii. Reagents are flowed over immobilized antibodies on the channel surface-Detection is achieved via Fluorescence, Chemiluminescence, or Electrochemical signals	i. Multi-step protocols (adding reagents, washing, and separating) are complex to very complex to automate ii. As the sample is much lesser than required level often gives False Neg error, Unbounded protein clogs the flow path
SARS-CoV-2 (Virus) Test	i. microfluidic chips were developed to perform rapid RT-PCR (Polymerase Chain Reaction) or antigen tests ii. high-surface-area micro-pillars to capture viral particles from swabs, followed by on-chip thermal cycling for RNA amplifications	i. Extremely limited number of RNA copies in early-stage samples, need to amplify genomic content (Operating Cost very High) ii. Often RBC Clogging (45%) with high HCT variation is noticed, Unbounded protein creates Bio-Fouling

Bridging the Gaps : Possible Current and Future Solution

The main objective of the literature review is to find out the limitations of the commonly used devices and find-out possible solutions

1. Materials and Manufacturing gaps

Microfluidic devices mostly have large surface to volume ratio, surface finish operation for creating proper clinical environment. any alternatives of the PDMS

(a) **Surface Engineering Solutions:** To reduce the inherent "waviness" of microchannels and prevent molecular absorption, industry now utilizes UV-Post Printing Hydrophobic Coatings and Omniphobic Lubricant-Infused Surfaces (SLIPS)[10]

(b) **The Thermoplastic Transition:** As the Soft lithography for the PDMS preparation is very expensive compared to these thermoplastic material as it is of no use after curing Polycarbonate (PC): it remains stable in PCR (DNA amplification) at the high temperatures. Easily 3D printed or **embossed** for application ready

PMMA(Polymethyl methyl Acrylic) : Offers superior mechanical strength and optical transparency compared to PDMS, making it the premier cost-effective choice for optical-based diagnostics

2. Low Concentration of the Analytes (Immunoassay)

Microfluidic systems suffer from low detection limits of **Analytes(Nucleic acid / Antigen)** due to small sample volumes and short optical path lengths

(a) **Field Amplified Sample Stacking (FASS):** preconcentrating analytes prior to capillary electrophoresis separation. Putting the sample in Low conductivity than buffer create Strong electric field and Velocity discontinuity at the Conductivity interference (applicable for conductive and charged particles only) [7]

(b) **Isoelectric Focusing:** commonly used for concentrating and separating proteins, peptides, proteins at their respective isoelectric points in the presence of a pH gradient, Analyte stops at Interface where net pH is zero

(c) **Membrane-Based Osmotic Preconcentration:** fluid (water)flux through the semi permeable membrane draw the solution with high osmolarity then removing the water concentrated Analyte retention Fault: Whole blood → Protein adsorption → Flux drops 80% [7]

3. Paper-Based Microfluidics (μ PADs)

While paper-based devices are low-cost, biocompatible they face significant physical barriers.

- (a) **Solving Evaporation and Barrier Breach:** High surface Area of microfluidic system like paper leads more than 30% sample loss via evaporation and Wax or AKD solution leakage some time create big problems, to overcome this Paramfilm Wax laser Melting (91% less penetration error in hydrophobic layer to direct the lateral flow)
- (b) **Use of Fluidic Logic gates:** Implementation of **Tunable-Delay Shunts** (as pioneered by Toley McKenzie). These utilize sugar-based bridges or dissolvable membranes. It works as reagent A is not able to move until B completely dissolves into, helps to eliminate incubation time to bond with probes
- (c) **Gradient based Filtration:** Upstream has large pores compares to bottom ones this **gradual narrowing pores** helps to trap smaller cellular debris at varying depths which reduces clogging enables accurate analysis of high viscous samples.
- (d) **Thread based Microfluidic devices** are used as for liquid transporting channels it does not need any patterned barriers and it can be used in **Twinning** for better mixing

Nowadays 3D-fluidic networks were fabricated using papers by patterning multiple layers of paper with fluidic networks. It is worth noting that creating a 3D fluidic network in paper is much more convenient than in traditionally monolithic microfluidic devices with multiple functional zones and multiplexing tests

4. Capillary blood vs Venous blood

Blood is **Coloidal Suspension** not a solution and Capillary fluid is mixture of Arterial, venous, interstitial,intercellular content and also a function of body temperature, hydration. That's why it differs from person to person locality to locality

- (a) **Solving Plasma Skimming Problem using Arterialization:** Red Blood Cells (RBCs) migrate toward the center of the vessel (the axial zone), while a cell-free "Plasma Layer" forms near the walls known as **Plasma Skimming**.Applying a thermal patch (heat) or localized vasodilatory drugs before taking the sample. This dilates the Arterial walls to increase the fluid flow so HCT concentration increases (used in medical treatment)[5]
- (b) **Mathematical Corelation Factor:** When physical sampling cannot be perfected, Multivariate Calibration is used Kuwa et al. and Ginsberg, formulated algorithms to bridge the gap. It requires information of the raw concentration

of glucose , time period between meal and sample collection. "Correlation of capillary and venous blood glucometry"

Apart from that we should look into those biomarkers where the corelation factors are higher than that of the glucose (50 – 60%)

Discussion

The main barrier preventing widespread adoption of microfluidic devices in healthcare is not lack of innovation, but translation difficulty. Many devices perform exceptionally in controlled laboratory conditions but struggle under real clinical variability.

Future improvements should focus on: i. Whole-blood validation ii. Standardized manufacturing pipelines iii. Robust surface chemistry iv. AI-assisted signal correction v. Integration with digital health infrastructure

Potential Insights for Future Aspects:

1. Electrolyte Monitoring for Chronic Kidney Diseases patients: Peoples with CKD (**Hyperkalemia**) have high K⁺ in their blood due to renal potassium excretion. After kidney dialysis period this electrolyte imbalance can be life- threatening thus need to be closely monitoired. While reading those Journals about the corelation between capillary and venous blood, It was found that generally the **Electrolytes (K,Na etc.) mostly available in Interstitial Fluid.According to there paper it is about 0.95 of the venous blood** which is very close for becoming a great **Biomarkar**[9]

2. **Non Invasive Optical Sensing** will play a huge role in future Because of Deep learning (Image processing techniques). A start up 'Luminovah' is working on Neo Natal jaundice curing using optometry. where specific frequency Photons hits the bilirubin molecules (which is insoluable in water) in skin capillaries. After imparting the photons the folded protein structure breaks and become water soluble.

Conclusion

Microfluidic devices have the potential to revolutionize healthcare by making diagnostics faster, cheaper, and more accessible. Despite significant progress, challenges remain in material compatibility, biofouling, manufacturing scalability, and regulatory validation.

Solutions lie in multidisciplinary collaboration between engineers, clinicians, material scientists, and regulatory experts. With proper design-for-manufacture strategies and clinical validation (specially corelation with blood) , microfluidic systems can successfully transition from laboratory prototypes to reliable medical devices.

References

1. Relevant discussions on Classroom microfluidics technologies.




Figure 1: Fluidized logic gates using sugar [?]. From: **Tunable-Delay Shunts for Paper Microfluidic Devices** (Bhushan J. Toley & Brittney McKenzie).




Figure 2: A FASS-sweeping combined mechanism: background electrolyte (BGE), low conductivity sample matrix (SM), '+', '-', or neutral. From: Cunlu Zhao & Zhengwei Ge, *Microfluidic Techniques for Analytes Concentration*.




Figure 3: Thread based: (A, B) Twisted pair of an untreated and a plasma-treated cotton thread. Sample mixing zone demonstrating the manipulation of liquid wicking along threads. From: Xu Liu & Wei Shen, *Thread as a Versatile Material for Low-Cost Microfluidic Diagnostics*.




Figure 4: Electrolyte concentration in ISF and Venous Blood, Shuwen Ma, Jiaqi Li, "Microneedle-based interstitial fluid extraction for drug analysis: Advances, challenges, and prospects"

- Whitesides, G. M., "The origins and the future of microfluidics," Nature.
- Daniel T. Chiu, Andrew J. deMello . "Challenges, and Opportunities for Microfluidics in the Chemical and Biological Sciences"
- Sotirios Papamatthaiou, Pavlos Menelaou, Bilal El Achab Oussallam, Despina Moschou. "Recent advances in bio-microsystem integration and Lab-on-PCB technology"
- Sana Sheybanikashani, AJian Zhou, bcand Ian Papautsky, "Blood microfluidics: progress and challenges"
- Ashis Kumar Sen, Debjani Paul and Suman Chakraborty, "Microfluidics and nanofluidics in India– some recent advancements and futuristic perspective"
- Kaimin Wu, Xuliang He , "Recent progress of microfluidic chips in immunoassay"
- Xiaole Maoa, Tony Jun Huanga, "Microfluidic diagnostics for developing world"
- Shuwen Ma , Jiaqi Li Lixia Pei, "Microneedle-based interstitial fluid extraction for drug analysis: Advances, challenges, and prospects"
- Celeste Chidiac, "The microfluidics revolution: manufacturing techniques and innovative applications"