



Alere Afinion 2 – ASTM Connectivity Protocol

No restriction

Doc ID: TP13/80/2094/2103/2722-14642
Master Doc ID: TP13/80/2094/2103/2722-13154
Version no: 1.0
Authorised: 02.09.2016

Alere Technologies AS
Kjelsåsveien 161, P.O. Box 6863 Rodeløkka, NO-0504 Oslo, Norway

1	ABOUT THIS DOCUMENT	2
1.1	Abbreviations.....	2
2	PHYSICAL TRANSMISSION OF MESSAGES.....	3
2.1	ASTM Socket transfer (High Level)	3
2.2	ASTM Socket transfer (Low Level)	3
3	MESSAGE STRUCTURE	4
3.1	ASTM message structure	4
3.2	Patient measurement results to LIS	4
4	SEGMENTS.....	5
4.1	Legend.....	5
4.2	H - Record.....	5
4.3	Acknowledge-Message.....	5
4.4	P - Record	5
4.5	O - Record.....	6
4.6	R - Record.....	6
5	EXAMPLES	7
5.1	Result-Message (ASTM)	7
<i>5.1.1</i>	<i>Example 1.....</i>	<i>7</i>
<i>5.1.2</i>	<i>Example 2.....</i>	<i>7</i>
<i>5.1.3</i>	<i>Example 3.....</i>	<i>7</i>
<i>5.1.4</i>	<i>Example 4.....</i>	<i>7</i>
<i>5.1.5</i>	<i>Example 5.....</i>	<i>7</i>
<i>5.1.6</i>	<i>Example 6.....</i>	<i>8</i>
<i>5.1.7</i>	<i>Example 7.....</i>	<i>8</i>

1 ABOUT THIS DOCUMENT

This document describes the Data Connectivity Protocol (patient and control records) for LIMS connectivity based on the ASTM E1381 ("Standard Specification for Low-Level Protocol to Transfer Messages Between Clinical Laboratory Instruments and Computer Systems") and ASTM E1394 ("Standard Specification for Transferring Information Between Clinical Instruments and Computer Systems") in the Alere Afinion 2 Analyzer. Even though these standards are obsolete by now, they are still used by Alere Afinion 2 Analyzer. It contains a description of the protocol and the data format transmitted from the Alere Afinion 2 Analyzer as well as various communication examples.

1.1 Abbreviations

LIS Laboratory Information System
LIMS Laboratory Information Management System

2 PHYSICAL TRANSMISSION OF MESSAGES

2.1 ASTM Socket transfer (High Level)

Each ASTM message will be transmitted without low level characters. The transmission of the data will be controlled by the TCP/IP socket layers. The ACK of the receiver is to mark, if the LIS could save the data successful.

Sender (Analyzer)	Direction	Receiver (LIS)
ASTM-result-message	-->	
	<--	<ACK>

wherein:

<ACK> ... ASCII 0x06

2.2 ASTM Socket transfer (Low Level)

The serial interface is based on the ASTM 1381 – 95 "Standard Specification for Low-Level Protocol to Transfer Messages Between Clinical Laboratory Instruments and Computer Systems".

Sender (Analyzer)	Direction	Receiver (LIS)	Description
<ENQ>	-->		Notifying receiver that there is information to send.
	<--	<ACK>	LIS sends back an ACK, if it is ready.
<STX>1H Message<ETX><CS><CR><LF>	-->		Header record
	<--	<ACK>	
<STX>2P Message<ETX><CS><CR><LF>	-->		Patient record
	<--	<ACK>	
<STX>3O Message<ETX><CS><CR><LF>	-->		Order record
	<--	<ACK>	
<STX>4R Message<ETX><CS><CR><LF>	-->		First result record
	<--	<ACK>	
<STX>5R Message<ETX><CS><CR><LF>	-->		Second result record
	<--	<ACK>	
.			
.			
.			
<STX>?L Message<ETX><CS><CR><LF>	-->		Terminator record
	<--	<ACK>	
<EOT>	-->		Message complete transmitted.

If one message will be answered by a <NAK>, this message will be repeated 3 times. After the third <NAK>, the transfer will be finished. Next try to transmit this measurement will start after 10 minutes.

If one message is longer than 240 characters, it is not broken into multiple messages, i.e. <ETB> is never used.

Wherein:

<ENQ> ... ASCII 0x05

<STX> ... ASCII 0x02

<ETX> ... ASCII 0x03

<EOT> ... ASCII 0x04

<ACK> ... ASCII 0x06

<NAK> ... ASCII 0x15

<ETB> ... ASCII 0x17

3 MESSAGE STRUCTURE

3.1 ASTM message structure

The ASTM high level message structure is based on the ASTM 1394-97 "Standard Specification for Transferring Information Between Clinical Instruments and Computer Systems". An overview of the protocol is described in the table below.

Sender (PCC)	Direction	Receiver (LIS)	Description
H Message<CR><LF> P Message<CR><LF> O Message<CR><LF> R Message<CR><LF> R Message<CR><LF> . . R Message<CR><LF> L 1 N<CR><LF>	-->		Record
	<--	<ACK>	

3.2 Patient measurement results to LIS

Message structure:

ASTM segment	Description
H	Message Header
P	Patient Identification
O	Observation Request
R	Observation Result

4 SEGMENTS

4.1 Legend

- Interpretation: additional description
- Req.:
 - F ... fixed value,
 - C ... configured value via user interface,
 - A ... data comes from analyzer,
 - X ... calculated values (e.g. date/time),
 - R ... required from LIS,
 - O ... optional from LIS
- ASTM Field: number of ASTM field
- Source of data: field of the Alere Afinion 2 Analyzer data record, where the data comes from.

4.2 H - Record

Field name	Interpretation	Req.	ASTM Field	Source of data
Field separator	Field separator byte	F	H.2.1	
Encoding characters	Other field separator characters	F	H.2.1	
Sending application	Model name (always "Alere Afinion 2 Analyzer")	F	H.5.1	
Sending facility	DeviceID of measuring device.	A	H.5.3	Serial number of footer
Receiving application	Name of the receiving application / dept. (configurable)	C	H.10.1	Configured value
Processing ID	P ... patient measurement results Q ... quality control results	A	H.12.1	
Version ID	ASTM-version used	F	H.13.1	ASTM: "1"
Date / time of message	Date and time of message creation	X	H.14.1	current time stamp

ASTM-Example: H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||EPR||P|1|20100608185448|

4.3 Acknowledge-Message

ASTM-Example: <ACK>

4.4 P - Record

Field name	Interpretation	Req.	ASTM Field	Source of data
Set ID - Patient ID	PID segment number	F	P.2.1	
Patient Identifier List	(Local) patient ID	A	P.4.1	P- ID of header

ASTM-Example: P|1||43|||U|

4.5 O - Record

Field name	Interpretation	Req.	ASTM Field	Source of data
Set ID - Observation Request	OBR segment number	F	O.2.1	
Filler Order Number		A	O.4.1	RUN# of header
Universal Service ID	Name of assay	A	O.5.4	Name of assay of header
Specimen action code	Constant value	F	O.12.1	N
Specimen Source	Alere Afinion 2 Analyzer ASTM description 0 O // other 1 C // Blood capillary 2 V // Blood venous	A	O.16.2	Assay variant of footer
Charge to practice	Reagent lot	A	O.24.2	LOT# of header
Result Status	Always "F" (final result)	F	O.26.1	

ASTM-Example: O|1||43|^^^CRP|||||N||||^O|||||^10124809||F|

4.6 R - Record

Field name	Interpretation	Req.	ASTM Field	Source of data
Set ID - Observational Simple	R – Segment number	F	R.2.1	
Observation Identifier	Test Device ID	A	R.3.4	Test Name of sub record
Observation Value	Measurement value	A	R.4.1	Result of sub record
Units	Unit	A	R.5.1	Unit of sub record
Abnormal Flags	Flags are not generally standardized. The recommendation is: <: less than measurement lower limit >: higher than measurement upper limit L: less than normal range H: higher than normal range LL: less than extreme range HH: higher than extreme range I: result ambiguous	A	R.7.1	Only '<' and '>' will be supported by the Alere Afinion 2 Analyzer. The Observation Value shall not be presented if abnormal flag is present. See precaution below for ACR and Lipid Panel.
Observation result status	Always "F" (final result)	F	R.9.1	
Responsible Observer	Operator ID of the user, which the measurement has done.	A	R.11.1	operator ID of footer
Date/Time of Analysis	Measurement time	A	R.13.1	Date/Time of header

ASTM-Example: R|1|^ ^^CRP|16|mg/L||||F||||20100608142352|

Precaution for ACR and Lipid Panel:

The ACR observation value shall be interpreted as not valid if albumin and/or creatinine are below or above measurement limits.

The values for LDL, Non-HDL and Chol/HDL shall be interpreted as not valid if either of Chol, HDL or Trig are below or above measurement limits.

An invalid result is indicated with either a comparator flag (> or <) passed along with the value, or "--- " in the observation value field, indicating that the calculation of the value was not possible. See examples 5.1.3, 5.1.4, 5.1.6 and 5.1.7.

5 EXAMPLES

5.1 Result-Message (ASTM)

5.1.1 Example 1

```
H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||||EPR|P|1|20100608185448|
P|1|43|||U|
O|1|43|^^^CRP|||||N|||^O|||||^10124809|F|
R|1|^^^CRP|16|mg/L|||F|||20100608142352|
L|1|N
```

5.1.2 Example 2

```
H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||||P|1|20120329111326|
P|1|ADCC PATIENT STX|||U|
O|1|1|^^^ACR|||||N|||^O|||||^10142193|F|
R|1|^^^ACR|5.6|mg/g|||F|||20100608140517|
R|2|^^^Alb|8.0|mg/L|||F|||20100608140517|
R|3|^^^Creat|17.4|mg/dL|||F|||20100608140517|
L|1|N
```

5.1.3 Example 3

Invalid ACR value indicated with "<" flag.

```
H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||||P|1|20120329111326|
P|1|2|||U|
O|1|2|^^^ACR|||||N|||^O|||||^10142193|F|
R|1|^^^ACR|<5.6|mg/g|||F|||20100608140536|
R|2|^^^Alb|<5.0|mg/L|||F|||20100608140536|
R|3|^^^Creat|33.0|mg/dL|||F|||20100608140536|
L|1|N
```

5.1.4 Example 4

Invalid ACR value indicated with "---" in the observation value field.

```
H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||||P|1|20120329111326|
P|1|7|||U|
O|1|7|^^^ACR|||||N|||^O|||||^10142193|F|
R|1|^^^ACR|---|mg/g|||F|||20100608140626|
R|2|^^^Alb|<5.0|mg/L|||F|||20100608140626|
R|3|^^^Creat|<16.4|mg/dL|||F|||20100608140626|
L|1|N
```

5.1.5 Example 5

```
H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||||P|1|20120329111440|
P|1|123|||U|
O|1|47|^^^Lipid Panel|||||N|||^O|||||^10162193|F|
R|1|^^^Chol|2.60|mmol/L|||F|||20120222143007|
R|2|^^^LDL|1.39|mmol/L|||F|||20120222143007|
R|3|^^^HDL|0.80|mmol/L|||F|||20120222143007|
R|4|^^^Trig|0.90|mmol/L|||F|||20120222143007|
R|5|^^^non-HDL|1.80|mmol/L|||F|||20120222143007|
R|6|^^^Chol/HDL|3.3|||F|||20120222143007|
L|1|N
```

5.1.6 Example 6

Invalid values indicated with "<" flag.

```
H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||||P|1|20120329111456|
P|1||123|||||U|
O|1||54|^^^Lipid Panel|||||N||||^O||||||^10162193||F|
R|1|^^^Chol|<2.59|mmol/L||<|F|||20120222143142|
R|2|^^^LDL|<0.16|mmol/L||<|F|||20120222143142|
R|3|^^^HDL|0.79|mmol/L|||F|||20120222143142|
R|4|^^^Trig|3.62|mmol/L|||F|||20120222143142|
R|5|^^^non-HDL|<1.80|mmol/L||<|F|||20120222143142|
R|6|^^^Chol/HDL|<3.3||<|F|||20120222143142|
L|1|N
```

5.1.7 Example 7

Invalid values indicated with "---" flag.

```
H|\^&|||Alere Afinion 2 Analyzer^^AF0000030|||||P|1|20120329111440|
P|1||123|||||U|
O|1||50|^^^Lipid Panel|||||N||||^O||||||^10162193||F|
R|1|^^^Chol|>12.90|mmol/L||>|F|||20120222143045|
R|2|^^^LDL|---|mmol/L|||F|||20120222143045|
R|3|^^^HDL|>2.59|mmol/L||>|F|||20120222143045|
R|4|^^^Trig|3.00|mmol/L|||F|||20120222143045|
R|5|^^^non-HDL|---|mmol/L|||F|||20120222143045|
R|6|^^^Chol/HDL|---||||F|||20120222143045|
L|1|N
```