

PATIENT 1
RESULT 1
RESULT 2
RESULT 3
TERMINATOR

And this message has to be divided into sub-messages, it will be possible to have the following options:

Option Only one header/terminator frame disabled (default)

First message

HEADER
PATIENT 1
RESULT 1
TERMINATOR

Second message

HEADER
RESULT 2
RESULT 3
TERMINATOR

Option Only one header/terminator frame enabled

First message

HEADER
PATIENT 1
RESULT 1

Second message

RESULT 2
RESULT 3
TERMINATOR