


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 040663 Lot #: M4953

Tx Parameters: 9.04MHz, 271mv

* Distance Range: 11.98 - 12.31 mm


* Mean Distance : 12.12 mm

UR ratio: 0.662;

*** Power : 7.50 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.12mm**

