


Normalized Angular Intensity Profile



Test Results:

SW0041 1.2

SN: 040651 Lot #: M4953

Tx Parameters: 9.06MHz, 271mv

* Distance Range: 11.95 - 12.32 mm


* Mean Distance : 12.12 mm

UR ratio: 0.665;

*** Power : 7.07 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.12mm**

