


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 040648 Lot #: M4953

Tx Parameters: 9.09MHz, 271mv

*** Distance Range: 11.93 - 12.37 mm**


*** Mean Distance : 12.13 mm**

UR ratio: 0.678;

*** Power : 7.33 W**

*** : For informational purpose only;
not contribute to UR computation.**

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.13mm**

