


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 040641 Lot #: M4953

Tx Parameters: 9.04MHz, 271mv

* Distance Range: 12.01 - 12.22 mm


* Mean Distance : 12.08 mm

UR ratio: 0.785;

*** Power : 7.42 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.08mm**

