


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 040679 Lot #: M4953

Tx Parameters: 9.06MHz, 271mv

* Distance Range: 11.90 - 12.42 mm


* Mean Distance : 12.16 mm

UR ratio: 0.661;

*** Power : 8.73 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.16mm**

