


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 040069 Lot #: M4933

Tx Parameters: 8.93MHz, 271mv

*** Distance Range: 11.96 - 12.25 mm**


*** Mean Distance : 12.08 mm**

UR ratio: 0.589;

*** Power : 8.39 W**

*** : For informational purpose only;
not contribute to UR computation.**

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.08mm**

