


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: E00229 Lot #: M4908

Tx Parameters: 8.98MHz, 325mv

*** Distance Range: 11.63 - 12.46 mm**


*** Mean Distance : 12.12 mm**

UR ratio: 0.840;

*** Power : 9.08 W**

*** : For informational purpose only;
not contribute to UR computation.**

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.12mm**

