


Normalized Angular Intensity Profile



Test Results:

SW0041 1.2

SN: 040662 Lot #: M4953

Tx Parameters: 9.04MHz, 271mv

* Distance Range: 11.92 - 12.28 mm


* Mean Distance : 12.11 mm

UR ratio: 0.701;

*** Power : 6.89 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.11mm**

