


Normalized Angular Intensity Profile



Test Results:

SW0041 1.2

SN: 040676 Lot #: M4953

Tx Parameters: 9.06MHz, 271mv

* Distance Range: 11.98 - 12.35 mm


* Mean Distance : 12.15 mm

UR ratio: 0.808;

*** Power : 7.04 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.15mm**

