


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 040675 Lot #: M4953

Tx Parameters: 9.12MHz, 271mv

* Distance Range: 11.94 - 12.41 mm


* Mean Distance : 12.16 mm

UR ratio: 0.679;

*** Power : 7.81 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.16mm**

