


Normalized Angular Intensity Profile



Test Results:

SW0041 1.2

SN: 040447USCSB1

Lot #: M4948

Tx Parameters: 9.09MHz, 271mv

* Distance Range: 11.99 - 12.38 mm


* Mean Distance : 12.14 mm

UR ratio: 0.765;

*** Power : 8.40 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.14mm**

