


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 040606 Lot #: M4953

Tx Parameters: 9.06MHz, 271mv

* Distance Range: 11.98 - 12.38 mm


* Mean Distance : 12.13 mm

UR ratio: 0.476;

*** Power : 7.60 W**

* : For informational purpose only;
not contribute to UR computation.

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.13mm**

