


Normalized Angular Intensity Profile



Test Results: SW0041 1.2

SN: 028755 Lot #: M4426

Tx Parameters: 9.08MHz, 325mv

*** Distance Range: 11.75 - 12.59 mm**


*** Mean Distance : 12.08 mm**

UR ratio: 0.631;

*** Power : 6.46 W**

*** : For informational purpose only;
not contribute to UR computation.**

Pressure Map



***Tracked Xdcr Wobbling, mean distance is 12.08mm**

