


五种去噪方法与 FDCT 真值对比

原始 LDCT
PSNR 36.8836 dB | SSIM 0.90915




均值滤波
PSNR 35.1052 dB | SSIM 0.94158



高斯滤波
PSNR 36.1205 dB | SSIM 0.94220



自适应中值滤波
PSNR 36.7682 dB | SSIM 0.92422




NLM
PSNR 39.2518 dB | SSIM 0.95269



小波软阈值
PSNR 37.6683 dB | SSIM 0.93691



FDCT 真值
参考真值



最优去噪
NLM

PSNR 39.2518 dB
SSIM 0.95269

参考：配对 FDCT