NoctiTrace HS-2 Home Sleep Test Kit: Patient Quick-Start and Return Leaflet

Mae Ping Sleep and Breathing Centre, Chiang Mai · Patient leaflet PIL-SLP-07, version 3.1 · Issued 1 August 2026 · Replaces version 3.0 of 14 January 2026

1. About this leaflet and your kit

Your doctor has asked for a home sleep test. You will take a NoctiTrace HS-2 Home Sleep Test Kit home and record your breathing, oxygen level, pulse and body position while you sleep in your own bed. The kit belongs to the Mae Ping Sleep and Breathing Centre and is lent to you for a short, fixed period.

The NoctiTrace HS-2 records only. It does not treat snoring or sleep apnoea and nobody watches the recording while you sleep.

What is in the grey kit case

- **Recorder unit (HS-2):** a small black box with one round button and one status light. The kit number (for example "MP-HS2-0148") is printed on its back.
- **Chest effort belt:** an elastic belt with a clip that holds the recorder on your chest.
- **Nasal cannula:** a thin clear tube with two soft prongs that sit just inside your nostrils. It is single-use and sealed in its own bag; two are supplied, one for each night.
- **Finger oxygen sensor:** a soft clip on a short grey cable, worn on the index finger of your non-dominant hand.
- **USB charging cable** with plug adapter, a **night diary card** and a padded **return pouch**.

2. When your test takes place

Kits are collected in person from the Sleep Test Desk, Building B, ground floor, on Monday to Thursday between 14:00 and 16:30. We do not hand out kits on Fridays, at weekends or on public holidays. At collection, Anuchit Thepkaew, our Kit Loan Coordinator, or one of the desk nurses will fit the kit on you once so you can see how it goes on, and you will sign the loan form.

Your two test nights are the night of the collection day and the night immediately after it. The recorder is programmed to accept a maximum of two recordings; it will not start a third night. If you cannot do the test on the planned nights, do not move the nights yourself: call the Sleep Test Desk before 16:00 on the collection day and we will arrange a new collection date.

3. Preparing for each test night

- Charge the recorder during the afternoon until the status light is steady green. A charge lasts one night only.
- Do not drink alcohol after 18:00 and avoid coffee, tea, energy drinks and cola after 14:00 on both test days.
- Take your usual medicines as normal, including sleeping tablets you take every night. Tell the desk nurse at collection about any new medicine started in the last two weeks.

Fictional document created for testing a question-answering system. All names, figures and organisations are invented.

- Remove nail polish, gel nails or false nails from the index finger that will carry the sensor, because they block the light the sensor uses.

4. Putting the kit on

Fit the kit in your bedroom about 15 minutes before you plan to switch off the light. Follow the steps in this order:

1. Wash and dry your face; do not use face cream or oil around the nose.
2. Fasten the chest effort belt around your chest, just below the armpits, over a thin cotton shirt or directly on the skin. It should be snug but you must be able to slide two fingers under it.
3. Clip the recorder onto the belt in the middle of your chest, with the button facing upwards towards your chin.
4. Open a new nasal cannula bag. Place the prongs in your nostrils, loop the tube over both ears and slide the small bead under your chin until the tube stays in place. Push the end of the tube firmly into the port on the top of the recorder.
5. Put the finger sensor on the index finger of your non-dominant hand, with the cable running along the back of your hand. Plug it into the side socket of the recorder.
6. Press and hold the round button for 3 seconds until the light flashes blue three times. The recording has now started.
7. Write the time you switched off the light on the night diary card, then try to sleep normally.

Note. Use a fresh cannula on the second night. Never wash or reuse a cannula.

5. During the night

While the recorder is working correctly, the light gives one short green flash every 10 seconds. You can sleep in any position. If you wake up to use the toilet, leave the kit on; note on the diary card any time you are awake for more than about 20 minutes.

A night counts as a valid recording when there are at least 4 hours of usable signal from both the nasal cannula and the finger sensor. The sleep physician needs only one valid night to report the result.

If the status light changes

- **Slow amber flash:** a sensor has lost contact. Check that the cannula prongs are in your nostrils and the finger clip is on firmly, then push both plugs in again.
- **Fast amber flash:** the battery is low. Connect the charging cable to a wall socket and keep wearing the kit; the recorder keeps recording while it charges.
- **Steady red light:** the recorder has stopped and cannot be restarted that night. Take the kit off, go back to sleep and write the time on the diary card. Use the second night as planned.

Stopping the recording in the morning

When you get up for the day, press and hold the button for 3 seconds until the light flashes blue twice, then remove the finger sensor, cannula and belt in that order. Write your wake-up time on the diary card.

Fictional document created for testing a question-answering system. All names, figures and organisations are invented.

6. Who may and may not use the kit

The NoctiTrace HS-2 is approved by the Centre for adults aged 18 and over. Children and teenagers are always tested in the sleep laboratory. The kit is also not used for patients who need oxygen at home overnight, or for patients on a home breathing machine (CPAP or similar), unless the sleep physician has written "home test on therapy" on the referral. People with a heart pacemaker may use the kit.

The test must cover your usual main sleep period. If you work night shifts and normally sleep during the day, you may record your daytime sleep instead, provided you tell the Sleep Test Desk when booking and the coordinator marks the loan form "day-sleep recording". Without this set-up a daytime recording cannot be scored reliably and will usually need to be repeated.

7. Returning the kit

Return the complete kit, including the diary card and the charging cable, by 12:00 on the day after your second test night. There are two ways to return it:

- **In person:** on working days between 08:00 and 12:00, bring the case to the Sleep Test Desk counter in Room B214, Building B, second floor (the desk moves upstairs in the mornings, when the ground floor is used for clinics). Before 08:00 or if the desk is closed, use the grey drop box marked "Sleep Kits" beside the main entrance of Building A; it is emptied at 08:30 and 12:00.
- **By courier:** only for patients living outside Chiang Mai province who chose this option on the loan form. Seal the case in the padded return pouch and hand it to the courier who collects it from your address on the morning after the second night. Keep the tracking slip until your deposit is refunded.

Deposit and charges

A refundable deposit of 3,000 baht is paid by card or bank transfer when you collect the kit. Until 30 June 2026 the deposit was 2,000 baht; patients who collected a kit before 1 July 2026 keep the old amount. The deposit is refunded to the same card or account within 7 working days after the kit is checked. The following amounts are deducted when they apply:

- Late return: 300 baht for each working day, or part of a day, after the 12:00 deadline.
- Missing finger oxygen sensor: 1,800 baht.
- Missing chest effort belt: 450 baht. Missing charging cable or adapter: 250 baht.
- Damaged or lost recorder unit: the whole deposit is kept, and the Centre may ask for the remaining repair cost.

8. Your results

The recording is downloaded at the Centre and scored by a sleep technologist, then reported by Dr. Phakhawadee Srisombat or another sleep physician within 10 working days of the kit being returned. Results are given at a follow-up appointment, not over the phone. If neither night reaches 4 hours of usable signal, one repeat test is offered free of charge.

Contacts. Sleep Test Desk (kit loans, bookings, returns): 053 118 460, Monday to Friday 08:00 to 16:30. Urgent kit problems out of hours: text your kit number to 081 774 2093 (answered 19:00 to 22:00).

Fictional document created for testing a question-answering system. All names, figures and organisations are invented.