



Su. Bhaskar Instrument Service

324, Mastermind-I, Royal Palms, Aarey Milk Colony,
Goregaon (E), Mumbai – 400 065. TeleFax: (022) 287 666 09

-:CALIBRATION CERTIFICATE:-

Certificate No. : CC/HD/0112/24-25 ✓

Instrument : HOLIDAY DETECTOR ✓

Voltage Range : 5KV-25KV ✓

Model : Analogue (AN-25)

Serial No. : 201183 ✓

Standard Meter : Fluke Meter [17B+] Sr. No. 48753887WS
Fluke HV Probe [80K-40] Sr. No. 51850035
Traceable to IDEMI, Mumbai (CC/ECL/0803/23-24)

-: RESULTS :-

Reading on Detector	Reading on Standard
05.0 KV	<u>05.3KV</u>
10.0 KV	<u>10.3KV</u>
15.0 KV	<u>15.2KV</u> ✓
20.0 KV	<u>20.1KV</u>
25.0 KV	<u>25.1KV</u>

Accuracy: better than ± 5 % of Full Scale

Date: 28.06.2024 ✓

Calibration Valid up to: 27.06.2025 ✓

Checked By,

Sr. Engineer Quality Control



Su. Bhaskar Instrument Service