UNITED CALIBRATIONS PVT. LTD.

ISO/IEC 17025:2005 Accredited *ISO 9001:2015 Certified

NABL ACCREDITATION
Cert No - CC - 2537

Calibration Certificate

CERTIFICATE NO.	UCPL/2021/H01705/01	CALIBRATION DATE	20.11.2020
		NEXT DUE DATE (As per Customer Requirement)	19.11.2021
		PAGE NO..	1 of 1
RECEIPT ENTRY NO.	UCPL/-IW-H01705	DATE OF RECEIPT	20.11.2020
		RECEIPT CONDITION	Visually Ok
CUSTOMER	M/s. NIHVA TECHNOLOGY PVT. LTD., 187 G BLOCK 1ST FLOOR, THERMAX CHOWK, MIDC CHINCHWAD, PUNE - 411019.		

UUC Details

Description	:	COATING THICKNESS GAUGE	Make	:	Microprocessor
Size/Range	:	0-20-1000 µm	Sr. No.	:	N843381
L.C.	:	1 µm	ID. No.	:	-
Location	:	------	Accuracy	:	-

Reference Used : -
Procedure No. Used : UCPL/CP/MECH/22
Environmental Condition : Temperature: 20 ± 2°C & Humidity: 40 to 60 % Rh

Details of Standards / Equipments used for Calibration

Description	ID. No.	Calibrated By	Certificate No..	Valid Upto
Coating Thickness Foils	UCPL/CTF/01	Shreyash Enterprises	SE.19.12.0002	01.12.2020

Traceability: Traceable to National / International standards through NABL Accredited calibration laboratory.

CALIBRATION RESULTS (All readings are in µm)

Length Measurement		
Standard Value	UUC* Value	Error
11.000	11.600	0.600
25.000	23.800	-1.200
52.000	50.900	-1.100
75.000	75.200	0.200
126.000	124.000	-2.000
255.000	252.000	-3.000
476.000	468.000	-8.000
691.000	682.000	-9.000

*UUC:Unit Under Calibration

Uncertainty of Measurement (at 95.45% of confidence level & coverage factor k=2): ± 0.005 mm

Vikas Kute **CALIBRATED BY (Calibration Engineer)**	P.D.Wadikar **APPROVED BY (Lab Manager)**

. The calibration results reported in this certificate are valid at the time of and under stated condition of the measurements.
. The calibration certificate shall not be reproduced except in full without our prior permission in writing.
. This Certificate relates only to the item Calibrated.

United Calibrations Pvt. Ltd. * Pune - 26 *

UCPL/F/MECH/03-22

Flat No-2, Shyam Enclave, Nr. Priyadarshani school, Indrayani Nagar, Bhosari, Pune, 411026
+91-8411009211/8411009212
Email ID: info@unitedcalibrations.com, Website: www.unitedcalibrations.com

REDMI NOTE 5 PRO
MI DUAL CAMERA

UNITED CALIBRATIONS PVT. LTD.

ISO/IEC 17025:2005 Accredited *ISO 9001:2015 Certified

NABL ACCREDITATION Cert No - CC - 2537

Calibration Certificate

CERTIFICATE NO.	UCPL/2021/H01705/02	CALIBRATION DATE	20.11.2020
		NEXT DUE DATE (As per Customer Requirement)	19.11.2021
		PAGE NO..	1 of 1
RECEIPT ENTRY NO.	UCPL/-IW-H01705	DATE OF RECEIPT	20.11.2020
		RECEIPT CONDITION	Visually Ok
CUSTOMER	M/s. NIHVA TECHNOLOGY PVT. LTD., 187 G BLOCK 1ST FLOOR, THERMAX CHOWK, MIDC CHINCHWAD, PUNE - 411019.		

UUC Details

Description	:	Coating Thickness Foils	Quantity: 5 Nos. Foils	Make	:	-
Size/Range	:	50 to 755 µm		Sr. No.	:	-
L.C.	:	NA		ID. No.	:	-
Location	:	------		Accuracy	:	-

Reference Used : -
Procedure No. Used : UCPL/CP/MECH/14
Environmental Condition : Temperature: 20 ± 2°C & Humidity: 40 to 60 % Rh

Details of Standards / Equipments used for Calibration

Description	ID. No.	Calibrated By	Certificate No..	Valid Upto
Digital External Micrometer	UCPL/DM/01	Shreyash Enterprises	SE.19.09.0189	06.09.2021

Traceability: Traceable to National / International standards through NABL Accredited calibration laboratory.

Calibration Results: All readings in µm

Sr. No.	Nominal Thickness	Departure from Nominal Thickness	Variation in Thickness
1	50.000	52.000	2.000
2	124.000	126.000	2.000
3	288.000	290.000	2.000
4	406.000	407.000	1.000
5	755.000	756.000	1.000

*UUC:Unit Under Calibration

Uncertainty of Measurement (at 95.45% of confidence level & coverage factor k=2): ± 1.8 µm

United Calibrations Pvt. Ltd. PUNE 26

Vikas Kute	P.D.Wadikar
CALIBRATED BY (Calibration Engineer)	**APPROVED BY (Lab Manager)**

- The calibration results reported in this certificate are valid at the time of and under stated condition of the measurements.
- The calibration certificate shall not be reproduced except in full without our prior permission in writing.
- This Certificate relates only to the item Calibrated.

UCPL/F/MECH/03-14

Flat No-2, Shyama Enclave, Nr. Priyadarshani school, Indrayani Nagar, Bhosari, Pune, 411026
Mob: +91- 8411009211/8411009212
Email ID: info@unitedcalibrations.com , Website: www.unitedcalibrations.com

REDMI NOTE 6 PRO
MI DUAL CAMERA