QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date		ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QCL/22-23/ E1379	20/12/2022	CC274722000015119F	19/12/2022	18/12/2023	1 of 1

MRROA No. :QC1299 Date of MRROA 20/12/2022 **Site Request No.** : QC-177 Dated: 19/12/2022

Customer : M/s Nihva Technologies Pvt.Ltd.
187/ G Block,First Floor,Thermax Chowk,Chinchwad Pune 411 019.

Lab/Site : Calibrated at Site.
U.U.C. Details : Digital Voltmeter
Make UT Lamination **Model** : --
Range : 0-99.9VDC **L.C.** : 0.1 VDC
Identification No. : WM-06 **Sl. No.** : --
Accuracy : Not specified by customer **Location** : Tig/Arc Welding M/C
Calibration Method : QS/CPE/01 **Standard Referred** : IS 13875

Test Purpose : To check the error of instrument
Observations : **Pre Calibration Status** : Worth calibrating,
: **Post Calibration Status** : As follows

Electrotechnical Calibration: Direct Current

Sl. No.	Range		Master Instrument	Unit Under Calibration	Error	Expanded Uncertainty
			V	V	V	(±) %
1	99.9	V	20.00	20.0	0.00	0.29
			50.00	49.9	-0.10	0.12
			90.00	89.9	-0.10	0.07

Standards / Equipment used for calibration:			Traceable to National / International standards through unbreakable chain of NABL accredited calibration laboratory.		
Sr. No	Standard / Equipment	Id. No.	Cert. No.	NABL Cert. no.	Validity
1	Universal Calibration System	QS/EM/24	NI/2206/050/001	CC-2294	16/06/2023

Environmental Condition : Temperature 25° C, Relative Humidity 55% RH.
(The reported uncertainty is at coverage factor k=2 corresponds to a confidence level of 95.45%.)

END

Calibrated & Prepared By Mr. Hitesh Chavan
Technical Assistant

Mr. K.G.Bhole
Authorized Signatory

User discretion:
Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.

QRICS PUNE