QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date	ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QCL/22-23/ E1375 20/12/2022	CC274722000015115F	19/12/2022	18/12/2023	1 of 1

MRROA No. :QC1299 Date of MRROA 20/12/2022 **Site Request No.** : QC-177 Dated: 19/12/2022

Customer : M/s Nihva Technologies Pvt.Ltd.
187/ G Block,First Floor,Thermax Chowk,Chinchwad Pune 411 019.

Lab/Site : Calibrated at Site.
U.U.C. Details : Digital Ammeter
Make : UT Lamination
Range : 0-315 ADC/0-75 mVDC
Identification No. : WM-06
Accuracy : Not specified by customer
Calibration Method : QS/CPE/01

Model : --
L.C. : 1 ADC
Sl. No. : --
Location : Tig/Arc Welding M/C
Standard Referred : IS 13875

Test Purpose : To check the error of instrument
Observations : **Pre Calibration Status** : Worth calibrating,
: **Post Calibration Status** : As follows

Electrotechnical Calibration: Direct Current

Sl. No.	Range		Master Instrument	Master Instrument	Unit Under Calibration	Error	Expanded Uncertainty
			ADC	mVDC	ADC	ADC	(±) %
1	315	ADC	63.0	15.00	63	0.0	0.92
			158.0	37.62	159	1.0	0.37
			284.0	67.62	285	1.0	0.26

Standards / Equipment used for calibration: **Traceable to National / International standards through unbreakable chain of NABL accredited calibration laboratory.**

Sr. No	Standard / Equipment	Id. No.	Cert. No.	NABL Cert. no.	Validity
1	Universal Calibration System	QS/EM/24	NI/2206/050/001	CC-2294	16/06/2023

Environmental Condition : Temperature 25° C, Relative Humidity 55 % RH.
(The reported uncertainty is at coverage factor k=2 corresponds to a confidence level of 95.45%.)

END

Calibrated & Prepared By Mr. Hitesh Chavan, Technical Assistant

Mr. K.G.Bhole
Authorized Signatory

User discretion:

Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.

QRICS PUNE