QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date		ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QC/22-23/10137	20/12/2022	CC274722000015134F	20/12/2022	19/12/2023	1 of 2

MRROA No. : QC1296 Date of Receipt 19/12/2022 **Customer DC. No.** : By Hand dt: 19/12/2022

Customer : M/s Nihva Technologies Pvt. Ltd.,
187, G-Block, 1st Floor, Thermax Chowk, Chinchwad, Pune-411019

Calibration Location : In Lab
UUC* Details : Measuring Tape (Trapezoid Type Ring) **Make** : Freemans
Range : 15 Mtrs **L.C.** : 10 mm
Identification No. : NTPL/ST-07 **Sl. No.** : Nil
Id No. taken from : As per instrument
Standard Referred : IS 1269

Test Purpose : To check the error in accordance to the standard (Calibration Method : QS/CPM/66)

Environmental Condition : 20 ± 2° C, Humidity 40 – 60%

Observations **: Pre Calibration Status** : Worth Calibrating,

Standards / Equipment used for calibration:			Traceable to National / International standards through unbreakable chain of NABL accredited calibration laboratory.		
Sr. No	**Standard / Equipment**	**Id. No.**	**Cert. No.**	**NABL Cert. No.**	**Validity**
1.	Tape & Scale Unit	QS/ME/75	SCL/22-23/11/446-001 dt 20/11/2022	CC-2282	19/11/2023

Note: *UUC =Unit Under Calibration

Calibrated By
Ankita Bhuvad
Technical Assistant

QRICS PUNE CALIBRATION

Authorized By
N. M. Shinde
Technical Manager

Regd. Office & Labs : "Paras Complex", GP/67, Office No. 2 & 7 to 10 (Basement), Thermax Chowk, Chinchwad, Pune - 411 019.
Customer Care : 020-27371831, 07276710888 & Accounts : 020-27370204, Suggestion, Feedback & Complaint : 9822069138, E-mail : qrics@hotmail.com, web : www.qrics.in

P.T.O.