QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date		ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QC/22-23/10135	20/12/2022	CC274722000015132F	20/12/2022	19/12/2023	1 of 1

MRROA No. : QC1296 Date of Receipt 19/12/2022 **Customer DC. No.** : By Hand dt: 19/12/2022

Customer : M/s Nihva Technologies Pvt. Ltd.,
187, G-Block, 1st Floor, Thermax Chowk, Chinchwad, Pune-411019

Calibration Location : In Lab
UUC* Details : Steel Tape (L Type)
Range : 5 Mtrs
Identification No. : NTPL/ST-05
Id No. taken from : As per instrument
Standard Referred : IS 1269

Make : Freemans
L.C. : 1 mm
Sl. No. : Nil

Test Purpose : To check the error in accordance to the standard (Calibration Method : QS/CPM/66)

Environmental Condition : 20 ± 2° C, Humidity 40 – 60%

Standards / Equipment used for calibration: Traceable to National / International standards through unbreakable chain of NABL accredited calibration laboratory.

Sr. No	Standard / Equipment	Id. No.	Cert. No.	NABL Cert. No.	Validity
1.	Tape & Scale Unit	QS/ME/75	SCL/22-23/11/446-001 dt 20/11/2022	CC-2282	19/11/2023

Observations : **Pre Calibration Status** : Worth calibrating
: **Post Calibration Status** : As follows reading average value of three readings,

Mechanical Calibration: Dimension (Basic Measuring Instruments, Gauges etc.)

Parameter distance between(in mtr.)	Readings taken at master values (in mm)	Values of UUC (in mm)	Error (in mm)
0-1.0	999.955	1000.300	0.345
0-2.0	1999.955	2000.025	0.070
0-3.0	2999.955	3000.045	0.090
0-4.0	3999.955	3999.920	-0.035
0-5.0	4999.955	4999.830	-0.125

NOTE

Maximum permissible error for total length in steel tape measure as per IS 1269(Part 2) is as below.(where 'L' is in meter)

Class I	Class II	Class III
± (0.1+0.1 x L)	± (0.3+0.2 x L)	± (0.6+0.4 x L)

1. For steel tapes of length upto 5 mtr. Permissible values are as per class I & II only.
2. For steel tapes of length above 5 mtr. Permissible values are as per class I, II and III.
3. For tapes in service the permissible value is equal to TWICE the values indicated in the above table.

Uncertainty of Measurement : ± 980 µm

(The reported uncertainty at above reading is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 corresponds to a confidence level of 95.45 %.)

Note: *UUC =Unit Under Calibration

END

Calibrated By Ankita Bhuvad, Technical Assistant

QRICS PUNE CALIBRATION

Authorized By N. M. Shinde, Technical Manager

User discretion:

Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.

Regd. Office & Labs : "Paras Complex", GP/67, Office No. 2 & 7 to 10 (Basement), Thermax Chowk, Chinchwad, Pune - 411 019.
Customer Care : 020-27371831, 07276710888 & Accounts : 020-27370204, Suggestion, Feedback & Complaint : 9822069138, E-mail : qrics@hotmail.com, web : www.qrics.in

P.T.O.