QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date	ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QCL/22-23/ E1378 20/12/2022	CC274722000015118F	19/12/2022	18/12/2023	1 of 1

MRROA No. :QC1299 Date of MRROA 20/12/2022 **Site Request No.** : QC-177 Dated: 19/12/2022

Customer : M/s Nihva Technologies Pvt.Ltd.
187/ G Block, First Floor,Thermax Chowk,Chinchwad Pune 411 019.

Lab/Site : Calibrated at Site.
U.U.C. Details : Digital Ammeter
Make : Migatronic
Range : 0-350 ADC/0-75 mVDC
Identification No. : WM-02
Accuracy : Not specified by customer
Calibration Method : QS/CPE/01

Model : --
L.C. : 1 ADC
Sl. No. : --
Location : CO2 Welding M/C
Standard Referred : IS 13875

Test Purpose : To check the error of instrument
Observations : **Pre Calibration Status** : Worth calibrating,
: **Post Calibration Status** : As follows

Electrotechnical Calibration: Direct Current

Sl. No.	Range		Master Instrument	Master Instrument	Unit Under Calibration	Error	Expanded Uncertainty
			ADC	mVDC	ADC	ADC	(±) %
1	350	ADC	70.0	15.00	70	0.0	0.92
			175.0	37.50	174	-1.0	0.37
			315.0	67.50	313	-2.0	0.26

Standards / Equipment used for calibration: | **Traceable to National / International standards through unbreakable chain of NABL accredited calibration laboratory.**

Sr. No	Standard / Equipment	Id. No.	Cert. No.	NABL Cert. no.	Validity
1	Universal Calibration System	QS/EM/24	NI/2206/050/001	CC-2294	16/06/2023

Environmental Condition : Temperature 25° C, Relative Humidity 55 % RH.
(The reported uncertainty is at coverage factor k=2 corresponds to a confidence level of 95.45%.)

END

Calibrated & Prepared By Mr. Hitesh Chavan, Technical Assistant

Mr. K.G.Bhole
Authorized Signatory

User discretion:
Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.

QRICS PUNE

Regd. Office & Labs : "Paras Complex", GP/67, Office No. 2 & 7 to 10 (Basement), Thermax Chowk, Chinchwad, Pune - 411 019.
Customer Care : 020-27371831, 07276710888 & Accounts : 020-27370204, Suggestion, Feedback & Complaint : 9822069138, E-mail : qrics@hotmail.com, web : www.qrics.in
Finance : financeqclpl_1989@hotmail.com, Grivenence Feedback & Complaint : grivanceqrics@hotmail.com

P.T.O.