QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date		ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QC/22-23/10132	20/12/2022	CC274722000015130F	20/12/2022	19/12/2023	2 of 2

Observation : **Post Calibration Status** : As follows

Mechanical Calibration: Dimension (Basic Measuring Instruments, Gauges etc.)

The following actual readings are rounded mean value of three readings.

Sl.No	Parameters	Acceptance criteria	Observations
1.0	External measurement		
1.1	0.00 mm		0.00 mm
1.2	21.02 mm	± 20 microns	21.02 mm
1.3	51.02 mm		51.02 mm
1.4	101.02 mm	± 30 microns	101.02 mm
1.5	130.00 mm	± 30 microns	130.00 mm
2.0	Internal measurement		
2.1	20.00 mm		20.00 mm
2.2	50.00 mm	± 20 microns	50.00 mm
2.3	100.00 mm		100.00 mm
2.4	130.00 mm	± 30 microns	130.00 mm
3.0	Error at tip of		
3.1	External Jaw	-	0.00 mm
3.2	Internal Jaw	-	0.00 mm
4.0	Depth Measurement		
4.1	50.00 mm	± 0.3 mm	50.02 mm
4.2	100.00 mm	± 0.3 mm	100.04 mm
4.3	130.00 mm	± 0.5 mm	130.04 mm

Note : 1. The above external and Internal measurements are taken at the middle of the external and Internal Jaws.
2. Only metric scale is calibrated.
3. If the gap is seen on internal jaws it does not affect the measurement.

Uncertainty of Measurement : ± 18.1 µm

(The reported uncertainty at above reading is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 corresponds to a confidence level of 95.45%.)

*** END ***

Calibrated By
N. M. Shinde
Technical Assistant

Authorized By
Ankita Bhuvad
Technical Manager

User discretion:

Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.