QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CALIBRATION CERTIFICATE

Certificate No./Date		Calibration Date	Next Due Date (As per customer)	No. of Pages
QC/22-23/10134	20/12/2022	20/12/2022	19/12/2023	1 of 1

MRROA No. : QC1296 Date of Receipt 19/12/2022 **Customer DC. No.** : By Hand dt: 19/12/2022

Customer : M/s Nihva Technologies Pvt. Ltd.,
187, G-Block, 1st Floor, Thermax Chowk, Chinchwad, Pune-411019

Calibration Location : In Lab
UUC* Details : Level Bottle
Range : 300 mm base
Identification No. : NTPL/LB-01
Id No. taken from : As per instrument
Calibration Procedure : QS/CPM/25

Make : Freemans
L.C. : NA
Sl. No. : Nil

Test Purpose : To check the accuracy and repeatability on the working range.
Environmental Condition : 20 ± 2° C, Humidity 40 – 60%

Standards / Equipment used for calibration:			Traceable to National / International standards through unbreakable chain of NABL accredited calibration laboratory.		
Sr. No.	**Standard / Equipment**	**Id. No.**	**Cert. No.**	**NABL Cert. No.**	**Validity**
1.	Granite Surface Plate	QS/ME/09	QCL/21-22/07134 dt: 08/10/2021	CC-2747	07/10/2023
2.	Spirit level	QS/ME/66	QCL/22-23/04156 dt : 31/07/2022	CC-2747	30/07/2024

Observations : **Pre Calibration Status** : Worth Calibrating
: **Post Calibration Status** : As follows

1) Level bottle set with master spirit level.
2) Level bottle set horizontally with our master spirit level observed – 0° 0' 0"

Uncertainty of Measurement : ± 9 µm/mtr
(The reported uncertainty at above reading is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 corresponds to a confidence level of 95.45%.)

Note: UUC* =Unit Under Calibration

*** END ***

sd
Calibrated By N.M. Shinde
(Technical Assistant)

QRICS PUNE CALIBRATION

Ankita Bhuvad
(Authorized Signatory)

User discretion:
Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.

1. The uncertainty of measurement evaluated and reported in the certificate includes the resolution of the device under calibration.
2. This certificate refers only to the item submitted and the results reported there-in are valid at the time of and under stated conditions of calibration.
3. This certificate shall not be reproduced, except in full, unless written permission for a publication of an approval abstract is obtained from the Proprietor, QRICS CALIBRATION, Pune.
4. The validity mentioned in this certificate is as per client's request.
5. Please review this certificate before use of instrument / equipment.