QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date		ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QC/22-23/10130	20/12/2022	CC274722000015128F	20/12/2022	19/12/2023	1 of 1

MRROA No. : QC1296 Date of Receipt 19/12/2022 **Customer DC. No.** : By Hand dt: 19/12/2022

Customer : M/s Nihva Technologies Pvt. Ltd.,
187, G-Block, 1st Floor, Thermax Chowk, Chinchwad, Pune-411019

Calibration Location : In Lab
UUC* Details : Engineers Square/Right angle **Make** : Tower
Range : 180 x 110 mm **L.C.** : NA
Identification No. : NTPL/RA/01 **Sl. No.** : Nil
Id No. taken from : As per instrument
Calibration Method : QS/CPM/20 **Standard Referred** : IS 2103
Test Purpose : To check the squareness & parallelism on working surface

Environmental Condition : 20 ± 2° C, Humidity 40 – 60%

Standards / Equipment used for calibration:			Traceable to National / International standards through unbreakable chain of NABL accredited calibration laboratory.		
Sr. No.	**Standard / Equipment**	**Id. No.**	**Cert. No.**	**NABL Cert. No.**	**Validity**
1.	Granite Surface Plate	QS/ME/09	QCL/21-22/07134 dt: 08/10/2021	CC-2747	07/10/2023
2.	Plunger Dial (Millimess Type)	QS/ME/14	QCL/21-22/10829 dt: 09/01/2022	CC-2747	08/01/2024
3.	Master Cylinder	QS/MM/42	P/44/210007 dated 08/03/2022	CC-2082	07/03/2025

Observations **: Pre Calibration Status** : Worth Calibrating
: Post Calibration Status : As follows

Mechanical Calibration: Dimension (Basic Measuring Instruments, Gauges etc.)

Sr. No.	Blade Length In mm	Perpendicularity Actual in microns	Parallelity Actual in microns
1.	180	32	61

Note : Parallelity tolerance of blade is three times the value given for perpendicularity tolerance.

Specified Perpendicularity Tolerance as per IS: 2103

Length in mm	Perpendicularity tolerance in µm		
	Grade A	Grade B	Grade C
50	3	-	-
100	3	7	15
150	4	8	18
200	4	9	20
300	5	11	25
500	-	15	35
750	-	20	43

Uncertainty of Measurement : ± 9.5µm (Squareness) & ± 4.9 µm (Parallelism)
(The reported uncertainty at above reading is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 corresponds to a confidence level of 95.45%.)

Note: *UUC=Unit Under Calibration

*** END ***

Calibrated By N.M. Shinde, Technical Assistant

Authorized By Ankita Bhuvad, Technical Manager

User discretion:

Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.

Regd. Office & Labs : "Paras Complex", GP/67, Office No. 2 & 7 to 10 (Basement), Thermax Chowk, Chinchwad, Pune - 411 019.
Customer Care : 020-27371831, 07276710888 & Accounts : 020-27370204, Suggestion, Feedback & Complaint : 9822069138, E-mail : qrics@hotmail.com, web : www.qrics.in

P.T.O.