QRICS CALIBRATION

(Mechanical, Electro-Technical & Thermal)

CC-2747

CALIBRATION CERTIFICATE

Certificate No./Date		ULR No	Calibration Date	Next Due Date (As per customer)	No. of Pages
QC/22-23/10137	20/12/2022	CC274722000015134F	20/12/2022	19/12/2023	2 of 2

Observations : **Post Calibration Status** : Since this tape is not with 'zero away from trapezoid ring' as per standard following readings are taken from the outer face of trapezoid ring

Mechanical Calibration: Dimension (Basic-Measuring Instrument, Gauges etc.)

Parameter distance between outer face of ring (in mtr.)	Readings taken at master values (in mm)	Values of UUC (in mm)	Error (in mm)
0-0.1	99.990	100.225	0.235
0-1.0	999.955	999.870	-0.085
0-2.0	1999.955	1999.930	-0.025
0-3.0	2999.955	2999.735	-0.220
0-4.0	3999.955	3999.645	-0.310
0-5.0	4999.955	4999.470	-0.485
0-6.0	5999.955	5999.205	-0.750
0-7.0	6999.955	6999.005	-0.950
0-8.0	7999.955	7998.675	-1.280
0-9.0	8999.955	8998.845	-1.110
0-10.0	9999.955	9998.585	-1.370
0-11.0	10999.955	10998.285	-1.670
0-12.0	11999.955	11998.260	-1.695
0-13.0	12999.955	12998.000	-1.955
0-14.0	13999.955	13997.620	-2.335
0-15.0	14999.955	14997.480	-2.475

NOTE

Maximum permissible error for total length in steel tape measure as per IS 1269(Part 2) is as below.(where 'L' is in meter)

Class I	Class II	Class III
± (0.1+0.1 x L)	± (0.3+0.2 x L)	± (0.6+0.4 x L)

1. For steel tapes of length upto 5 mtr. Permissible values are as per class I & II only.
2. For steel tapes of length above 5 mtr. Permissible values are as per class I, II and III.
3. For tapes in service the permissible value is equal to TWICE the values indicated in the above table.

Uncertainty of Measurement : ± 8.2 mm

(The reported uncertainty at above reading is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 corresponds to a confidence level of 95.45%.)

*** END ***

Calibrated By
Ankita Bhuvad
Technical Assistant

QRICS PUNE CALIBRATION

Authorized By
N. M. Shinde
Technical Manager

User discretion:

Note: 1) Any error / Complaint must be resolved with lab within 3 days from the receipt of certificate in written.