S A ELECTRONIC

PRIVATE LIMITED

Plot no. T-165/3/11, Priyadarshani Mahila
Swayamrojgar Indl. prem. co-op. soc. Ltd,
M.I.D.C Bhosari, Pune-411026
Tel : +91-20-27111581/2/3
Mail : saelectronics@vsnl.net
Visit at : www.saelectro.com

CERT. NO.: SA/0011/21-22

CALIBRATION CERTIFICATE

NAME OF THE CUSTOMER : **M/S. AS MECHFLUID SOLUTIONS PVT.LTD**

EQUIPMENT PARTICULARS : PRESSURE GAUGE

RANGE : 0 – 600 BAR

MAKE : WIKA

MODEL NO : 232.50.160(BOTTOM)

ACCURACY : +/- 1% FS

DATE OF CALIBRATION : 22/02/2022 **DUE ON : 21/02/2023**

STANDARDS USED FOR CALIBRATION:

INSTRUMENTS SERIAL NO. / MODEL NO.	*CERTIFICATE NO.*	*TRACEABILITY TO*	*VALIDITY*
110EPLO2/S-10/0 TO 600 BAR	*WP21/786 DT.21/07/2021*	*WIKA INST. INDIA PVT. LTD, PUNE*	*20/07/2022*
C1504201 / PC-UC-45B	*ARAI/CAL/2107/729, ARAI/CAL/2107/694*	*ARAI, PUNE*	*19/07/2022*

AMBIANT TEMP. : 26 °C HUMIDITY: 58%

OBSERVATIONS :

Sr. No.	Range	Standard Input in BAR				
	0 - 600 BAR	00.0	200	400	500	600
900112TZ		00.0	200	399	501	600
900112TB		00.0	200	403	497	600
900112TY		00.0	203	395	505	600
900112U6		00.0	197	398	495	600
900112U1		00.0	200	405	500	600
900112U2		00.0	198	404	500	600
900112U3		00.0	203	400	495	600
900112U9		00.0	200	395	504	600
900112TP		00.0	197	403	500	600
900112TZ		00.0	201	400	496	600

Remark: 1) All measurement standards used for Calibration are Traceable to National Standards.
2) The Instrument is calibrated and the error is under acceptable limit.

CALIBRATED BY:
PO NO. : 172388; DT: 08/12/2021

APPROVED BY:

S A ELECTRONIC PRIVATE LIMITED

S A ELECTRONIC

PRIVATE LIMITED

Plot no. T-165/3/11, Priyadarshani Mahila
Swayamrojgar Indl. prem. co-op. soc. Ltd,
M.I.D.C Bhosari, Pune-411026
Tel : +91-20-27111581/2/3
Mail : saelectronics@vsnl.net
Visit at : www.saelectro.com

CERT. NO.: SA/0010/21-22

CALIBRATION CERTIFICATE

NAME OF THE CUSTOMER : **M/S. AS MECHFLUID SOLUTIONS PVT.LTD**

EQUIPMENT PARTICULARS : PRESSURE GAUGE

RANGE : 0 – 600 BAR

MAKE : WIKA

MODEL NO : 232.50.160(BOTTOM)

ACCURACY : +/- 1% FS

DATE OF CALIBRATION : 22/02/2022 **DUE ON : 21/02/2023**

STANDARDS USED FOR CALIBRATION:

INSTRUMENTS SERIAL NO. / MODEL NO.	*CERTIFICATE NO.*	*TRACEABILITY TO*	*VALIDITY*
110EPLO2/S-10/0 TO 600 BAR	*WP21/786 DT.21/07/2021*	*WIKA INST. INDIA PVT. LTD, PUNE*	*20/07/2022*
C1504201 / PC-UC-45B	*ARAI/CAL/2107/729, ARAI/CAL/2107/694*	*ARAI, PUNE*	*19/07/2022*

AMBIANT TEMP. : 26 °C HUMIDITY: 58%

OBSERVATIONS :

Sr. No.	Range	Standard Input in BAR				
	0 - 600 BAR	00.0	200	400	500	600
900112U4		00.0	200	399	501	600
900112U7		00.0	200	403	497	600
900112U8		00.0	203	395	505	600
900112UA		00.0	197	398	495	600
900112TN		00.0	200	405	500	600
900112TH		00.0	198	404	500	600
900112TS		00.0	203	400	495	600
900112TU		00.0	200	395	504	600
900112TT		00.0	197	403	500	600
900112TQ		00.0	201	400	496	600

Remark: 1) All measurement standards used for Calibration are Traceable to National Standards.
2) The Instrument is calibrated and the error is under acceptable limit.

CALIBRATED BY:
PO NO. : 172388; DT: 08/12/2021

APPROVED BY:

S A ELECTRONIC PRIVATE LIMITED