Arndt Schimmelmann, Ph.D. Biogeochemical Laboratories, Indiana University

Certificate of Analysis

Hydrogen and Carbon Stable Isotope Ratio Reference Materials: Fatty Acid Ester Mixture F8

Compound	mean δD ‰	σ ‰	range SD ‰	n	mean δ ¹³ C ‰	σ ‰	range δ ¹³ C ‰	n	Conc. mg/mL
Tetradecanoic acid methyl ester #14M	-231.2	1.4	-229.3 to -232.3	4	-29.98	0.02	-29.96 to -29.99	3	0.084
Tetradecanoic acid ethyl ester #n14E	-231.2	2.7	-228.1 to -234.6	7	-29.13	0.03	-29.10 to -29.16	3	0.098
Hexadecanoic acid methyl ester #n16M	-166.8	1.7	-164.8 to -168.6	4	-29.90	0.03	-29.87 to -29.94	3	0.098
Hexadecanoic acid ethyl ester #IU 16E	-211.0	1.7	-209.5 to -213.5	4	-30.92	0.02	-30.90 to -30.95	3	0.092
Octadecanoic acid methyl ester #n18M	-206.2	1.7	-204.0 to -208.2	5	-23.24	0.01	-23.23 to -23.35	4	0.10
Octadecanoic acid ethyl ester #18E	-214.2	0.7	-213.2 to -214.9	4	-28.22	0.01	-28.22 to -28.24	3	0.098
Icosanoic acid methyl ester #2	-166.7	0.3	-166.4 to -167.1	3	-30.68	0.02	-30.66 to -30.71	3	0.10
Icosanoic acid ethyl ester #20E2	-195.5	1.2	-193.8 to -196.6	4	-26.10	0.03	-26.08 to -26.13	3	0.084

Notes

First, commercially available and custom-synthesized esters (\geq 99%) were tested for purity using GC and GC-mass spectrometry at Caltech. Subsequently, stable isotope ratios were determined at Indiana University by combusting multiple aliquots of the compounds off-line according to Schimmelmann et al. (1999) *Geochimica et Cosmochimica Acta* **63**, 3751-3766 (http://dx.doi.org/10.1016/S0016-7037(99)00221-5). Elemental hydrogen gases were measured and reported isotopically according to Coplen's (1996) guidelines relative to VSMOW (zero ‰) and normalized to SLAP (-428 ‰) (Coplen, T. B., 1996, New guidelines for reporting stable hydrogen, carbon, and oxygen isotope-ratio data. *Geochimica et Cosmochimica Acta* **60**, 3359-3360). Carbon isotope ratios from carbon dioxide gases are reported relative to the VPDB scale where NBS 19 and L-SVEC are defined as exactly +1.95 and -46.6 per mil, respectively (Coplen et al., 2006. New guidelines for δ¹³C measurements. *Analytical Chemistry* **78** (7), 2439-2441; http://dx.doi.org/10.1021/ac052027c). Mass-spectrometric data in customary δ-notation have a precision of ± 1.5 ‰ for δD and ± 0.05 ‰ for δ¹³C.

The data in the Table shown above represent the mean value, one sigma standard deviation, the range, the number of analyses, and the concentration in milligram per milliliter of solution in cyclohexane.

Primary reference materials used for isotopic characterization were VSMOW, SLAP, NBS 19, and L-SVEC. Isotopic and statistical analyses were performed at Indiana University by Arndt Schimmelmann.

Remedies for any claimed defect in these products will be limited to product replacement or refund of the purchase price. In no event shall Indiana University, the Biogeochemical Laboratories, or Caltech and their staff be liable for any damages, including but not limited to incidental or consequential damages. By accepting shipment of reference materials you agree to this limitation of liability.

This is a certified reference material (CRM) that can be traced to primary standards VSMOW, SLAP, NBS 19, and L-SVEC.

CERTIFIED 15 July 2008.

Arndt Schimmelmann, Ph.D., Senior Scientist, Indiana University, Department of Geological Sciences, Biogeochemical Laboratories, 1001 **East** 10th Street, Bloomington, Indiana 47405-1405, U.S.A, phone (812) 855-7645, fax (812) 855-7961,e-mail: aschimme@indiana.edu, personal home page: http://geology.indiana.edu/schimmelmann/index.html