ALGO® 3i

Newborn Hearing Screener

Patent Pending

MEDICAL DESIGN EXCELLENCE AWARDS® 2004 WINNER

'04 INDUSTRIAL DESIGN EXCELLENCE AWARDS BRONZE

so small...



...such a WONDER

The AABR® technology you can trust in a hand-held screener

- Highest clinical accuracy[1] — supported by numerous published studies within respected peer-reviewed journals
- A complete screen of the hearing pathway in one simple step

Fast and simple operation

- Unlike other screeners, the ALGO 3i device screens both ears simultaneously
- Unique SpeedScreen™ function allows you to conduct a test while entering patient information
- SmartHelp system offers helpful tips targeted to your needs
- Colored, graphical displays guide you easily through the screening process
- Fully automated, objective pass/refer results that require no interpretation
- Completely standardized, pre-set screening parameters – you don't have to make adjustments

Maximum adaptability

- Screen in a variety of settings with the versatile ALGO 3i SoftClip™ attachment
- Several language options are available for the screening process

Wireless data transfer and data management

- Wireless printing of screening results
- Wireless transfer of patient screening data to a PC
- Compatibility with a variety of data management systems

[1] *References on file at Natus*

ALGO® 3i
Newborn Hearing Screener
Patent Pending



The ALGO 3i Newborn Hearing Screening System includes:

- ALGO 3i screener
- Carrying case (back pack)
- Sensor cable
- Earphone cable
- Label Printer
- Rollstand Mount
- Battery pack and battery charger
- Instructional Materials (on CD)
- ALGO Check Kit
- Data Utility Software (DUS) and data export adapter





Specifications:

Sweep Rates
37 clicks/second (nominal) or
34 clicks/second (nominal) alternating

Power Supply

Battery	7.4V 1800mAH, Li Ion
Full Charge Operation	minimum 4 hours run time
Recharge Time	≈4 hours
Battery Charger	Input: 100-240V~, 50-60 Hz, Output: 12V ⎓ at 3.3A

Device Electrical Ratings
12V ⎓ at 1A

Nominal Dimensions

Screener	7.9 x 3.8 x 2.5 inches (H x W x D) 20.2 x 9.58 x 6.33 cm

Nominal Weight
Screener 2.0 lbs (0.93 kg)
(including battery and cables)

Safety and Compliance Standards
UL2601-1
CAN/CSA-C22.2 601-1
EN60601-1
EN60601-1-2

Classifications
Type BF applied part

Ordinary equipment for protecting against ingress of water (IPX0)

Not suitable for use in the presence of flammable anaesthetic mixture with air, oxygen or nitrous oxide

Continuous operation

Operating Temperature:
41° F (5° C) - 100° F (38° C)

Storage Temperature:
-22° F (-30° C) - 131° F (55° C)

Humidity:
5-90% non-condensing

Operating Altitude:
sea level - 6000 feet (1820 meters)

natus®

Natus Medical Inc. • San Carlos, CA USA
www.natus.com
customer _ service@natus.com
+1 650-802-0400, 800-255-3901
Fax: +1 650-802-0401

CE 0197

P/N 051167D