

5.2. Strategy


The screening strategy is used to assess diabetes risk. The Follow-up strategy can be used for treatment follow-up of diabetic patients or patients with chronic kidney disease (CKD).

Select the strategy you wish.

5.3. Medical Doctor choice

Up to 4 Medical Doctors (MD) can be chosen. The doctor selected last remains selected until later changed (green background). Simply pointing at another doctor's name will select a specific folder for their patients.


Click on the MD you wish. *The background of the selected MD becomes green.*



5.3.1. Screening strategy

In the screening strategy, the head electrodes are optional (see notes in chapter 3.8).

Uncheck the Headset bow if you wish to use EZSCAN without the head electrodes.




Headset checkbox

5.3.1.1. More results

In the screening strategy, the probability to have Insulin resistance and IGT scales are displayed.



For a patient with high risk, the probability to have diabetes complications is also displayed.



5.3.1.2. Clinical interpretation of the results

The EZSCAN results can be interpreted based on the results of a clinical trial conducted in Chennai, India.


The aim of this clinical study was to determine the ability of EZSCAN to detect abnormalities in carbohydrate metabolism as compared to a 2 Hours OGTT (oral glucose tolerance test).

This study showed a good correlation between the results of EZSCAN as compared to the OGTT.

To know more about this study go to Appendix.

The results of this study (Glucose and Insulin blood levels during an OGTT according to EZSCAN results and classification) can be displayed.

Double click on the background of the color graph (not on the colored bubble).



NOTE

The bottom part of this pop-up window is colored according to diagnosis ellipse and contains important useful explanations to understand the action of insulin and the relationship with the resulting blood sugar level.


5.3.2. Follow-up strategy

NOTE

In the follow-up strategy, the use of the **head electrodes is mandatory**. If you uncheck the Headset check box, you will be automatically transferred to the screening strategy.

Click on Follow-up checkbox.


In the expert mode, the probability to have Feet sudomotor impairment and the probability to have cardiac neuropathy are displayed:



Follow-up checkbox

The probability risk to develop Chronic Kidney Disease (CKD) can also be displayed:

Click on CKD checkbox.



CKD checkbox

6. Print the results

The following instructions are available for both standard and expert mode.


6.1. Configure the print report

1. Press the “Configure” button to open the Configuration window where the print report details can be specified.

2. Double-click on each field to prompt the alphanumerical keyboard to appear on screen.


3. Enter the relevant information into the fields:

   NAME OF THE HOSPITAL
   NAME OF THE DEPARTMENT
   NAME OF THE DOCTOR



6.2. Configure the printer usage


Two sets of printing parameters can be selected by the user.




6.2.1. Print style

Two print styles are available.

1. Check the 'Background' checkbox if you want a background.
2. Choose the screen style you wish: 'Screen' or 'Classic' style.



'Screen' style



'Classic' style

6.2.2. Print auto

Check the 'Auto' button check-box for automatically printing a test report after the scan completion…

NOTE

Be sure of the printer installation and its configuration settings before selecting the "Auto" mode.

OR

Uncheck the 'Auto' button check-box for selectively manually printing a report only upon pressing the 'PRINT' button.

6.2.3. Preview and Print


If you prefer to preview the report on screen before deciding to print it:

1. Check the ‘Preview’ option check-box.
2. Click on the Print button.
3. Click on  .

6.2.4. PDF report

1. Check the ‘pdf’ option check-box.
2. Click on the Print button.
3. By default, the file name is automatically filled with “Name of the patient Date Time” but you can modify it if you want.
4. Save the pdf.

pdf
checkbox



Press the “Configure” button again to close the Configuration window.

7. Retrieve previous patient data

All patients' data are automatically saved into specific folder on the Master Unit's hard disk. Data are arranged by Medical Doctor (MD1, MD2, etc.), date and time when the scan was done from the most recent to the most ancient.


Three methods are available to look for a patient's scan result:

Method 1

1. Double-clicking on the 'Name' field to erase the previous patient's name *(the 'Name' field will become empty).*

2. Press the 'READ 1' button: the whole content of the selected Medical Doctor folder will automatically be displayed in the new window *(all the patients previously scanned).*

3. Double-Click on the patient's name you want to retrieve the data.



Method 2

1. Double-clicking on the 'Name' field to erase the previous patient's name *(the 'Name' field will become empty).*

2. Enter the patient's name you are looking for in the 'Name' field *(double check the spelling)*, then press the 'READ 1' button.

   Only this patient's previous scan results will be displayed in the new window.

Method 3

1. Double-clicking on the 'Name' field to erase the previous patient's name (the 'Name' field will become empty)

2. Enter the patient's first letter(s) name you are looking for in the 'Name' field (double check the spelling), then press the 'READ 1' button.

   All the patients whose names have the same beginning letters will be displayed.

Examples:

Enter 'M' in the 'Name' field and press the 'READ 1' button.

The new window will display all the patients whose names begin with an M.

NOTES

The 'Name' field is NOT case sensitive, and can contain digits (such as Patient ID/IDENTITY CODE).

Since the 'Name' field can contain a maximum of 30 characters, it is advisable to use the 'Comment' field to enter additional patient information such as ID code, etc…

8. Share patient's data


8.1. Transfer data to a USB key

Thanks to this function, the data can be transferred from one EZSCAN to another.

All patients' data are saved into a specific database folder on the Master Unit's hard disk, which can be found on C:\Program Files\IMPETO\.


To copy all the data base:

1	DOUBLE-CHECK ON THE TARGET PC THAT IT IS FREE OF VIRUS.
2	Go to C:\Program Files\IMPETO\
3	Copy the 'DataBase18F2.13' entire folder to the external USB flash disk.



To copy one or more data:

1	DOUBLE-CHECK ON THE TARGET PC THAT IT IS FREE OF VIRUS.
2	Go to C:\Program Files\IMPETO\
3	Click on ‘DataBase18F2.13’
4	Click on MD1, MD2…according to where the data is stored.
5	Copy the Data to the external USB flash disk.




8.2. Access data through office local network


< Requires the installation of EZSCAN software on the remote computer >

>>> Check with your distributor.

9. Install a USB printer



1	DOUBLE-CHECKING ON ANOTHER PC THAT THE USB PRINTER IS FREE OF VIRUS.
2	Insert a USB key or a USB-compatible CD-ROM reader with the appropriate printer drivers for Windows XP or Windows 7.
3	Follow the standard Windows procedure to install the new printer drivers and configure the printer default options.



4 Connect the USB printer to one of the 2 USB ports available at the bottom part of the Master Unit's rear panel.

10. Connect a video projector



For a large audience or demonstration purpose, it is possible to connect an external larger screen or a video projector on the VGA-type PC standard video connector located on the rightmost part of the Master Unit's rear panel.

11. Switch of EZSCAN



Always end EZSCAN session and turn off the Master Unit only by pressing the on/off button on the front of the Master Unit.

NOTE

The procedure above is causing the device to 'hibernate' so that the same display will be found the next time the Master Unit is powered on.

To fully reset the operation, switch off the machine following standard Windows "turn off" procedure on the display screen, after closing eZscan application software.

CAUTION

REMOVING THE POWER CORD FROM MASTER UNIT OR FROM WALL SOCKET BEFORE COMPLETE UNIT SHUT OFF, MAY IRREMEDIABLY CORRUPT DATA ON THE UNIT'S HARD DISK.

12. Disconnect the electrodes

Grab the larger part of each connector and slide it softly backwards.






13. Troubleshooting

13.1. Check the cable connections

1. Press the “Configure’ button at the right side of the screen.

2. Press the “Test device” button (need to point to this) at the left side of the new window.



The lower part of this test panel will enable you to check the correct connection of the electrodes:

Pressing either –HANDS+ or +HANDS- buttons, should illuminate both sides of the hands electrodes.

Pressing the “Left” button close to the +HANDS- button should illuminate the left hand electrode only.

Pressing the “Right” button close to the –HANDS+ button should illuminate the right hand electrode only.



Pressing either –FEET+ or +FEET- buttons, should illuminate both sides of the feet electrodes.

Pressing the “Right” button close to the –FEET+ button should illuminate the right foot electrode only.


Pressing the “Left” button close to the +FEET- button should illuminate the left foot electrode only.

NOTE

The current aux electrode has no LED attached but should not need to be tested if the hands and feet electrodes are correctly connected to the Master Unit.

Failure to illuminate the correct electrode(s) should prompt you to double-check the physical cable connections and possibly swap the wrong cables before calling for assistance.

13.2. Test the EZSCAN hardware



1. Press the 'Auto Test' button to initiate the same complete Test and Calibration program which has been launched before leaving the factory; its print report (PDF format) is sitting on the desktop screen.

2. Press the "Exit" button, then the "Configure" button in order to get back to the EZSCAN home page.

NOTE

This test must be launched with NO PATIENT installed on electrodes.


13.3. Calibrate the touch screen


In case the pointing capability on the touch screen does not appear accurate enough, follow the following procedure to calibrate or 'fine-tune' the EZSCAN touch screen operation:

1 Click on the ‘ELO Touch’ icon on the left to launch the ‘Touch Utility’ program





2 Click on the ‘Align’ icon to launch a simple 3-step procedure





3 When requested, press with your finger on the 3 targets presented sequentially






4 Press ‘Confirm’ box.

14. Upgrade EZSCAN firmware

Impeto Medical will regularly provide software upgrades to ensure continuous enhancement of EZSCAN products. Check with your distributor or on www.impeto-medical.com website for their availability on a regular basis.



1	Close EZSCAN program.
2	AFTER DOUBLE-CHECKING ON ANOTHER PC THAT IT IS FREE OF VIRUS, insert a USB Flash Disk into one of the 2 USB ports available on the front panel below the screen.
3	Click on the self-extractible software upgrade file for automatic upgrading.
4	Click on EZSCAN icon in the middle of the desktop screen.

15. Precaution of use and technical specifications

Principle of EZSCAN

EZSCAN® is based on two well-known principles, reverse iontophoresis and electrochemistry. EZSCAN collects physiological data by means of Chronoamperometry and processes them with analysis software:

1. Reverse iontophoresis: this technology applies a low DC voltage between electrodes located on specific areas of the skin with high density of appendage ducts in order to evaluate the electroporation level of subdermal concentrations (Chimadzhev *et al*., 1998),
2. Electrochemistry: this technology applies a sudden step of DC low voltage between two electrodes in a fluid to obtain data related to the concentration of this solution.

EZSCAN® devices use both technologies, coupled and complemented with exclusive software which analyzes cutaneous conductance data collected through the Chronoamperometry measurements.

The EZSCAN system is composed of a EZSCAN electronic device integrated into a touch-screen Windows XP-based Industrial-grade PC computer and connected to 6 electrodes placed on the feet, the hands and head, as well as EZSCAN acquisition control software also responsible for archiving the measurements taken on all patients.

General operation of EZSCAN

The patient positions his bare feet on the feet electrodes, applies the palms of his hands to the hands electrodes, and the operator can affix forehead electrodes to the head.

After entering the information specific to the patient (name, age, gender, height, weight), the operator will ask the software to activate the electronic circuitry of the Chronoamperometry device, which will apply DC reference voltages successively to electrode couples and will measure the evolution in time of the electrical current passing through the electrodes.

Several successive cycles of measurements are carried out in an automatic way and all the measured values are recorded on the hard disk. Data-processing is then performed to compute the conductance on each electrode.

The number of measurements made by means of the EZSCAN device, as well as the number of the recorded patients is only limited by the capacity of the integrated hard disk.

At the end of the measurements cycle, which lasts approximately 2 minutes, the expert can visualize a synthesis of the reprocessed data of the patient and with the possibility of printing a sheet of results to include in the file of the patient.

No control is accessible to the patient and the use of the EZSCAN system is strictly reserved to physicians in the framework of a professional use in medical clinic or during visits in hospital.

Precautions of use and maintenance

Transport

The system EZSCAN is delivered under the conditions of packing described previously.

If it is necessary to pack, to transport or deliver EZSCAN after its use, it is recommended to arrange all its elements in their housing and position of origin.

Pay particular attention to carefully arrange the electrode cables in the locations especially designed for them in the protection foams, not to damage them or weaken them.

The EZSCAN box can be stored upright or laid down.
The EZSCAN device does not comprise any accumulator.

Recycling

At the end of its lifetime, the EZSCAN system must be returned to Impeto Medical's authorized distributor, which will return it to Impeto Medical, in order to ensure the recycling of certain components.

The components and the accessories of EZSCAN are free of mercury and of components containing this element.

Protection against the very wet environments

It is advised not to use EZSCAN in a very wet environment.

EZSCAN is not protected from the effects of the discharges of an external defibrillator, nor against high frequency currents or strong electromagnetic disturbances. The use of portable telephones or wireless fixed telephones can cause disturbances of the signal.

Electromagnetic compatibility

The use of portable and mobile RF communication devices (for example: cellular telephones) can influence the analysis carried out during the recording, as the recorded signals can be disturbed by electromagnetic interferences.

The device should not be used in the presence of ionizing radiations (x-rays, gamma rays …) because those could erase the internal storage.

For the tables concerning the electromagnetic emissions and the immunity of the recorder, see the appendices of this handbook.

Maintenance of the device and electrodes

No particular maintenance is necessary.

The external case and the cables of the electrodes can be cleaned using a slightly wet tissue or soaked with soapy water. Do NOT use any detergent product, nor alcohol or acetone.

To avoid corrosion like ensuring the best conditions of hygiene, the feet and hands electrodes must be cleaned immediately after each patient with an anti-bacterial and fungicidal solution.

Warranty

EZSCAN has one full year warranty period (12 months from the date of installation, parts and labor). After the warranty period has expired, customers can subscribe a service contract for the maintenance of their equipment.

Feet and hands electrodes, as well as their associated cables will have their wear and tear and hence are not covered by the warranty. Take care to maintain them in good condition of aspect and operation to prolong their lifespan to the maximum, then to order the replacement parts at Impeto Medical or its authorized distributors.

WARNING! The warranty is null and void if EZSCAN was opened and repaired by any unauthorized person. Warranty will hold well only against manufacturing defects and certainly not for any mechanical damage due to mishandlings. Refer to manual for proper use of the equipment.

Description of the pictograms affixed on the case

The following pictograms are affixed on EZSCAN:



Indicates that the parts applied of the device are of type BF (IEC 60601-1).

III Class of electric protection.

CE 0499 EZSCAN is in conformity with European Directive 93/42/EEC.

Use of the electrodes

The data relating to the use of the electrodes are provided as an indication.

To ensure high quality recordings, a particular care must be taken to the preparation of the skin of the patient and to the installation of the electrodes.

It is also advisable to check if the electrode cable connectors are well inserted in the corresponding sockets at the rear of the EZSCAN Master Unit.

Clean the skin of the patient to establish a good electrode/skin contact.

WARNING! Use only the original electrodes provided by Impeto Medical.

Impeto Medical cannot guarantee the results of measurements carried out with electrodes coming from other manufacturers, even if they have a similar appearance!

When it is not used, the EZSCAN system can be stored without being necessary to disconnect the electrode cables.

• **Forehead electrodes installation**

Put in place the head gear with the reusable electrodes and adjust its size to the patient head. The blue side of the cable will be connected to the electrode placed on the right side of the patient's head the white side on the left.

• **Positioning on the electrodes plates**

The patient must position the sole of his bare feet in the center of the feet electrodes and his hands well flat in the center of the hands electrodes, hands must be applied firmly. Except in the event of characterized overweight, the patient will be preferably standing upright on the plate of the feet electrodes.

Interrupting or stopping the recording

The recorder will stop automatically at the end of the measurements cycle, after approximately 2 minutes.

To stop the recorder BEFORE the end of the measurements cycle, simply click on the same button which is now labeled "STOP" during the cycle of measurement.

Recommendations to the operator

As for any medical examination, it is recommended to check the environmental conditions and the condition of the patient before carrying out a measurement:

- Room temperature between 18 and 25 degrees Celsius,
- Patient not having carried out an intense effort during 2 hours preceding the examination (except in the event of effort test),
- Cardiac holter measurements going on can be altered during the 2 minutes time scan.

Contra-indications

The repeated use of EZSCAN does not create any side effect, but no specific tests have been carried out to date under the following particular conditions, so it is advised not to proceed to measurements on the following patients:

- People with any open or bleeding wound at any electrode contact surface location,
- People with any type of implantable device,
- People with missing leg(s) or hand(s),
- Pregnant Women or women who are uncertain about a possible pregnancy, Subjects under the age of 21.

Calibration

EZSCAN performs automatically a self-calibration procedure before each patient. So it assures that each electrode got enough contacts to proceed for the measurements. Therefore, there is no need to calibrate the device before operation. However, it is possible to double-check the correct operation of the elements of the whole system by carrying out, in case of doubt or at least once in a year a calibration procedure using the “Calibration Test Box” accessory available from Impeto Medical.

Accessories and disposable parts

The following elements are available for EZSCAN:

Code	Description
SuMU	EZSCAN Master Unit
APH2	Hands plate with two electrodes and its connecting cable
APF2	Feet plate with two electrodes and its connecting cable
HDGR	Head gear with two electrodes
SCC2	Head electrodes cable with 2 clips
UMSu	Installation manual
ASAFE	Bottle of ANIOS Surfa’Safe spray cleaner

Technical specifications

• Description

EZSCAN is a device capable of measuring galvanic skin responses and is designed for the medical professional.

EZSCAN is a class IIa medical device (according to the EEC directive 93/42) and a class II medical device (according FDA 21CFR 882.1540).

• Characteristics of acquisition

Storage support:	80 Gb Hard disk
Mode:	No compression
Frequency of acquisition:	100 Hz
Resolution:	1 nanoSiemens
Display units:	microSiemens
Display precision:	+ or – 1%

• Electrical characteristics

Measurement voltage:	1V to 4V DC typical
Frequency:	0 Hz (DC current)
Dynamic resolution:	10 bits
Power requirement of the unit:	100 - 240V AC + or – 3% variation at 50 - 60Hz freq
Power supply:	250 W power supply in conformity with the EN60601-1

• Sealing

Index of protection:	IP41

• Mechanical characteristics

Master Unit alone	13 kg
Master Unit dimensions	38 x 38 x 30 cm
Weight of the Master Unit + electrodes:	15 kg
Weight of the transportation box:	16 kg
Dimensions of the shipping box:	54 x 46 x 39 cm

Storage and conditions of use of the apparatus and electrodes

	Temperature	Moisture	Pressure (without condensation)
Operation	15° to 35°C	20% to 80%	700 to 1060 hPa
Storage and transport	-0° to 45°C	20% to 80%	500 to 1060 hPa

Table 1

Guidance and manufacturer's declaration – electromagnetic emissions		
EZSCAN is intended for use in the electromagnetic environment specified below. The customer or the user of EZSCAN should assure that it is used in such an electromagnetic environment.		
Emissions test	**Compliance**	**Electromagnetic environment – guidance**
RF emissions CISPR 11	Group 1	EZSCAN uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	EZSCAN is suitable for use in all establishments, including domestic establishments and those directly connected to the public low voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Complies	

Table 2

Guidance and manufacturer's declaration – electromagnetic immunity			
EZSCAN is intended for use in the electromagnetic environment specified below. The customer or the user of the EZSCAN should assure that it is used in such an electromagnetic environment.			
Immunity test	**IEC 60601 test level**	**Compliance level**	**Electromagnetic environment guidance**
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	± 2 kV for power supply lines ± 1 kV for input/output lines	± 2 kV for power supply lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV line(s) to line(s) ± 2 kV line(s) to earth	±1 kV mode differential ±2 kV mode common	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5 % *UT* (>95 % dip in *UT*) for 0,5 cycle 40 % *UT* (60 % dip in *UT*) for 5 cycles 70 % *UT* (30 % dip in *UT*) for 25 cycles <5 % *UT* (>95 % dip in *UT*) for 5 s	0 Vac during 10 ms 92 Vac during 100 ms 161 Vac during 1 s 0 Vac during 5 s	Mains power quality should be that of a typical commercial or hospital environment. If the user of the EZSCAN requires continued operation during power mains interruptions, it is recommended that the EZSCAN be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m for 50Hz and for 60Hz	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE *UT* is the a.c. mains voltage prior to application of the test level			

Table 3 - for ME equipment and ME systems that are not life-supporting

Guidance and manufacturer's declaration – electromagnetic immunity			
The EZSCAN is intended for use in the electromagnetic environment specified below. The customer or the user of the EZSCAN should assure that it is used in such an electromagnetic environment.			
Immunity test	**Immunity test**	**Immunity test**	**Immunity test**
			Portable and mobile RF communications equipment should be used no closer to any part of the EZSCAN, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 V rms	**Recommended separation distance**
		3 V/m	$d = 1.17 \sqrt{P}$
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2,5 GHz		$d = 1.17 \sqrt{P}$ 80 MHz to 800 MHz $d = 2.33 \sqrt{P}$ 800 MHz to 2,5 GHz where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey (1) should be less than the compliance level in each frequency range. (2) Interference may occur in the vicinity of equipment marked with the following symbol: 
NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.			
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

(1) Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the EZSCAN is used exceeds the applicable RF compliance level above, the EZSCAN should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the EZSCAN.

(2) Over the frequency range 150 kHz to 80 MHz, field strengths should be less than [*V*1] V/m.

Table 4 – for ME equipment and ME systems that are not life-supporting

Recommended separation distances between portable and mobile RF communications equipment and the EZSCAN			
EZSCAN is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the EZSCAN can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the EZSCAN as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter W	**Separation distance according to frequency of transmitter** m		
	150 kHz to 80 MHz $d = 1.17 \sqrt{P}$	**80 MHz to 800 MHz** $d = 1.17 \sqrt{P}$	**800 MHz to 2,5 GHz** $d = 2.33 \sqrt{P}$
0,01	0.12	0.12	0.23
0,1	0.37	0.37	0.74
1	1.16	1.16	2.33
10	3.69	3.69	7.37
100	11.6	11.6	23.3
For transmitters rated at a maximum output power not listed above, the recommended separation distance *d* in metres (m) can be determined using the equation applicable to the frequency of the transmitter, where *P* is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer. NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies. NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

GRAND-DUCHE DE LUXEMBOURG

Société Nationale de Certification et d'Homologation s.à r.l.

Notified Body
Organisme notifié - Benannte Stelle
N° 0499

EC Certificate

Certificat CE / EG Zertifikat

Full Quality Assurance System Approval Certificate according to Annex II (excluding section 4) of Medical Devices Directive (93/42/EEC)

Certificat de Système complet d'Assurance de Qualité selon Annexe II (à l'exclusion du point 4) de la Directive relative aux Dispositifs Médicaux (93/42/CEE)

Zertifikat des vollständigen Qualitätssicherungssystems nach Anhang II (ausgenommen Abschnitt 4) der Richtlinie über Medizinprodukte (93/42/EWG)

We certify that the Full Quality Assurance System of the company:
Nous certifions par la présente que le Système Complet d'Assurance de la Qualité de la société:
Wir bestätigen, daß das vollständige Qualitätssicherungssystem der Firma:

IMPETO MEDICAL

17, rue Campagne Première
FR-75014 PARIS

is in conformity with the following requirements:
est conforme aux exigences suivantes:
den folgenden Anforderungen entspricht:

Annex II (excluding section 4) of Medical Devices Directive (93/42/EEC)
Annexe II (à l'exclusion du point 4) de la Directive relative aux Dispositifs Médicaux (93/42/CEE)
Anhang II (ausgenommen Abschnitt 4) der Richtlinie über Medizinprodukte (93/42/EWG)

The scope of the Full Quality System is detailed in annex, which is a constituent part of this certificate.
Le domaine d'application du Système Complet d'Assurance de la Qualité est détaillé dans l'annexe qui est partie constituante du présent certificat.
Die Zweckbestimmung des vollständigen Qualitätssicherungssystems ist im Anhang aufgeführt und ist Bestandteil des vorliegenden Zertifikats.

Certificate *Certificat / Zertifikat* **No. 0643945-03**

The certification is valid until 2014-10-19 and is subject to a yearly surveillance.
La certification est valide jusqu'au 2014-10-19 et est soumise à une surveillance annuelle.
Die Zertifizierung ist gültig bis zum 2014-10-19 und unterliegt einer jährlichen Überwachung.

Luxembourg, 2009-10-20

Société Nationale de Certification et d'Homologation s.à r.l.
11, rue de Luxembourg ✉ B.P. 23
L-5201 Sandweiler / Luxembourg
☎ (Int+352) 35 72 14-250 Fax (Int+352) 35 72 14-244
F2-1.M2e4 / V1 / 2009-08-27

Claude LIESCH
Directeur

GRAND-DUCHE DE LUXEMBOURG

Société Nationale de Certification et d'Homologation s.à r.l.

Notified Body
Organisme notifié - Benannte Stelle
N° 0499

Annex of Certificate No.
Annexe du certificat no. / Anhang zum Zertifikat Nr.

0643945-03

IMPETO MEDICAL

Scope: Design, manufacture and final inspection of
Domaine d'application: Conception, fabrication et contrôle final de
Zweckbestimmung: Auslegung, Fertigung und Endkontrolle von

Analyseur Chronoampèrométrique Numérique

EZSCAN

This annex is only valid if attached to the certificate mentioned above.
La présente annexe est seulement valable en relation avec le certificat mentionné ci-dessus.
Dieser Anhang hat nur Gültigkeit in Verbindung mit dem oben genannten Zertifikat.

Luxembourg, 2009-10-20

Société Nationale de Certification et d'Homologation s.à r.l.
11, rue de Luxembourg ✉ B.P. 23
L-5201 Sandweiler / Luxembourg
☎ (Int+352) 35 72 14-250 Fax (Int+352) 35 72 14-244
F2-1.M2a / V1 / 2009-08-27

Claude LIESCH
Directeur

Appendix

EZSCAN, A NEW NON-INVASIVE TECHNOLOGY, HAS THE SAME SENSITIVITY TO DETECT ABNORMALITIES IN INSULIN RESISTANCE SUBJECTS AS THE 45MN OGTT GLUCOSE VALUES

PRESENTED AT EASD 2009

A. Ramachandran[1], A. Moses[2], C. Snehalatha[1], J. Deslypere[3]
[1]India Diabetes Foundation, Chennai, India, [2]Moses Diabetes Centre, Chennai, India, [3]SGS, Singapore, Singapore

AIMS

The EZSCAN is a new patented technology which uses low level DC voltage inducing reverse iontophoresis, together with chronoamperometry to evaluate the functions of tissues in specific locations of the body. The whole test takes only 3 minutes, does not require patient preparation and does not cause inconveniences for the subjects. The aim was to determine the ability of EZSCAN to detect abnormalities in carbohydrate metabolism as compared to a 45 minutes OGTT.

METHODS

A cross-sectional study was done in Chennai, India [1] in subjects who were not known to have any abnormalities in carbohydrate metabolism. Moreover all subjects underwent an OGTT 8 months before the study, showing a normal glucose tolerance status. A frequent sampling OGTT with measurements of glucose and insulin levels at 10 time points was done. Testing with the EZSCAN, a new non-invasive device to test carbohydrate metabolism [2-4], was also performed. A reading of < 40% was considered normal on the EZSCAN Insulin Resistance scale. Subjects were divided in three groups: normal (green color reading onthe EZSCAN; value < 40%) and abnormal (yellow color reading: value between 40% and 65% and orange/red color reading: value > 65%). The criterion of Abdul-Ghani et al. [5] was used for the cut-off values of the 45 min OGTT (8.6mmol/l).

RESULTS

A total of 30 subjects were assessed with the following demographics:

- Age: 41 ± 11 years
- BMI: 28 ± 6 $kg \cdot m^{-2}$
- M/F ratio: 1
- OGTT $_{45mins}$ (mmol/l): 8.66 ± 2.35
- OGTT $_{2hours}$ (mmol/l): 7.21 ± 2.21

Of the 11 subjects who were found to have abnormal $OGTT_{45mins}$ results (Fig.1A), 4 had IGT and 2 had DM. Five of these 6 cases tested positive with the EZSCAN (83% sensitivity), while 1 case (23 yrs old Type I DM) was not detected. There was a good correlation between the results of the EZSCAN as compared to the $OGTT_{45mins}$ (r=0.55, p=0.002).

Graphical plots of the time profile of glucose (Fig.1B) and insulin (Fig.1C) levels show that the EZSCAN measurements (green=normal EZSCAN reading and yellow, orange/red=abnormal EZSCAN reading) correlate better with the presence of insulin resistance rather than with insulin deficiency.




Fig.1 : Mean values of Glucose and Insulin concentrations over time as function of EZSCAN classification

CONCLUSION

The EZSCAN is a promising new tool able to detect the presence of early carbohydrate metabolism dysfunctions in insulin resistant subjects. Compared to tests currently done in daily practice, the EZSCAN is fast and noninvasive and it can be done at any moment of the day without fasting need.

REFERENCES

[1] Ramachandran A. et al. Submitted for publication, 2009.
[2] Brunswick P. et al. *Exploration de l'innervation des glandes sudoripares chez le diabétique.* Présentation à Congrès Conjoint Alfediam-SFE, Diabète – Endocrinologie, Marseille, p 44, 2007.
[3] Chizmadzhev YA et al. *Electrical properties of skin at moderate voltages: Contribution of appendageal macropores.* Biophys J 74:843-856, 1998.
[4] Quinton, P. M., *Cystic Fibrosis: Lessons from the Sweat Gland.* Physiology 22: 212-225, 2007.
[5] Abdul-Ghani, M. A., Lyssenko, V., Tuomi, T., DeFronzo, R. A., Groop, L., *Fasting Versus Postload Plasma Glucose Concentration and the Risk for Future Type 2 Diabetes: Results from the Botnia Study.* Diabetes Care 32: 281-286, 2009.

Impeto Medical SAS - info@impeto-medical.com - www.impeto-medical.com
17, rue Campagne Première, 75014 PARIS FRANCE - Tel: +33 1 4320 1640 - Fax: +33 1 4322 8075

impetomedical

IMPETO MEDICAL

17, rue Campagne Première

75014 PARIS – FRANCE

Tel: +33 (0)1 4320 1640

Fax: +33 (0)1 4322 8075

Mail: info@impeto-medical.com

CE 0499

www.impeto-medical.com