

Table of contents

\USER

SD

VASO

Neurovascular Coupling

[rslh_ep3d_vaso_v220920_singleShot_fatSat-on](#)

\\USER\SD\VASO\Neurovascular Coupling\rslh_ep3d_vaso_v220920_singleShot_fatSat-on

TA: 0:26 PM: REF Voxel size: 0.9×0.9×0.9 mmPAT: 3 Rel. SNR: 1.00 : 26dc5a59

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	16
FoV read	133 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
TR 1	42.0 ms
TR 2	2221 ms
TE 1	15.00 ms
Averages	1
Multi-echo Shots	1
Filter	Distortion Corr.(3D)
Coil elements	AC

Contrast - Common

TR 1	42.0 ms
TR 2	2221 ms
TE 1	15.00 ms
Multi-echo spacing	38.38 ms
Magn. preparation	Non-sel. HSN IR
TI 1	912 ms
TI 2	1542 ms
Flip angle	33 deg
Fat suppr.	Fat sat.
Magn. Prep. Shots	1

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	10
Pause after meas. 1	0.0 s
Pause after meas. 2	0.0 s
Pause after meas. 3	0.0 s
Pause after meas. 4	0.0 s
Pause after meas. 5	0.0 s
Pause after meas. 6	0.0 s
Pause after meas. 7	0.0 s
Pause after meas. 8	0.0 s
Pause after meas. 9	0.0 s

Resolution - Common

FoV read	133 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
Base resolution	148
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Acc. factor PE	1
Ref. lines PE	24
Acc. factor 3D	3
Ref. lines 3D	16
CAIPI 3D Shift	1
Reference Scan Mode	GRE/separate
CAIPIRINHA mode	Free

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	16
FoV read	133 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
TR 1	42.0 ms
TR 2	2221 ms
Multi-slice mode	Interleaved
Series	Ascending
Multi-echo Shots	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm

Geometry - AutoAlign

Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	Fat sat.

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	133 mm
R >> L	133 mm
F >> H	15 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	297.191255 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	200.000 V

Sequence - Part 1

Introduction	On
Dimension	3D
Reordering	Linear
Asymmetric echo	Off
Contrasts	1
Multi-slice mode	Interleaved

Sequence - Part 1

Echo spacing	1.01 ms
Bandwidth	1126 Hz/Px

Sequence - Part 2

EPI factor	37
Segmentation	3
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
Turbo factor	15

Sequence - Special

PATRef FA	3 deg
RF duration	1480 us
RF BWT product	8
Ernst T1	1200 ms
PATRef prep. shots	10
Volume dummy shots	0
Dummy Measurements	0
Invert PE	Off
Min. TE if PF	On
Echo Time Shift	On
Ramp Sampling	On
NORDIC	Off
SVDPC	On
Sym VASO	Off
Alternate RO	Off
Invert RO	Off
Water Exc.	-none-
External PC	per Series
Saturation RF	per Shot
EPI rise time factor	1.10
Mosaic DICOMs	On
Modify Ice Config	Off
HSN RF power scale	2.00
Inversion Delay	650 ms
Relaxation Delay	0 ms
Var. FA /MAGEC	4

Sequence - Assistant

Mode	Off
------	-----