

<b>Table of contents</b>
--------------------------

\\USER				
	SD			
		VASO		
			Neurovascular Coupling	
				<a href="#">rslh_ep3d_vaso_v220920_multiShot_fatSat-off</a>

\\USER\SD\VASO\Neurovascular Coupling\rslh\_ep3d\_vaso\_v220920\_multiShot\_fatSat-off

TA: 0:32 PM: REF Voxel size: 0.9×0.9×0.9 mmPAT: 3 Rel. SNR: 1.00 : 26dc5a59

**Properties**

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

**Routine**

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	16
FoV read	133 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
TR 1	24.7 ms
TR 2	2822 ms
TE 1	8.97 ms
Averages	1
Multi-echo Shots	1
Filter	Distortion Corr.(3D)
Coil elements	AC

**Contrast - Common**

TR 1	24.7 ms
TR 2	2822 ms
TE 1	8.97 ms
Multi-echo spacing	20.98 ms
Magn. preparation	Non-sel. HSN IR
TI 1	808.2 ms
TI 2	1178.7 ms
Flip angle	33 deg
Fat suppr.	None
Magn. Prep. Shots	2

**Contrast - Dynamic**

Averages	1
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	10
Pause after meas. 1	0.0 s
Pause after meas. 2	0.0 s
Pause after meas. 3	0.0 s
Pause after meas. 4	0.0 s
Pause after meas. 5	0.0 s
Pause after meas. 6	0.0 s
Pause after meas. 7	0.0 s
Pause after meas. 8	0.0 s
Pause after meas. 9	0.0 s

**Resolution - Common**

FoV read	133 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
Base resolution	148
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

**Resolution - iPAT**

PAT mode	CAIPIRINHA
Acc. factor PE	1
Ref. lines PE	24
Acc. factor 3D	3
Ref. lines 3D	16
CAIPI 3D Shift	1
Reference Scan Mode	GRE/separate
CAIPIRINHA mode	Free

**Resolution - Filter Image**

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

**Resolution - Filter Rawdata**

Raw filter	Off
Elliptical filter	Off

**Geometry - Common**

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	16
FoV read	133 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
TR 1	24.7 ms
TR 2	2822 ms
Multi-slice mode	Interleaved
Series	Ascending
Multi-echo Shots	1

**Geometry - AutoAlign**

Slab group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm

**Geometry - AutoAlign**

Initial Rotation	0.00 deg
Initial Orientation	Transversal

**Geometry - Saturation**

Saturation mode	Standard
Fat suppr.	None

**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

**System - Miscellaneous**

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

**System - Adjustments**

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

**System - Adjust Volume**

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	133 mm
R >> L	133 mm
F >> H	15 mm
Reset	Off

**System - pTx Volumes**

B1 Shim mode	TrueForm
Excitation	Slab-sel.

**System - Tx/Rx**

Frequency 1H	297.191255 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	200.000 V

**Sequence - Part 1**

Introduction	On
Dimension	3D
Reordering	Linear
Asymmetric echo	Off
Contrasts	1
Multi-slice mode	Interleaved

**Sequence - Part 1**

Echo spacing	1.05 ms
Bandwidth	1126 Hz/Px

**Sequence - Part 2**

EPI factor	19
Segmentation	6
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
Turbo factor	15

**Sequence - Special**

PATRef FA	3 deg
RF duration	1480 us
RF BWT product	8
Ernst T1	1200 ms
PATRef prep. shots	10
Volume dummy shots	0
Dummy Measurements	0
Invert PE	Off
Min. TE if PF	On
Echo Time Shift	On
Ramp Sampling	On
NORDIC	Off
SVDPC	On
Sym VASO	Off
Alternate RO	Off
Invert RO	Off
Water Exc.	-none-
External PC	per Series
EPI rise time factor	1.10
Mosaic DICOMs	On
Modify Ice Config	Off
HSN RF power scale	2.00
Inversion Delay	650 ms
Relaxation Delay	0 ms
Var. FA /MAGEC	4

**Sequence - Assistant**

Mode	Off
------	-----