

10/23/2019 → time PAR
sensors logging
@ 15 min int.

~~next to each other and logging
at 15 min interval~~

10/24/2019

Wind sensor logging
PAR, RH, soil temp logging

CB - Thermistor & scales all logging. WOO! 1700!

CB - Downloaded scale data & thermistor data.
Replaced East Rpi ~~the~~, since it had frozen
for 2nd time recently (stopped on 10/26 AM).

- Changed Water timer from 19:00 to 18:05-20:05!
(since LEDs turn off at 20:00).

- Pulled leaf from W-6 (on scale) since it was ~~at~~
creating gap between stem and plastic wrap.

Box fans turned on 1600 real time

Not able to calibrate PSY-1s

10/30

Calibrating
psychrometers
& placing on
leaves

software $\frac{1}{4}$
PSY #1 (No on body of logger)
2.281 solution = -2.98 / -3.01
0.642 solution = -0.74

pot on W: ~~W-6~~ W-6

PSY #4

- 0.642 solution = -0.38
- 2.281 solution =

Went into low power mode and
won't come out

PSY #2 can't get coms with 1st

PSY #3

- 0.642 solution =

Went into low power mode

NO setup taking