LAB ALPHA - Tensile Test Certificate

ISO 6892-1, method B, room temperature

Specimen ID: AL-001

Material: Al 6061-T6

Test date: 2026-09-14

Yield strength (Rp0.2): 276 MPa

Tensile strength (Rm): 311 MPa

Elongation at break (A): 12.4 %

Remarks: Fracture inside gauge length. Nothing to report.