


WALL-MOUNTED NOCOSPRAY® 2

With wall-crossing system

EN



IMP-PX-200
16/10/2020
Version B

OXYPHARM

829 rue Marcel Paul
94500 Champigny-sur-Marne
commercial@oxypharm.net
T: +33 1 45 18 78 70

www.oxypharm.net



I.	INTRODUCTION.....	3
I.1.	General information	3
I.2.	Classification / Regulation.....	3
I.3.	Warranty.....	3
I.3.1.	Clause	3
I.3.2.	Purpose	3
I.3.3.	Transport.....	3
I.3.4.	Duration	3
I.4.	Technical characteristics	4
II.	PRECAUTIONS FOR USE.....	4
III.	DESCRIPTION.....	6
III.1.	Wall-mounted NOCOSPRAY® 2 description.....	6
III.2.	Wall-mounted NOCOSPRAY® 2 dimensions	6
IV.	INSTRUCTIONS FOR USE	7
IV.1.	Positioning of the device.....	8
IV.2.	Installation of the bottle	8
IV.3.	Diffusion duration and consumption	9
IV.4.	Functioning in standard mode	10
IV.4.1.	Use without remote control.....	10
IV.4.2.	Use with remote control	11
IV.4.2.1.	Pairing with Nocospray (« blank » remote control).....	11
IV.4.2.2.	Functioning with the remote control	12
IV.4.2.3.	Pairing with another Nocospray (« re-pairing »)	13
IV.5.	Functioning in timer mode.....	14
IV.5.1.	General informations	14
IV.5.2.	Timer mode features	15
IV.5.3.	Back on standard mode.....	15
V.	TRACEABILITY SYSTEM	16
V.1.	Software installation.....	16
V.2.	Transferring treatments historical	17
V.3.	READING AND MODIFYING DATA	18
V.3.1.	General information	18
V.3.2.	General settings modification	19
V.3.3.	Export records (treatment historical)	19
V.3.4.	USB mode exit	19
VI.	TROUBLESHOOTING.....	20
VII.	HYGIENE AND MAINTENANCE	20
VIII.	EMC COMPLIANCE BOARD.....	21
IX.	RECYCLING	23

I. INTRODUCTION

I.1. General information

This user manual described the functioning of the wall-mounted NOCOSPRAY® 2 with wall-crossing system diffuser and involve essentials information ensuring the safe use.

The documentation is valid to the extent that the customer has not made any changes to the product.

Peruse this manual before any use of the NOCOSPRAY® 2 diffuser. It was written to ensure you master all functionalities.

I.2. Classification / Regulation

Electric classification : Class I

I.3. Warranty

The warranty certificate should be submitted to the dealer or to OXY'PHARM company, Z.A des Grands Godets, 829 rue Marcel Paul, 94500 Champigny sur Marne, France - to justify the warranty.

I.3.1. Clause

The following provisions are not exclusive to the benefit of the legal warranty provided by articles 1641 and following of the Civil code, relative to hidden defects.

I.3.2. Purpose

The warranty consists in the free supply of the parts recognized as defective by our technical services.

Repairs required as a result of a false manipulation, abnormal use, negligence or overload of the device, as well as those required by the variations of the electrical network, surges or faulty facilities, etc... cannot be taken under warranty.

In addition, if an abnormal corrosion of flexible parts, hose, seals... was found meaning the use of another liquid, the warranty would also be void.

I.3.3. Transport

The device travels at risk of the user: in case of damage during transport, the recipient must file a claim with the carrier before accepting delivery of the device.

The device must be stored in a dry place and transported in its original packaging.

I.3.4. Duration

The NOCOSPRAY® 2 diffuser is guaranteed for a period of 2 years, from date of invoice.

I.4. Technical characteristics

Voltage :	230V mono	115V mono
Power:	1000W	1000W
Frequency :	50 à 60 Hertz	50 à 60 Hertz
Nominal intensity:	4,5A	9A

Weight: 5,9Kg
Package weight: 8,3Kg
Compatible bottle: 1 Litre
Maximum room volume to be treated: 1000m³
Average flow of the liquid: 1000 ml/h
Environmental conditions:

In condition of use :

- Temperature between 0°C and 40°C
- Ambient humidity between 10% and 90%
- Atmospheric pressure of 800 to 1060 hPa

Transport and storage conditions :

- Temperature between 0°C and 70°C
- Ambient humidity between 10% and 95%
- Atmospheric pressure of 700 to 1060 hPa

II. PRECAUTIONS FOR USE

The manufacturer responsibility could under no circumstances be engaged for all material or corporal incident as a result of abnormal or faulty use of the device, opposing to recommended precautions in the none limitative list below:



Use only with OXY'PHARM product



- **EM device modification is forbidden.**
- To avoid risk of electric shock, this device must only be connected to a supply main with protective earth.
- Make sure that electric power matches regular standards.
- Hand-held RF communication devices may affect EM devices.
- This device may only use by professionals. According to requirements of norm NF EN 60601-1, it may cause radio interference and disturb operation of devices located close to the equipment.
- The use of accessories, transducers and wired other than those specified, with the exception of equipment sold by OXY'PHARM as replacement parts for internal components, may result in increased emissions or decreased immunity of the device.
- The device cannot be used close to other devices or piled. In case this is not possible, check the normal operating in configuration which it will be used.
- Never fill the bottles otherwise than with the product of the OXY'PHARM range.
- Do not introduce any object or liquid in the nose of NOCOSPRAY® 2.
- Check presence of nozzle and nozzle holder (cf. n°2 §III.1).
- **Remove the bottle from NOCOSPRAY® 2 for transportation.**
- **Any intervention must be carried out by staff trained on the system.**

Safety instructions

The device diffuses a chemical substance and makes an important sound power, therefore it is necessary to pay attention to following recommendations.

IMPORTANT RECOMMENDATIONS

EN



- Use only with NOCOTECH products.



- Always consult the accompanying documents.



- For indoor use only.



- Do not inhale the product.



- Do not direct the spray toward an incandescent source.



- Do not smoke close to the device.




- Grounded connections.



- No one should be in the room while the machine is operating
- Do not use in the presence of a flammable anesthetic mixture
- Handle the machine and bottle with care
- If a person must enter the room while the machine is running or during the required contact time, they must be equipped with individual protection equipment :
 - Noise-cancelling headphones.
 - Independent respirator.
 - Hydrogen peroxide resistant personal protective equipment.
- The individual should exit the room and seal it again as soon as possible.


III. DESCRIPTION

III.1. Wall-mounted NOCOSPRAY® 2 description




1	Main switch	6	Nose of the diffuser (venturi)
2	Fuses (5x20 10 ampères T)	7	Bottle
3	Power cable	8	Nozzle and nozzle holder
4	USB tracking	9	Control board
5	Ring	10	Filtering foam

III.2. Dimensions and assembly




The nozzle must protrude from the bulkhead.

YES



NO



IV. INSTRUCTIONS FOR USE

IV.1. Positioning of the device

- The NOCOSPRAY® 2 must be placed on a hard and flat surface.
- The nozzle must not be recessed in the partition (imperative).
- The thickness of the bulkhead must not be greater than 4 cm.
- For regular consumption, the bottle should be placed at 20 ° from the horizontal.

IMPORTANT RECOMMENDATION



Before starting a cycle, check that there is no obstacle to less than 2 meters from the device.

IV.2. Installation of the bottle



Follow instructions below to install the bottle.

Unscrew the bottle.	Screw the bottle (after having written date of opening on it), pay attention not to tighten it too tight, to prevent any irreversible damage on the seal.
---------------------	---


**Always check if the bottle contains enough product for the treatment
(the device will not indicate whether there is sufficient liquid)**

IV.3. Diffusion duration and consumption

Volume (m ³)	Consumption (ml)	Duration of diffusion
10	10	36 seconds
15	15	54 seconds
20	20	1 minute & 12 seconds
25	25	1 minute & 30 seconds
30	30	1 minute & 48 seconds
35	35	2 minutes & 6 seconds
40	40	2 minutes & 24 seconds
45	45	2 minutes & 42 seconds
50	50	3 minutes
62.5	62.5	3 minutes & 45 seconds
75	75	4 minutes & 30 seconds
87.5	87.5	5 minutes & 15 seconds
100	100	6 minutes
112.5	112.5	6 minutes & 45 seconds
125	125	7 minutes & 30 seconds
137.5	137.5	8 minutes & 15 seconds
150	150	9 minutes
162.5	162.5	9 minutes & 45 seconds
175	175	10 minutes & 30 seconds
187.5	187.5	11 minutes & 15 seconds
200	200	12 minutes
225	225	13 minutes & 30 seconds
250	250	15 minutes
275	275	16 minutes & 30 seconds
300	300	18 minutes
325	325	19 minutes & 30 seconds
350	350	21 minutes
375	375	22 minutes & 30 seconds
400	400	24 minutes
425	425	25 minutes & 30 seconds
450	450	27 minutes
475	475	28 minutes & 30 seconds
500	500	30 minutes
525	525	31 minutes & 30 seconds
550	550	33 minutes
575	575	34 minutes & 30 seconds
600	600	36 minutes
650	650	39 minutes
700	700	42 minutes
750	750	45 minutes
800	800	48 minutes
850	850	51 minutes
900	900	54 minutes
950	950	57 minutes
1000	1000	1 hour

IV.4. Functioning in standard mode

IV.4.1. Use without remote control



Connect the power cable and put the main switch on « I »
(The light switches on)




Select the volume by dragging your finger on the clear part of the touchpad. (The range is between 10m³ and 1000m³)



The light indicates the volume chose, if two lights are illuminated, the selected volume is the middle point between the two values.



After selecting the volume, press and hold the start button in the middle of the touchpad to start the diffusion cycle. A beeping sound will begin.



A beeping sound will continue for 15seconds and then the diffusion will begin. Leave and close off the room before the end of these 15 seconds.



The device can be stopped by pressing the start button again (device beeps when treatment ends).





IV.4.2. Use with remote control

The remote control is used to start the device from outside the room, replacing the start button on control touch pad

IV.4.2.1. Pairing with Nocospray (« blank » remote control)


When receiving the remote control and before first use, it needs to be paired with Nocospray® appliance.

Pairing method :


		
<p>Nocospray needs to be turned off but plugged in</p>	<p>Press remote control button (+/- 5 seconds) remaining very close to the Nocospray device that needs to be paired until the Led flashes</p>	<p>While maintaining pressure on the remote control button, switch the Nocospray that needs to be paired on (both actions at the same time)</p>
		
<p>Remote control Led starts flashing quickly and the paired Nocospray beeps several times to signal that pairing was successful</p>		

IV.4.2.2. Functioning with the remote control

EN



Connect the power cable and put the main switch on « I »
(The light switches on)




Select the volume by dragging your finger on the clear part of the touchpad. (The range is between 10m^3 and 1000m^3)




The light indicates the volume chose, if two lights are illuminated, the selected volume is the middle point between the two values.



Press the remote control button to start the treatment.



A beeping sound will continue for 15 seconds and then the diffusion will begin.
Leave and close off the room before the end of these 15 seconds.



The device can be stopped by pressing the start button again (device beeps when treatment ends).







IV.4.2.3. Pairing with another Nocospray (« re-pairing »)



The remote control can be paired to only one Nocospray® at a time.

However, a remote control already paired can be 're-paired' to another Nocospray®









Re-pairing method :

		
<p>Check that all Nocospray close to the remote control are well switched off including the one to be re-paired</p>	<p>Plug in the Nocospray to be re-paired.</p>	<p>Press long on remote control button (+/- 30 secondes) remaining very close to the Nocospray device that needs to be re-paired until the Led flashes</p>
		
<p>During these 30 seconds, the Led light flashes and stops alternately. This phase is repeated 3 times</p>	<p>Once the Led light on the remote control flashes regularly and without stopping, switch the Nocospray that needs to be re-paired on, while maintaining pressure on the remote control button (both actions at the same time)</p>	<p>The Led light on the remote control starts flashing quickly and Nocospray beeps several times to signal that pairing was successful</p>

IV.5. Functioning in timer mode

IV.5.1. General informations

Plugged into Nocospray® device and on mains socket, the timer enables postponing or repetition of treatments. Timer mode should be **selected to use Nocospray® with a timer**.

		
<p>Connect the power cable and put the main switch on « 1 »</p>	<p>To initiate timer mode, drag along the clear part of touchpad with your finger to the 1000m³ setting. Then continue in clockwise direction for 3 full rotations until three (3) lights are illuminated</p>	<p>Timer mode is on when 3 Led lights light up as follows :</p> <ul style="list-style-type: none"> - Led 100m³ and 500m³ light up - Led 250m³ flashes.
		
<p>Select the volume. As the selection is progressive, if two LEDs are lit, it is the intermediate value which is selected. Note: selected volume appears only when the keyboard is touched/pressed</p>	<p>Put the main switch on « 0 »</p>	<p>Plug the Nocospray socket into the timer and the timer in the mains socket</p>
		
<p>Set the timer, referring to the specific instructions for use. Nb : it can be set before it is plugged in</p>	<p>The timer will start and stop the Nocospray according to the setting</p>	


IV.5.2. Timer mode features




If Nocospray is on timer mode when you switch it on, you can modify the volume or set the device back on standard mode **during 1 minute.**

After this minute, Nocospray starts beeping for 15 seconds before starting treatment.

When treatment is over, it is strongly advised to set the device back standard before switching it off.




When treatment is over, volume can be changed as long as the appliance is still switched on



The device cannot be stopped by pressing the start button on touchpad.

IV.5.3. Back on standard mode



To set the appliance back on standard mode, touch the pad and drag your finger counterclock-wise back to the 10m3 position and then continue for an additional 3 full rotations

V. TRACEABILITY SYSTEM

V.1. Software installation


Click on “NocoSprayTracker_Setup.exe” on the CD ROM or USB key supplied. Allow the program to install the software if Windows asks it.

EN


1°) Click on « Next »




2°) Click again on « Next »




3°) Click on « Install »




4°) Click on « Finish ».



If the following window appears, click on « Fermer »





Nocospray Tracker software runs:



EN


V.2. Transferring treatments historical

A mini USB connector enables transferring the treatments historical to a computer.

		<p>No treatment available in USB mode.</p> <p>To exit USB mode, unplug the cable from the computer</p>
<p>Plug the USB cable stick</p>	<p>The 10m3 and 1000m3 LEDs flash and the 250 m3 lights up.</p>	


Connect Nocospray 2 on the USB port of your computer:

An icon on bottom right might appear to indicate that the device driver is running.



If your computer doesn't allow an automatic installation, the driver « .inf » is provided with the software, please contact your system manager.

Mention « Connected to COM4» on bottom left of the page, shows that Nocospray is well connected.



V.3. READING AND MODIFYING DATA

V.3.1.General information

Nocospray tracker software will launch automatically once installed. If it does not it can be initiated manually.

- « Device Name »: Name of the device (editable parameter)
- « Device date »: Date and time of the device (to be set the first time – it will automatically update afterwards)
- « Software version »: Version of Nocospray Tracker software
- « Device version »: Version of the intern software of Nocospray



Each line indicates each treatment. It is possible to organize each treatment following its number, date, time, volume, duration, room name, operator name or comment.

« Treatment »: treatment number

« Date »: Date of the treatment

« Time »: Hour of the treatment

« Volume »: Volume setting of the Nocospray

« Duration »: Duration of the treatment (this parameter is automatically calculated)

« Operator name »: Name of the operator who made the treatment.

« Room »: Name of the room where treatment has been launched


« Comment »: Other comment

: Enter parameters « Operator name », « Room » and « Comment »

WARNING: Once those parameters are saved, they cannot be modified

V.3.2.General settings modification

This icon gives you access to the following window:



Device settings

Device Name : Nocospray V2

Device date

	lun.	mar.	mer.	jeu.	ven.	sam.	dim.
5	30	31	1	2	3	4	5
6	6	7	8	9	10	11	12
7	13	14	15	16	17	18	19
8	20	21	22	23	24	25	26
9	27	28	1	2	3	4	5
10	6	7	8	9	10	11	12

Device hour 23 : 55


Cancel Save

On this window device name « Device Name », date and hour of the device can be modified.


V.3.3.Export records (treatment historical)




This icon enables exporting the all recorded treatments to a PDF file as long as they have not been deleted. It brings you to the following window :




PDF preview

PDF folder :  PDF Name :

 Created on : 27/02/2017

Treatment	Date	Time	Volume	Duration	Operator name	Room	Comment
-----------	------	------	--------	----------	---------------	------	---------

Cancel Export

Using this icon , you can choose the folder where you want save the data (PDF file).

« PDF Name » : enter the name of the record file.

Then, click to « export » to record file.

V.3.4.USB mode exit

To exit USB mode, unplug the cable from the computer.

VI. TROUBLESHOOTING

The table below shows a list of common problems and likely solutions. If the problem persists, contact your reseller or our Customer Service Center.

EN



Before any intervention, switch off the device and disconnect the power cord

Failure	Resolution
The device does not switch on	<ul style="list-style-type: none">- Check the power cable connection- Check if the switch is in « I » position and if the light is on- Check fuses
The device works but the consumption is incorrect	<ul style="list-style-type: none">- Check that nozzle is attached securely, is in working, and not blocked- Check if the bottle contains enough liquid
The device has fell and has a malfunction	Contact the Customer Service Center

VII. HYGIENE AND MAINTENANCE


The NOCOSPRAY® 2 device requires regular maintenance.



- **Switch off the device before proceeding any cleaning.**
- **Do not use abrasive products, bleach, acetone or solvents for maintenance or cleaning operations of the device. These aggressive agents may damage the surface materials (discoloration, tasks, drips, microcracks...)**
- **For the cleaning of external surfaces, use a slightly damp cloth or wipes.**

VIII. EMC COMPLIANCE BOARD

EMC Compliance according to IEC 60601-1-2 (2007)			
The NOCOSPRAY 2 is intended for use in the electromagnetic environment specified below. The customer or the user of the NOCOSPRAY 2 should assure that it is used in such an environment.			
Emission test	Standard	Compliance	Electromagnetic environment
Emissions RF	CISPR 11	Group 1	The NOCOSPRAY 2 uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. The NOCOSPRAY 2 is suitable for use in all establishments, including domestic establishments and those directly connected to the public low voltage power supply network that supplies buildings used for domestic purposes.
Emissions RF	CISPR 11	Class B	
Harmonics emissions	IEC 61000-3-2	N/A	
Voltage/Flicker test	IEC 61000-3-3	N/A	
The NOCOSPRAY 2 is intended for use in the electromagnetic environment specified below. The customer or the user of the NOCOSPRAY 2 should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment
Electrostatic discharges IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical Fast Transients IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV line(s) to line(s) ±2 kV line(s) to earth	±1 kV line(s) to line(s) ±2 kV line(s) to earth	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5 % U_T (>95 % dip in U_T) for 0,5 cycle 40 % U_T (60 % dip in U_T) for 5 cycles 70 % U_T (30 % dip in U_T) for 25 cycles <5 % U_T (>95 % dip in U_T) for 5 s	<5 % U_T (>95 % dip in U_T) for 0,5 cycle 40 % U_T (60 % dip in U_T) for 5 cycles 70 % U_T (30 % dip in U_T) for 25 cycles <5 % U_T (>95 % dip in U_T) for 5 s	Mains power quality should be that of a typical commercial or hospital environment. If the user of the system requires continued operation during power mains interruptions, it is recommended that the system be powered from an uninterruptible power supply or a battery. U_T is the a.c. mains voltage prior to application of the test level.
Magnetic field at power frequency IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

Conducted RF IEC 61000-4-6	3 Vrms 150kHz-80MHz	3 Vrms	Portable and mobile RF communications equipment should be used no closer to any part of the NOCOSPRA Y 2, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance : $d = 1,67 \cdot \sqrt{P}$
Radiated RF IEC 61000-4-3	3V/m 80MHz-2.5GHz	3V/m	$d = 1,67 \cdot \sqrt{P} \text{ 80MHz-800MHz}$ $d = 2,33 \cdot \sqrt{P} \text{ 800MHz-2,5GHz}$ <p>where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey (a) should be less than the compliance level in each frequency range (b)</p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p> 
NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.			
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			
(a) Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the NOCOSPRA Y 2 is used exceeds the applicable RF compliance level above, the NOCOSPRA Y 2 should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the NOCOSPRA Y 2.			
(b) Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3V/m.			

Recommended separation distances between portable and mobile RF communications equipment and the NOCOSPRA Y 2.			
The NOCOSPRA Y 2 is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the NOCOSPRA Y 2 can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the NOCOSPRA Y 2 as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitter (m)		
	150kHz to 80MHz	80MHz to 800MHz	800MHz to 2.5GHz
0.01	0.117	0.117	0.233
0.1	0.369	0.369	0.737
1	1.167	1.167	2.330
10	3.690	3.690	7.368
100	11.67	11.67	23.300
For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.			
NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.			
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

IX. RECYCLING

IMPORTANT INFORMATION ABOUT THE ELIMINATION COMPATIBLE WITH THE ENVIRONMENT

This device is subject to separate collection for electrical and electronic equipment in accordance with Directive 2002/96 / EC (applicable in European Union countries and other European countries with separate collection systems).



This symbol that appears on the device indicates that at the end of its life cycle, the device will need to be treated separately from household waste.

The device will be assigned to a center of selective collection for electrical and electronic appliances, or reported to the dealer on the purchase of a new device.

The user is responsible for the delivery of the device at the end of life to the collection authorities or to a corporation duly empowered under penalty of sanctions provided for by the legislation on the elimination of waste. The waste supplies must be managed in compliance with the local regulations relating to the management of waste.

Proper differential collection, and the subsequent recycling, processing and environmentally compatible disposable of waste equipment avoids unnecessary damage to the environment and possible related health risks and also promotes recycling of the materials used in the appliance