# Protocol: Real-time RT-PCR assays for the detection of SARS-CoV-2 Institut Pasteur, Paris

This protocol describes procedures for the detection of SARS-CoV-2 for two RdRp targets (IP2 and IP4).

Based on the first sequences of SARS-CoV-2 made available on the GISAID database on January 11, 2020, primers and probes (nCoV\_IP2 and nCoV\_IP4) were designed to target the RdRp gene spanning nt 12621-12727 and 14010-14116 (positions according SARS-CoV, NC 004718).

As a confirmatory assay, we used the E gene assay from the Charité protocol<sup>1</sup>

## **Material**

## Kits:

Kit Extraction NucleoSpin Dx Virus Ref: Macherey Nagel 740895.50 SuperScript™ III Platinum® One-Step Quantitative RT-PCR System Ref: Invitrogen 1732-020

**Primers and probes** 

Filliers and probes				
Name	Sequences (5'-3')	Length (bases)	PCR product size	Ref.
RdRp gene / nCoV_IP2				
nCoV_IP2-12669Fw	ATGAGCTTAGTCCTGTTG	17		
nCoV_IP2-12759Rv	стсссттгаттаттат	18	108 bp	1
nCoV_IP2-12696bProbe(+)	AGATGTCTTGTGCTGCCGGTA [5']Hex [3']BHQ-1	21		
RdRp gene / nCoV_IP4				
nCoV_IP4-14059Fw	GGTAACTGGTATGATTTCG	19		
nCoV_IP4-14146Rv	CTGGTCAAGGTTAATATAGG	20	107 bp	1
nCoV_IP4-14084Probe(+)	TCATACAAACCACGCCAGG [5']Fam [3']BHQ-1	19		
E gene / E_Sarbeco				
E_Sarbeco_F1	ACAGGTACGTTAATAGTTAATAGCGT	18		
E_Sarbeco_R2	ATATTGCAGCAGTACGCACACA	20	125 bp	2
E_Sarbeco_P1	ACACTAGCCATCCTTACTGCGCTTCG [5']Fam [3']BHQ-1	20		

<sup>1/</sup> National Reference Center for Respiratory Viruses, Institut Pasteur, Paris.

Primer sets nCoV\_IP2 and nCoV\_IP4 can be multiplexed. Both reaction mixtures are described below.

PCR amplification regions (positions according to SARS-CoV, NC\_004718)

nCoV\_IP2 / 12621-12727 nCoV\_IP4 / 14010-14116 E gene / 26141-26253

## **NUCLEIC ACID EXTRACTION**

RNA is extracted from specimens using the NucleoSpin Dx Virus (Macherey Nagel ref. 740895.50). RNA extracted from 100  $\mu$ l of original sample, is eluted in 100  $\mu$ l of elution buffer.

<sup>2/</sup> Corman et al. Eurosurveillance<sup>1</sup>

# MIX PREPARATION FOR ALL SEPARATE PRIMER/PROBE COMBINATIONS

All primers and probes described below were validated under the following conditions.

## RT-PCR Mix kit:

Invitrogen Superscript™ III Platinum® One-Step gRT-PCR system (ref: 11732-088)

## Real-time PCR equipment:

• LightCycler 480 (96)

Adjustments may be required for the use of other kits or other real-time PCR instruments. All Assays used the same conditions. Primer and probe sequences, as well as optimized concentrations are shown in table below. A 25µl reaction was set up containing 5µl of RNA.

Simplex Mix	Vol (μl)	[final]
H <sub>2</sub> O PPI	3.60	
Reaction mix 2X	12.50	3 mM Mg
MgSO4 (50mM)	0.40	0.8 mM Mg
Forward Primer (10µM)	1.00	0.4 μΜ
Reverse Primer (10μM)	1.00	0.4 μΜ
Probe (10μM)	0.50	0.2 μΜ
SuperscriptIII RT/Platinum Taq Mix	1.00	
Final Volume	20.00	

Multiplex Mix (nCoV_IP2&IP4)	Vol (μl)	[final]
H <sub>2</sub> O PPI	1.3	
Reaction mix 2X	12.50	3 mM Mg
MgSO4 (50mM)	0.40	0.8 mM Mg
Forward Primer (10μM)	1.00	0.4 μΜ
Reverse Primer (10µM)	1.00	0.4 μΜ
Forward Primer (10μM)	1.00	0.4 μΜ
Reverse Primer (10µM)	1.00	0.4 μΜ
Probe (10μM)	0.4	0.16 μΜ
Probe (10μM)	0.4	0.16 μΜ
SuperscriptIII RT/Platinum Taq Mix	1.00	
Final Volum	e 20.00	

## **CONTROLS**

Each real-time RT-PCR assay includes in addition of unknown samples:

- Two negative samples bracketing unknown samples during RNA extraction (negative extraction controls)
- Positive controls (in duplicate); when using in vitro synthesized transcripts as controls include five quantification positive controls (in duplicate) including 10<sup>5</sup>, 10<sup>4</sup> and 10<sup>3</sup> copies genome equivalent (ge) of in vitro synthesized RNA transcripts.
- One negative amplification control.

# **AMPLIFICATION CYCLES (LIGHTCYCLER SYSTEM)**

Reverse transcription	55°C	20 min	x1	
Denaturation	95°C	3 min	x1	
Amplification	95°C	15 sec	x50	Acquisition
	58°C	30 sec	XSU	
Cooling	40°C	30 sec	x1	

## **SENSITIVITY**

## For the nCoV\_IP and E\_Sarbeco real-time RT-PCR

Sensitivity, in terms of 95% hit rate is about 100 copies of RNA genome equivalent per reaction (this amount of target sequences is always detected), the probability to detect lower amounts of virus decreases, but samples containing 10 copies could be detected with multiplex assay.

	Mult (Ct va	Simplex (Ct values)	
RNA copies Of transcript	nCoV_IP2	nCoV_IP4	E_Sarbeco
1,00E+07	21,67	21,97	24,72
1,00E+06	24,97	25,12	28,19
1,00E+05	28,00	27,88	30,96
1,00E+04	31,84	30,51	33,33

Ct values may vary from instrument to instrument by up to 2 cycles, while the interval between two dilutions steps is constant ( $\Delta$ Ct).

#### **SPECIFICITY**

Cross-reactivity with other respiratory viruses was tested with specimens known to be positive for a panel of respiratory viruses (influenza A(H1N1)pdm09, A(H3N2), B-Victoria, B-Yamagata; influenza C; RSV A, B; hBoV; hPIV; hMPV; HRV/enterovirus; adenovirus; hCoV (HKU1, OC43, 229E and NL63); MERS-CoV. None of the tested viruses showed reactivity with PCR2 and PCR4.

## POSITIVE CONTROL FOR SARS-CoV-2 REAL-TIME RT-PCR

One specific control has been designated.

Positive control for real-time RT-PCR is an *in vitro* transcribed RNA derived from strain BetaCoV\_Wuhan\_WIV04\_2019 (EPI\_ISL\_402124). The transcript contains the amplification regions of the *RdRp* and *E gene* as positive strand. Each microtube contains 10<sup>11</sup> copies of target sequences diluted in yeast tRNA, and lyophilised.

## Reconstitution of transcribed RNA

Add 100  $\mu$ l of RNase/DNAse-free H2O to obtain a solution at a concentration of  $10^9$  copies/ $\mu$ l. Store at -80°C. Dilute to prepare a master bank at  $2x10^6$  copies/ $\mu$ l. Store at -80°C.

From this prepare a working bank of reagent at  $2x10^4$  copies/ $\mu$ l in order to avoid freeze/thaw cycles. Working tubes may be stored at -20°C for less than one week.

Positive controls are available upon request (grippe@pasteur.fr)

## Aknowledgements

We gratefully acknowledge the Authors, the Originating and Submitting Laboratories for their sequence and metadata shared through GISAID (EPI\_ISL\_402119; EPI\_ISL\_402121; EPI\_ISL\_402120; EPI\_ISL\_402123; EPI\_ISL\_402124; EPI\_ISL\_402125).

## Reference

1- Corman VM, Landt O, Kaiser M, et al. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. Euro Surveill 2020;25.