BIOMEDICAL
INTERNACIONAL
IMPORTADORA DE INSUMOS Y EQUIPO MÉDICO

FORMULARIO C-1
ESPECIFICACIONES TÉCNICAS

Para ser llenado por la Entidad convocante *(Llenar las Especificaciones Técnicas de manera previa a la publicación del DBC)*		Para ser llenado por el proponente al momento de elaborar su propuesta
#	Características y condiciones técnicas solicitadas (*)	Característica Propuesta (**)
	ESPECIFICACIONES TECNICAS	TENSIOMETRO DIGITAL ADULTO
	ITEM / CANTIDAD / UNIDAD DE MANEJO / DETALLE DEL BIEN 1 / 1 / PZA / TENSIOMETRO • Tensiómetro de brazo automático con guía de ajuste del brazalete, detector de latido irregular, brazalete universal (22-42 cm). **MARCA:** ESPECIFICAR **PROCEDENCIA:** ESPECIFICAR •	MARCA: ADC PROCEDENCIA HECHO EN CHINA ENSAMBLADO EN EE.UU. MODELO: 6021N CANTIDAD 1 PIEZA - TENSIOMETRO DE BRASO AUTOMATICO CON GUIA DE AJUSTE DEL BRAZALETE, DETECTOR DE LATIDO IRREGULAR, BRAZALETE UNIVERSAL 22CM -42 CM - 4 PILAS "AA"

Katty Roxana Lilia Guzmán Quiroga
REPRESENTANTE LEGAL

BIOMEDICAL INTERNACIONAL
NIT: 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

FORMULARIO C-1
ESPECIFICACIONES TÉCNICAS

Para ser llenado por la Entidad convocante *(Llenar las Especificaciones Técnicas de manera previa a la publicación del DBC)*				Para ser llenado por el proponente al momento de elaborar su propuesta
#	Características y condiciones técnicas solicitadas (*)			Característica Propuesta (**)

ESPECIFICACIONES TECNICAS

ITEM	CANTIDAD	UNIDAD DE MANEJO	DETALLE DEL BIEN
2	1	PZA	OXIMETRO DE PULDO

MARCA: ESPECIFICAR

MODELO: ESPECIFICAR

PAÍS DE FABRICACIÓN: ESPECIFICAR

- OXIMETRO DE PULSO CON PANTALLA ADULTO Y 2 PILAS "AAA"

TENSIOMETRO ANEROIDE CON PEDESTAL ADULTO

MARCA:
ADC
PROCEDENCIA:
HECHO EN CHINA ENSAMBLADO EN EE.UU.
MODELO:
2200
CANTIDAD:
1 PIEZA

- OXIMETRO DE PULSO CON PANTALLA
- LEE SPO2 Y FRECUENCIA DEL PULSO
- PRESION : SPO2 ± 2% / HR ± 5 BPM
- 2 MODOS DE VISUALIZACION CAMBIABLES
- PANTALLA LED ROJA
- LLEVA CORTDON
- 2 PILAS "AAA"

Katty Roxana Lilia Guzmán Quiroga
REPRESENTANTE LEGAL

BIOMEDICAL
INTERNACIONAL
NIT: 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

BIOMEDICAL
INTERNACIONAL
IMPORTADORA DE INSUMOS Y EQUIPO MÉDICO

FORMULARIO C-1
ESPECIFICACIONES TÉCNICAS

Para ser llenado por la Entidad convocante *(Llenar las Especificaciones Técnicas de manera previa a la publicación del DBC)*		Para ser llenado por el proponente al momento de elaborar su propuesta
#	Características y condiciones técnicas solicitadas (*)	Característica Propuesta (**)
	ESPECIFICACIONES TECNICAS	**NEGATOSCOPIO DE DOS CUERPOS**

ITEM	CANTIDAD	UNIDAD DE MANEJO	DETALLE DEL BIEN
3	1	PZA	ESTETOSCOPIO DOBLE CAMPANA

- ALTA SENSIBILIDAD ACUSTICA PARA LA REALIZACION DE VALORACIONES GENERALES.
- DIAFRAGMA DE DOBLE FRECUENCIA EN EL LADO PEDIATRICO Y EN EL LADO ADULTO

MARCA:
ADC
PROCEDENCIA:
603 BK
CANTIDAD:
1 PIEZAS

- ALTA SENSIBILIDAD ACUSTICA PARA LA REALIZACION DE VALORACIONES GENERALES.
- DIAFRAGMA DE DOBLE FRECUENCIA EN EL LADO PEDIATRICO Y EN EL LADO ADULTO
- EL TUBO ES SUAVE Y FLEXIBLE

Katty Roxana Lilia Guzmán Quiroga
REPRESENTANTE LEGAL

BIOMEDICAL
INTERNACIONAL
NIT: 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

BIOMEDICAL
INTERNACIONAL
IMPORTADORA DE INSUMOS Y EQUIPO MÉDICO

FORMULARIO DE PRESENTACIÓN DE PROPUESTA

NOMBRE DE LA EMPRESA:	Nº de NIT o C.I.:
BIOMEDICAL INTERNACIONAL	2723410014
NOMBRE DEL REPRESENTANTE LEGAL O PROPIETARIO:	**Nº Telefono de Contacto:**
KATTY ROXANA LILIA GUZMAN QUIROGA	4522784

Nota: Poner unicamente el precio unitario ofertado en la columna "Precio Unitario", de los ítems ofertados por su empresa/persona.

Nº de Ítem	CONCEPTO	UNIDAD DE MEDIDA	Cantidad	Precio Unitario	Precio Total Ofertado
001	TENSIOMETRO	PIEZA	1	390,00	390,00
002	OXIMETRO	PIEZA	1	400,00	400,00
003	ESTETOSCOPIO	PIEZA	1	350,00	350,00
	TOTAL				1.140,00

DRA. KATTY ROXANA LILIA GUZMAN QUIROGA
GERENTE GENERAL

BIOMEDICAL
INTERNACIONAL
NIT: 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

Apostille

(Convention de La Haye du 5 Octobre 1961)

1. Country: United States of America

This public document

2. has been signed by **Milton Adair Tingling**

3. acting in the capacity of **County Clerk**

4. bears the seal/stamp of the **county of New York**

Certified

5. at New York City, New York 6. the 21st day of December 2020

7. by Deputy Secretary of State for Business and Licensing Services, State of New York

8. No. NYC-1522834

9. Seal/Stamp 10. Signature

BIOMEDICAL INTERNACIONAL
NIT: 2223410014 C/ Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

Whitney A. Clark

Whitney A. Clark

Deputy Secretary of State for Business and Licensing Services

STATE OF NEW YORK ★ DEPARTMENT OF STATE ★ EXCELSIOR

Apostille (REV: 09/25/12)

Form 1

State of New York
County of New York } *ss:*

No. 685202

I, Milton Adair Tingling, Clerk of the County of New York, and Clerk of the Supreme Court in and for said county, the same being a court of record having a seal, **DO HEREBY CERTIFY THAT**

GUILLERMO COBO

whose name is subscribed to the annexed original instrument has been commissioned and qualified as a NOTARY PUBLIC..... and has filed his/her original signature in this office and that he/she was at the time of taking such proof or acknowledgment or oath duly authorized by the laws of the State of New York to take the same: that he/she is well acquainted with the handwriting of such public officer or has compared the signature on the certificate of proof or acknowledgment or oath with the original signature filed in his/her office by such public officer and he/she believes that the signature on the original instrument is genuine.

IN WITNESS WHEREOF, I have hereunto set my hand and my official seal this 10th day of December, 2020

BIOMEDICAL INTERNACIONAL
NIT: 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

Milton Adair Tingling

County Clerk, New York County

COUNTY CLERK
NEW YORK COUNTY

ás,
os
de
es
la
o,
20
erce Drive
, NY 11788
)0-ADC-2670
31-273-9600
31-273-9659
dctoday.com
dctoday.com
EI • TOKYO

ERICAN DIAGNOSTIC CORPORATION

Carta de Autorización de Distribución

Lunes, 7 de diciembre de 2020

De nuestra consideración:

Mediante la presente, el fabricante **American Diagnostic Corporation (ADC) localizado en 55 Commerce Drive, Hauppauge, NY 11788 USA** certifica que:

La Empresa **BIOMEDICAL INT/KATTY GUZMAN, AVENIDA PAPA PAULO #775 ZONA RECOLETA COCHABAMBA, BOLIVIA,** es uno de nuestros distribuidores, no exclusivo autorizado a distribuir productos que fabricamos tales como estetoscopios, tensiómetros, termómetros, juegos de diagnóstico, accesorios, laringoscopios, etc. Y está autorizada a representarnos en la venta de nuestros productos y equipos tanto en compra directa como en licitaciones del sector público y privado.

Declaramos que los productos fabricados en nuestra planta estetoscopios, esfigmómetros, y equipos de diagnóstico son elaborados bajos estrictos, controles de conformidad, con todas las disposiciones pertinentes de la directiva del consejo de la FDA, sobre dispositivos médicos. Todo el apoyo a la información se mantiene bajo el control del fabricante.

Nuestros dispositivos médicos cumplen con los principios esenciales de seguridad y efectividad. Además, nos comprometemos a notificar a la autoridad de salud todos, los reportes de incidentes o eventos adversos, retiros voluntarios del mercado de dispositivos médicos del que tengamos conocimiento.

Debidamente autorizado para firmar la autorización por y en nombre de **American Diagnostic Corporation.**

Esta autorización es válida hasta diciembre 31, 2022.

Sinceramente,

Thelma Valle
Latin America Sales Manager
American Diagnostic Corporation

STATE OF NEW YORK
COUNTY OF NEW YORK }
SWORN TO BEFORE ME THIS
9th day of Dec. 2020

BIOMEDICAL INTERNACIONAL
NIT 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

GUILLERMO COBO
Notary Public, State of New York
No. 01CO6033970
Qualified in New York County
Commission Expires Dec. 6, 2021

55 Commerce Drive
Hauppauge, NY 11788
Customer Service: 1-800-ADC-2670
Voice: 1-631-273-9600
Fax: 1-631-273-9659
Email: info@adctoday.com
http://www.adctoday.com
NEW YORK • LONDON • TAIPEI • TOKYO

American Diagnostic Corporation



Tipo:	Punta del dedo
Rango de dedo:	8,0 mm-20,0 mm (0,31 "-0,79")
Rango de SpO2:	70-100%
Rango PR:	30-250 lpm
Modo de visualización PR:	Gráfico de barras
Precisión de SpO2:	'+/- 2% @ 80-99%, +/- 3% @ 70-80%
Monitor:	DIRIGIÓ
Precisión de relaciones públicas:	+/- 2 lpm
Requerimientos de energía:	dos pilas alcalinas AAA
El consumo de energía:	40 mA
Duración de la batería: 18	horas (continuo)
Dimensiones:	2.19 "x1.13" x1 (55x29x25 mm)

Características del 2200:

- Lee SPO2 y frecuencia de pulso
- Precisión: SPO2 ± 2% / HR ± 5 BPM
- 2 modos de visualización cambiables
- Pantalla LED roja brillante
- Se adapta a dedos con un grosor de 8 mm a 20 mm (0,31 "-0,79")
- Llevar cordón
- 2 pilas "AAA"

BIOMEDICAL INTERNACIONAL
NIT: 2172341001 4 C. Papa paulo # 775
Tel.: 4522784 - 4522785
Cochabamba - Bolivia

American Diagnostic Corporation



Las características 6021N:

- Mide la presión sistólica, diastólica y el pulso
- La clasificación de la British Hypertension Society (BHS) AA garantiza una precisión sin precedentes
- La tecnología Smart Logic ™ determina el nivel ideal de inflación
- Detector de latidos cardíacos irregulares
- Control automático de inflación y deflación
- Recuerda las últimas 99 lecturas con fecha y hora para dos usuarios
- El brazalete de D-Barra libre de látex libre de ajuste automático se adapta a rangos de brazos de 22 cm a 42 cm (8,7 "a 16,5")
- Apagado automático
- No hecho con látex de caucho natural
- 4 pilas "AA" incluidas
- Estuche Deluxe de almacenamiento
- Adaptador de CA opcional (se vende por separado)
- Cumple o excede EN 1060-1, y EN 1060-3

BIOMEDICAL
INTERNACIONAL
NIT: 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

American Diagnostic Corporation

ADSCOPE ® 603

Estetoscopio médico

Tipo Pechera:	Combinación
Pechera Material:	Inoxidable
Pechera Acabado:	Satín
Pechera Peso:	3.1 oz
Pechera Peso (g):	87,9 g
El diafragma Diámetro:	1-3 / 4"
El diafragma Diámetro (cm):	4,45 cm
Material del diafragma:	Epoxi / fibra de vidrio
Tipo Diafragma:	Tecnología AFD
Campana Diámetro:	1-3 / 8"
Campana Diámetro (cm):	3,5 cm
Tipo de auricular:	clínico
Material del audífono:	Inoxidable
Auricular Acabado:	Satín
Tubería Longitud:	21"
La longitud del tubo (cm):	53 cm
Largo total:	31"

BIOMEDICAL INTERNACIONAL
NIT: 2123410014 C/ Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

Los ADSCOPE ® 603 características de la serie:

- Combinación de diafragma / campana campana adulto mecanizado con precisión a tolerancias exactas de acero inoxidable quirúrgico para un rendimiento excepcional y durabilidad
- AFD tecnología de ADC proporciona una respuesta acústico mejorado
- Una ligera presión para, de baja frecuencia de tipo campana, una presión firme para alta, frecuencia del tipo de diafragma
- campana grande Extra (1 3/8" de diámetro) para la respuesta de baja frecuencia sin igual
- diafragma Ultra-sensibles (1 3/4" de diámetro) para una mayor amplificación y respuesta de alta frecuencia más nítidas
- Protección contra el frío de campana y de retención de diafragma llantas para la comodidad del paciente
- auricular Clínico cuenta con yugo de refuerzo moldeado en los flexibles binaurales 22" tubos de PVC y acero inoxidable fijos a 15° de ángulo para una mayor comodidad (modelo táctico con binaurales de aluminio)
- Tubing es suave y flexible y se ha realizado sin ftalatos desde 2014
- Adsoft ™ olivas de silicona Plus con nueva unión de encaje elástico para la máxima comodidad de uso y sello acústico
- Incluye kit de accesorios con dos pares de repuesto Adsoft ™ Plus olivas
- etiqueta de identificación Alcance incluido
- Pesa 5.8 oz (164 gr)
- Longitud total 31"
- Ensamblado, inspeccionado y empaquetado en los EE.UU.

CERTIFICATE OF REGISTRATION

American Diagnostic Corporation

55 Commerce Drive
Hauppauge
NY 11788 UNITED STATES

UL LLC®(UL) issues this certificate to the Firm named above, after assessing the Firm's quality system and finding it in compliance with:

ISO 13485:2016
EN ISO 13485:2016

Manufacture and distribution of digital sphygmomanometers, sphygmomanometer cuffs, sphygmomanometers, stethoscopes, electronic stethoscopes. The distribution of digital thermometers, infrared thermometers, diagnostic instruments, otoscopes, ophthalmoscopes, tuning forks, laryngoscopes, laryngoscope blades, bite sticks, neurological hammers, surgical instruments, valve-mask/face shield CPR products, ring cutters, penlights, disposable airways, tourniquets, shears and pulse oximeters.

UKAS MANAGEMENT SYSTEMS
4426

Authorized by

Deborah Jennings-Conner
Global Regulatory Director
Life and Health Sciences, UL LLC

Check Certificate Status: here

BIOMEDICAL INTERNACIONAL
NIT 272341001 4 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

File Number	A28888	Cycle Start	March 5, 2021
Certificate Number	3193.210305	Effective Date	March 5, 2021
Initial Issue Date	October 30, 2020	Expiry Date	March 4, 2024

This quality system registration is included in UL's Directory of Registered Firms and applies to the provision of goods and/or services as specified in the scope of registration from the address(es) shown above. By issuance of this certificate the firm represents that it will maintain its registration in accordance with the applicable requirements. This certificate is not transferable and remains the property of UL LLC.

UL REGISTERED FIRM

UL LLC
333 Pfingsten Road
Northbrook, IL 60062-2096 USA

UL and the UL logo are trademarks of Underwriters Laboratories Inc. © 2018

CERTIFICATE OF REGISTRATION

American Diagnostic Corporation

55 Commerce Drive
Hauppauge
NY 11788 UNITED STATES

REPs Facility ID: **F005182**

Additional Regulatory Requirements

Australia:
- Therapeutic Goods (Medical Devices) Regulations, 2002, Schedule 3 Part 1 (excluding Part 1.6) – Full Quality Assurance Procedure

Canada:
- Medical Devices Regulations – Part 1- SOR 98/282

Japan:
- MHLW Ministerial Ordinance 169, Article 4 to Article 68
- PMD Act (,as applicable)

United States:
- 21 CFR 820
- 21 CFR 803
- 21 CFR 806
- 21 CFR 807 – Subparts A to D
- 21 CFR 821 (where applicable)

BIOMEDICAL INTERNACIONAL
NIT: 1723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

File Number	A28888	Cycle Start Date	March 5, 2021
Certificate Number	3194.210305	Effective Date	March 5, 2021
Initial Issue Date	October 30, 2020	Expiry Date	March 4, 2024

This quality system registration is included in UL's Directory of Registered Firms and applies to the provision of goods and/or services as specified in the scope of registration from the address(es) shown above. By issuance of this certificate the firm represents that it will maintain its registration in accordance with the applicable requirements. This certificate is not transferable and remains the property of UL Medical and Regulatory Services of UL LLC.
Certificates may be verified by visiting the Online Certifications Directory on UL.com.

REGISTERED FIRM

UL Medical and Regulatory Services UL, LLC is an MDSAP Recognized Auditing Organization

UL LLC
333 Pfingsten Road
Northbrook, IL 60062-2096 USA

UL and the UL logo are trademarks of Underwriters Laboratories Inc. © 2011.

CERTIFICATE OF REGISTRATION

American Diagnostic Corporation

55 Commerce Drive
Hauppauge
NY 11788 UNITED STATES

REPs Facility ID: F005182

UL Medical Regulatory Services of UL LLC®(UL) issues this certificate to the Firm named above, after auditing the Firm's quality management system and finding it in conformance per the defined scope with respect to:

ISO 13485:2016

with additional regulatory requirements listed on final page of this certificate.

Manufacture and distribution of digital sphygmomanometers, sphygmomanometer cuffs, sphygmomanometers, stethoscopes, electronic stethoscopes. The distribution of digital thermometers, infrared thermometers, diagnostic instruments, otoscopes, ophthalmoscopes, tuning forks, laryngoscopes, laryngoscope blades, bite sticks, neurological hammers, surgical instruments, valve-mask/face shield CPR products, ring cutters, penlights, disposable airways, tourniquets, shears and pulse oximeters.

MDSAP
MEDICAL DEVICE SINGLE AUDIT PROGRAM

Authorized by

Deborah Jennings-Conner

Deborah Jennings-Conner
Global Regulatory Director
UL Life and Health Sciences
UL LLC

Check Certificate Status: here

BIOMEDICAL INTERNACIONAL
NIT 272341001 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

File Number	A28888	Cycle Start Date	March 5, 2021
Certificate Number	3194.210305	Effective Date	March 5, 2021
Initial Issue Date	October 30, 2020	Expiry Date	March 4, 2024

This quality system registration is included in UL's Directory of Registered Firms and applies to the provision of goods and/or services as specified in the scope of registration from the address(es) shown above. By issuance of this certificate the firm represents that it will maintain its registration in accordance with the applicable requirements. This certificate is not transferable and remains the property of UL Medical and Regulatory Services of UL LLC.
Certificates may be verified by visiting the Online Certifications Directory on UL.com.

UL REGISTERED FIRM

UL Medical and Regulatory Services UL, LLC is an MDSAP Recognized Auditing Organization

UL LLC
333 Pfingsten Road
Northbrook, IL 60062-2096 USA

UL and the UL logo are trademarks of Underwriters Laboratories Inc. © 2011.

U.S. FOOD & DRUG ADMINISTRATION
CENTER FOR DEVICES AND RADIOLOGICAL HEALTH

U.S. Food & Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
www.fda.gov

Certificate No. 10531-6-2020

CERTIFICATE TO FOREIGN GOVERNMENT

In order to allow the importation of United States products into foreign countries , the U.S. Food and Drug Administration (FDA) certifies the following information concerning the product(s) to be exported listed below:

Name of Product(s)	**Name of Manufacturer/Distributor, Address**
See Attached List	See Attached List
(Three Pages)	(One Page)

The product(s) described above (and the manufacturing/distribution site(s) which produces/distributes it) is subject to the jurisdiction of the FDA under the Federal Food, Drug, and Cosmetic Act.

It is certified that the above product(s) may be marketed in, and legally exported from, the United States of America at this time. The manufacturing plant(s) in which the product(s) is produced is subject to periodic inspections. The last such inspection showed that the plant(s), at that time, appeared to be in substantial compliance with current good manufacturing practice requirements for the product(s) listed above.

Sincerely,

CDR Cesar A. Perez, PhD, Director
DRP2: Division of Establishment Support
Office of Regulatory Programs
Office of Product Evaluation and Quality
Center for Devices and Radiological Health
U.S. Food and Drug Administration, DHHS

BIOMEDICAL INTERNACIONAL
NIT: 123410014 C/ Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

This certificate is valid from June 17, 2020 to June 16, 2022.

U.S. FOOD & DRUG ADMINISTRATION
CENTER FOR DEVICES AND RADIOLOGICAL HEALTH

U.S. Food & Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
www.fda.gov

Certificate No. 10531-6-2020
Certificate to Foreign Government - Name of Manufacturer/Distributor Attachment Page 1 of 1

Name of Owner Operator

AMERICAN DIAGNOSTIC
55 COMMERCE DR.
Hauppauge, NY USA 11788

Name of Manufacturer

AMERICAN DIAGNOSTIC CORP.
55 COMMERCE DR.
HAUPPAUGE, NY USA 11788

----END OF MANUFACTURER/DISTRIBUTOR LIST----

BIOMEDICAL
INTERNACIONAL
NIT 2723410014 C/Papa paulo # 77
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

U.S. FOOD & DRUG ADMINISTRATION
CENTER FOR DEVICES AND RADIOLOGICAL HEALTH

U.S. Food & Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
www.fda.gov

Certificate No. 10531-6-2020

Certificate to Foreign Government - Name of Product(s) Attachment Page 1 of 3

Name of Owner Operator

AMERICAN DIAGNOSTIC
55 COMMERCE DR.
Hauppauge, NY USA 11788

Name of Manufacturer

AMERICAN DIAGNOSTIC CORP.
55 COMMERCE DR.
HAUPPAUGE, NY USA 11788

Name of Product(s)

Pulse Oximeter 2100 Series
Pulse Oximeter 2150 Series
E-sphyg Sphygmomanometer 7002 Series
E-sphyg 2 Sphygmomanometer 9002 Series
Diagnostix Sphygmomanometer 703 Series
Diagnostix Sphygmomanometer 700 Series
Diagnostix Sphygmomanometer 720 Series
Prosphyg Sphygmomanometer, 760 series
Prosphyg Sphygmomanometer 774 series
Prosphyg Sphygmomanometer 775 series
Prosphyg Sphygmomanometer 776 series
Prosphyg Sphygmomanometer 777 series
Prosphyg Sphygmomanometer 770 Series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 788 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 788-641 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 778 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 778-603 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 768 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 768-641 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 768-618 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo, 768-670 series
Stethoscope and Sphygmomanometer Kit, Pro's Combo 769-641 series
Sphygmomanometer Kit, Multicuff 705 series
Sphygmomanometer Kit, Multicuff 731, 732 series
Sphygmomanometer Kit, Multicuff 740 series
Mobile Aneroid Sphygmomanometer 752M series
Clock Wall Aneroid Sphygmomanometer 750W series
Clock Desk Aneroid Sphygmomanometer 750D series
Prosphyg Sphygmomanometer Kit, 780, 785, 790 series
Advantage Sphygmomanometer Kit, 6005
Advantage Digital Sphygmomanometer, 6012 series
Advantage Digital Sphygmomanometer, 6021 series
Advantage Digital Sphygmomanometer, 6022 series
Advantage Digital Sphygmomanometer, 6023 series
Advantage Wrist Digital Sphygmomanometer 6015, 6016 series
ADSCOPE Electronic Stethoscope, 657, 658 series
ADSCOPE Pro Stethoscopes 600GP, 600, 601, 602, 603, 604, 605, 606, 608 series
ADSCOPE Pro Stethoscopes 612, 613, 614, 615, 617, 618, 619 series
ADSCOPE Pro Stethoscopes 655 series
ADSCOPE Sprague Stethoscopes 640, 641, 645, 646, 647 series
Proscope Institutional Stethoscopes 660, 662, 670, 675, 676 series
Proscope Disposable Stethoscopes 664, 665 series
Proscope Teaching Stethoscopes 661, 671 series
Essentials Pro Stethoscopes 607, 610 series
Tympanic Infrared Thermometers 421,424 series
Conductive Thermometer 427 series
Digital Stick Thermometers 412, 413, 415, 415FL, 417, 418N, 419, 422, 423, 426 series

BIOMEDICAL INTERNACIONAL
NIT: 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

U.S. FOOD & DRUG ADMINISTRATION
CENTER FOR DEVICES AND RADIOLOGICAL HEALTH

U.S. Food & Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
www.fda.gov

Certificate No. 10531-6-2020

Certificate to Foreign Government - Name of Product(s) Attachment Page 2 of 3

Thermometer Sheaths 416-50, 416-100, 416-1000, 416-5000 series
Pocket Diagnostic Otoscope and Ophthalmoscope Set Series 5110N, 5111N, 5112N, 5110E
2.5 Volt Diagnostic Otoscope and Ophthalmoscope Set Series 5212, 5215, 5210, 5211, 5211V
Light, Microscope, Dermascope (5312)
Macintosh Laryngoscopes 4070, 4071, 4071D-20, 4072, 4072D-20 series
Macintosh Laryngoscopes 4073, 4073D-20, 4074, 4074D-20 series
Miller Laryngoscopes 4080-0, 4080, 4080-0D-20, 4080D-20, 4081, 4081D-20 series
Miller Laryngoscopes 4082, 4082D-20, 4083, 4083D-20, 4084, 4084D-20 series
Wisconsin Laryngoscopes 4090, 4091, 40915, 4092, 4093, 4094 series
Macintosh Fiberoptic Laryngoscopes 4070F, 4071F, 4072F, 4073F, 4074F
Miller Fiberoptic Laryngoscopes 4080F, 4081F, 4082F, 4083F, 4084F series
Laryngoscope Fiberoptic Laryngoscope Handles 4065F, 4066F, 4067F series
Laryngoscope Handles 4065, 4065D-20, 4066, 4067 series
Laryngoscope Kits 4079, 4079DN, 4079F, 4089, 4089DN, 4089F series
Berman Airway, Oropharyngeal, Anesthesiology Kits 4000 series
Berman Airway, Oropharyngeal, Anesthesiology 4040, 4060, 4080, 4090, 4100, 4110 series
Guedel Airway, Oropharyngeal, Anesthesiology Kit 43000
Guedel Airway, Oropharyngeal, Anesthesiology 43040, 43050, 43060, 43070 series
Guedel Airway, Oropharyngeal, Anesthesiology 43080, 43090, 43100, 43110 series
Airway, Oropharyngeal, Anesthesiology Bitestick 4010 series
Mask, Oxygen, CPR Face Shields 4053, 4055, 4056 series
Forceps, Shears, Penlight, Scissor, Tourniquet Sets 104, 105, 115, 204, 205, 215 series
Stethoscope, Bandage Scissor, Penlight, Shears Nurse Organizer kits 116, 117, 216 series
Neurological Hammers 3691, 3693, 3695, 3697, 3698, 3699, 3600, 3694 series
Tuning Forks 500128, 500256, 500512 series
Tuning Forks 501024, 502048, 504096 series
Penlights 350, 351, 352, 353, 354, 355, 356, 358 series
Scissors, General, Surgical, Shears 320, 321 series
Scissors, General, Surgical 300, 301, 302, 3006, 3007, 3000G, 3010G, 323 series
Scissors, General, Surgical 3410, 3404SS, 3404SB, 3424, 3425, 3656, 308 series
Thumb Dressing and Plain Splinter Forceps 330, 3328, 332, 3319 series
Hemostatic Forceps 314, 3141, 310, 311, 312, 313, 3101, 3102 series
Magill Catheter and Tongue Forceps 315, 316, 317 series
Finger Ring Cutter (380, 381)
Tape, Measuring, Rulers and Calipers, EKG Accessories 394, 395, 396 series
Tourniquets 340, 341, 342 series
Sphygmomanometer, Thermometer, and Pulse Oximeter ADVIEW 2 9005 Series
Sphygmomanometer, Thermometer, and Pulse Oximeter ADVIEW 2 Platforms (9005D, 9005W, 9005M)
Sphygmomanometer, Thermometer, and Pulse Oximeter Modules (9005TO, 9005TR, 9005S, 9005SM)
Sphygmomanometer Gauges (800 series, 802 series, 804N, 805, 808CN, 808N, 809N, 809SPU)
Sphygmomanometer Cuff and Bladders 815 series, 830 series, 845 series
Sphygmomanometer Cuff and Bladders 840-770, 840-780 Series, 840-790 Series, 850CN Series
Sphygmomanometer Cuff and Bladders (850-6022, 850-6023, 850-6012N, 850-6021NSA, 850-6022N)
Sphygmomanometer Cuff and Bladders (850-6022NX, 850-6023N)
Sphygmomanometer Inflation Systems (865 Series)
Sphygmomanometer Bulbs and Valves 870N Series, 870LN, 872N Series, 872LN, 873N
Sphygmomanometer Bulbs and Valves 875N, 875NL, 878, 875PLF-10
Sphygmomanometer SPU Cuffs 8450 series, 8650 series
Sphygmomanometer Carry Cases (880, 880L, 887, 888, 8814BK, 883B, 883O, 884)
Stethoscope, Bandage Scissor, Penlight, and Shears Nurse Combo Kits 116-641 series
Stethoscope, Bandage Scissor, Penlight, and Shears Nurse Combo Kits 117-641 series
Stethoscope, Bandage Scissor, Penlight, and Shears Nurse Combo Kits 116-647 series
Stethoscope, Bandage Scissor, Penlight, and Shears Nurse Combo Kits 116-670 series
Stethoscope Accessories/Parts (218 Series, 698BK, 697 Series, 600-01GP to 640-19)
Infrared Thermometer, 429, 432 Series
Stethoscope Accessory, Tape Holder 219 Series
3.5 Volt Portable Diagnostic Otoscope and Ophthalmoscope Set 5400 Series

BIOMEDICAL INTERNACIONAL
NIT/ 2723410014 C. Papa paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia

U.S. FOOD & DRUG ADMINISTRATION

CENTER FOR DEVICES AND RADIOLOGICAL HEALTH

U.S. Food & Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
www.fda.gov

Certificate No. 10531-6-2020

Certificate to Foreign Government - Name of Product(s) Attachment Page 3 of 3

3.5 Volt Diagnostic Instrument Desk Otoscope and Ophthalmoscope Set 5500 Series

3.5 Volt Diagnostic Instrument Wall Otoscope and Ophthalmoscope Set 5600 Series

E-Sphyg 3 9003 Series NIBP Monitor

----------END OF PRODUCT LIST----------

BIOMEDICAL INTERNACIONAL
NIT 2723410014 C. Pape paulo # 775
Telf.: 4522784 - 4522785
Cochabamba - Bolivia