



شركة المير للخدمات الفنية ذ م م

ALMEER TECHNICAL SERVICES CO. WLL

CERTIFICATE OF CALIBRATION

Customer Details:

Al Sultan United Medical Co
Kuwait

Certificate Number : 26/05/1218/377281

Certificate Issue Date : 08-May-2026

Total Pages: Cal. Cert . 1 Page + Cal. Results



Unit Under Calibration (UUC) Details:

Customer Reference : ----
Description : True RMS Multimeter
Manufacturer : Fluke
Model Number : 177
Serial Number : 54950357
Range & Resolution : Multi
Device received condition : Good

Other Informations:

Ambient Temperature : 22.3°C
Relative Humidity : 40.8% RH
Calibration Location : Permanent Lab
Work Order No : 26/05/1218
Calibration Procedure : Fluke 177: (1 year) VER /5560A

Calibration Date : 08/05/2026
Recalibration Date : 07/05/2027
(As per contract)

Standards and test equipments used:

Asset ID	Description
SSI-990	Multiproduct Calibrator

Certificate Number	Due Date
F10081010	16/04/2028

Remarks:

The calibration has been executed using above mentioned standard equipments which are traceable to the international system of units (SI) through recognized National Metrology Institutes such as NIST, NPL, PTB and others.

The results reported in this certificate are valid at the time of and under the stated conditions of measurement for the particular item identified above.

This report shall not be reproduced, except in full, without the written approval of Almeer Technical Services Co. W.L.L.

Calibrated by


Sabu Mathew
(Instrument Technician)



Authorized by


Y. David
(Technical Manager)

ALMEER TECHNICAL SERVICES CO.W.L.L

Customer Reference -----
Description True RMS Multimeter
Serial Number 54950357

Certificate Number: 26/05/1218/377281
Calibration Date 8-May-2026

Calibration Results

Test Description	True Value	UUC Indicated	Error	Tolerance	Status
Root Difference Square guardbanding method used.					
AC VOLTAGE					
600 mV Range					
	300.0 mV @ 45 Hz	299.9 mV	-0.10 mV	3.30 mV	Pass
	300.0 mV @ 500 Hz	299.1 mV	-0.90 mV	3.30 mV	Pass
	300.0 mV @ 1 kHz	297.4 mV	-2.60 mV	6.30 mV	Pass
6 V Range					
	5.000 V @ 45 Hz	4.996 V	-0.0040 V	0.0530 V	Pass
	5.000 V @ 500 Hz	4.994 V	-0.0060 V	0.0530 V	Pass
	5.000 V @ 1 kHz	4.956 V	-0.0440 V	0.1030 V	Pass
60 V Range					
	50.00 V @ 45 Hz	49.96 V	-0.040 V	0.530 V	Pass
	50.00 V @ 500 Hz	50.09 V	0.090 V	0.530 V	Pass
	50.00 V @ 1 kHz	50.08 V	0.080 V	1.030 V	Pass
600 V Range					
	300.0 V @ 45 Hz	299.8 V	-0.20 V	3.30 V	Pass
	500.0 V @ 500 Hz	500.6 V	0.60 V	5.30 V	Pass
	500.0 V @ 1 kHz	500.5 V	0.50 V	10.30 V	Pass
1000 V Range					
	1000 V @ 45 Hz	997 V	-3.0 V	13.0 V	Pass
	1000 V @ 500 Hz	1002 V	2.0 V	13.0 V	Pass
	1000 V @ 1 kHz	1002 V	2.0 V	23.0 V	Pass
AC VOLTAGE FREQUENCY					
99.99 Hz Range					
	45.00 Hz @ 1 V	45.00 Hz	0.000 Hz	0.055 Hz	Pass
999.9 Hz Range					
	900.0 Hz @ 1 V	900.0 Hz	0.00 Hz	1.00 Hz	Pass
9.999 kHz Range					
	9.000 kHz @ 1 V	9.000 kHz	0.0000 kHz	0.0100 kHz	Pass
99.99 kHz Range					
	50.00 kHz @ 5 V	50.00 kHz	0.000 kHz	0.060 kHz	Pass
DC VOLTAGE					
6 V Range					
	5.000 V	5.000 V	0.0000 V	0.0065 V	Pass



ALMEER TECHNICAL SERVICES CO.W.L.L

Customer Reference -----
Description True RMS Multimeter
Serial Number 54950357

Certificate Number: 26/05/1218/377281
Calibration Date 8-May-2026

Calibration Results

Test Description	True Value	UUC Indicated	Error	Tolerance	Status
600 V Range	300.0 V	300.1 V	0.10 V	0.47 V	Pass
1000 V Range	1000 V	1000 V	0.0 V	3.5 V	Pass
	-1000 V	-1000 V	0.0 V	3.5 V	Pass
DC VOLTAGE FREQUENCY					
	45.00 Hz @ 3 V	45.00 Hz	0.000 Hz	0.055 Hz	Pass
	50.00 kHz @ 30 V	50.00 kHz	0.000 kHz	0.060 kHz	Pass
DC MILLIVOLTS					
	30.0 mV	30.0 mV	0.00 mV	0.23 mV	Pass
	-300.0 mV	-299.9 mV	0.10 mV	0.47 mV	Pass
	600.0 mV	599.9 mV	-0.10 mV	0.74 mV	Pass
RESISTANCE					
600 Ω Range	500.0 Ohms	499.9 Ohms	-0.10 Ohms	4.70 Ohms	Pass
6 k Ω Range	5.000 kOhms	5.002 kOhms	0.0020 kOhms	0.0460 kOhms	Pass
60 k Ω Range	50.00 kOhms	50.00 kOhms	0.000 kOhms	0.460 kOhms	Pass
600 k Ω Range	500.0 kOhms	500.0 kOhms	0.00 kOhms	4.60 kOhms	Pass
6 M Ω Range	5.000 MOhms	5.000 MOhms	0.0000 MOhms	0.0460 MOhms	Pass
50 M Ω Range	19.00 MOhms	18.99 MOhms	-0.010 MOhms	0.315 MOhms	Pass
CAPACITANCE					
1000 nF Range	900 nF	901 nF	1.0 nF	12.8 nF	Pass
10 μ F Range	9.00 μ F	9.00 μ F	0.000 μ F	0.128 μ F	Pass
100 μ F Range	90.0 μ F	90.0 μ F	0.00 μ F	1.28 μ F	Pass
1000 μ F Range					



ALMEER TECHNICAL SERVICES CO.W.L.L

Customer Reference
Description
Serial Number

True RMS Multimeter
54950357

Certificate Number: 26/05/1218/377281
Calibration Date 8-May-2026

Calibration Results

Test Description	True Value	UUC Indicated	Error	Tolerance	Status
	900 μ F	900 μ F	0.0 μ F	12.8 μ F	Pass
CONTINUITY					Pass
DIODE TEST	2.000 V	2.000 V	0.0000 V	0.0220 V	Pass
AC MILLIAMPS					
60 mA Range	3.00 mA @ 45 Hz 50.00 mA @ 1 kHz	3.01 mA 49.92 mA	0.010 mA -0.080 mA	0.075 mA 0.780 mA	Pass Pass
400 mA Range	400.0 mA @ 1 kHz	398.8 mA	-1.20 mA	6.30 mA	Pass
AC AMPS					
6 A Range	4.000 A @ 45 Hz	4.000 A	0.0000 A	0.0630 A	Pass
10A Range	9.00 A @ 1 kHz	9.01 A	0.010 A	0.165 A	Pass
DC MILLIAMPS					
60 mA Range	3.00 mA 50.00 mA	3.00 mA 49.91 mA	0.000 mA -0.090 mA	0.060 mA 0.530 mA	Pass Pass
400 mA Range	-400.0 mA	-399.3 mA	0.70 mA	4.30 mA	Pass
DC AMPS					
6 A Range	4.000 A	4.000 A	0.0000 A	0.0430 A	Pass
10 A Range	-9.00 A	-9.00 A	0.000 A	0.120 A	Pass

-- End of measurement results--

