



شركة المير للخدمات الفنية ذ م م

ALMEER TECHNICAL SERVICES CO. WLL

CERTIFICATE OF CALIBRATION

Customer Details:

Al Sultan United Medical Co
Kuwait

Certificate Number : 26/05/1164/376994

Certificate Issue Date : 05-May-2026

Total Pages: Cal. Cert . 1 Page + Cal. Results



Unit Under Calibration (UUC) Details:

Customer Reference : ---
Description : True RMS Multimeter
Manufacturer : Fluke
Model Number : 117
Serial Number : 43662770WS
Range & Resolution : Multi
Device received condition : Good

Other Informations:

Ambient Temperature : 22.3°C
Relative Humidity : 40.8% RH
Calibration Location : Permanent Lab
Work Order No : 26/05/1164
Calibration Procedure : Fluke 117: (1 year) VER /5560A

Calibration Date : 05/05/2026
Recalibration Date : 04/05/2027
(As per contract)

Standards and test equipments used:

Asset ID	Description
SSI-990	Multiproduct Calibrator

Certificate Number	Due Date
F10081010	16/04/2028

Remarks:

The calibration has been executed using above mentioned standard equipments which are traceable to the international system of units (SI) through recognized National Metrology Institutes such as NIST, NPL, PTB and others.
The results reported in this certificate are valid at the time of and under the stated conditions of measurement for the particular item identified above.

This report shall not be reproduced, except in full, without the written approval of Almeer Technical Services Co. W.L.L.

Calibrated by


Sabu Mathew
(Instrument Technician)



Authorized by


Y. David
(Technical Manager)

ALMEER TECHNICAL SERVICES CO.W.L.L

Customer Reference ----
Description True RMS Multimeter
Serial Number 43662770WS

Certificate Number: 26/05/1164/376994
Calibration Date 5-May-2026

Calibration Results

Test Description	True Value	UUC Indicated	Error	Tolerance	Status
Root Difference Square guardbanding method used.					
DISPLAY TEST					Pass
DISPLAY TEST					
BACKLIGHT TEST					Pass
BACKLIGHT TEST					
KEYPAD TEST					Pass
KEYPAD TEST					
RESISTANCE					
600 Ω Range	0.0 Ohms	0.0 Ohms	0.00 Ohms	0.20 Ohms	Pass
	500.0 Ohms	500.3 Ohms	0.30 Ohms	4.70 Ohms	Pass
6 k Ω Range	5.000 kOhms	5.000 kOhms	0.0000 kOhms	0.0460 kOhms	Pass
60 k Ω Range	50.00 kOhms	50.00 kOhms	0.000 kOhms	0.460 kOhms	Pass
600 k Ω Range	500.0 kOhms	499.9 kOhms	-0.10 kOhms	4.60 kOhms	Pass
6 M Ω Range	5.000 MOhms	4.999 MOhms	-0.0010 MOhms	0.0460 MOhms	Pass
40 M Ω Range	10.00 MOhms	10.01 MOhms	0.010 MOhms	0.520 MOhms	Pass
	30.00 MOhms	29.97 MOhms	-0.030 MOhms	1.520 MOhms	Pass
CONTINUITY					
Tone On @ 20 Ohms					Pass
Tone Off @ 250 Ohms					Pass
AC VOLTS					
6 V Range	5.000 V @ 45 Hz	4.995 V	-0.0050 V	0.0530 V	Pass
	5.000 V @ 1 kHz	5.002 V	0.0020 V	0.1030 V	Pass
60 V Range	50.00 V @ 45 Hz	49.95 V	-0.050 V	0.530 V	Pass
	50.00 V @ 1 kHz	49.87 V	-0.130 V	1.030 V	Pass



ALMEER TECHNICAL SERVICES CO.W.L.L

Customer Reference ----
Description True RMS Multimeter
Serial Number 43662770WS

Certificate Number: 26/05/1164/376994
Calibration Date 5-May-2026

Calibration Results

Test Description	True Value	UUC Indicated	Error	Tolerance	Status
600 V Range					
	600.0 V @ 45 Hz	599.5 V	-0.50 V	6.30 V	Pass
	600.0 V @ 1 kHz	601.9 V	1.90 V	12.30 V	Pass
FREQUENCY					
50 kHz Range					
	50.00 kHz @ 2 V	50.00 kHz	0.000 kHz	0.070 kHz	Pass
DC VOLTS					
6 V Range					
	0.000 V	0.000 V	0.0000 V	0.0020 V	Pass
	5.000 V	5.000 V	0.0000 V	0.0270 V	Pass
60 V Range					
	50.00 V	50.00 V	0.000 V	0.270 V	Pass
600 V Range					
	600.0 V	599.9 V	-0.10 V	3.20 V	Pass
	-600.0 V	-600.0 V	0.00 V	3.20 V	Pass
AC MILLIVOLTS					
600 mV Range					
	6.0 mV @ 45 Hz	6.0 mV	0.00 mV	0.36 mV	Pass
	600.0 mV @ 1 kHz	600.5 mV	0.50 mV	12.30 mV	Pass
DC MILLIVOLTS					
600 mV Range					
	10.0 mV	10.0 mV	0.00 mV	0.25 mV	Pass
	600.0 mV	600.1 mV	0.10 mV	3.20 mV	Pass
DIODE TEST					
	1.900 V	1.900 V	0.0000 V	0.0191 V	Pass
CAPACITANCE					
1000 nF Range					
	0 nF	0 nF	0.0 nF	2.0 nF	Pass
9999 μ F Range					
	900 μ F	899 μ F	-1.0 μ F	19.1 μ F	Pass
LoZ CAPACITANCE					
10 μ F Range					
	9.00 μ F	9.00 μ F	0.000 μ F	1.820 μ F	Pass



ALMEER TECHNICAL SERVICES CO.W.L.L

Customer Reference ----
Description True RMS Multimeter
Serial Number 43662770WS

Certificate Number: 26/05/1164/376994
Calibration Date 5-May-2026

Calibration Results

Test Description	True Value	UUC Indicated	Error	Tolerance	Status
DC AMPS					
10 A Range	10.00 A	10.00 A	0.000 A	0.130 A	Pass
AC AMPS					
6 A Range	5.000 A @ 45 Hz	4.997 A	-0.0030 A	0.0780 A	Pass
AUTO-V LoZ	0.5 V @ 45 Hz	0.5 V	0.00 V	0.31 V	Pass
AC Annunciator On	0.5 V	0.5 V	0.00 V	0.31 V	Pass
DC Annunciator On					Pass
VoltAlert					
Range Hi					
Range Hi					Pass
Range Lo					
Range Lo					Pass

-- End of measurement results--

