

Logo	Sensitization Test Plan	Doc #:	XXX-XXXXXX-XX
	Guinea Pig Maximization Test <Add Device Name>	Rev.:	XX
		Page:	1 of 2

1. Purpose

The purpose of this study is to evaluate the sensitization potential of a test article extract using the Guinea Pig Maximization Test (GPMT) on healthy young adult animals.

2. Testing Guidelines

- 2.1. Based on the requirements of the International Organization for Standardization 10993-10, Biological evaluation of medical devices – Part 10: Tests for irritation and skin sensitization.
- 2.2. Based on Guidance for Industry and FDA Staff, “Use of International Standard ISO 10993-1, "Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process", VI. Test-Specific Considerations, Section B
- Animal Selection: Healthy, male and/or female young adult guinea pigs.
 - If female animals are used: Animals must be nulliparous and non-pregnant to ensure accurate detection of sensitization responses.
 - Positive controls using the same source and strain of animals will be conducted:
 - At least every six months or concurrent with the test assays.
 - Positive control data from concurrent or periodic testing (within three months before or after device testing) will be included.
 - A minimum of five animals per positive control group is required to ensure reproducibility and sensitivity of the test system.

3. GLP Compliance

Good Laboratory Practice – This nonclinical laboratory study will be conducted in accordance with the United States Food and Drug Administration Good Laboratory Practice Regulations, 21 CFR Part 58.

4. Identification of Test and Control Articles

4.1. Test Article

The sponsor, <company name> will submit the test articles, <trade name>, to be evaluated.

Template Number: xxx-xxxx-xx Rev A	Printed copies are considered UNCONTROLLED documents after the date printed. Printed on December 4, 2024 CONFIDENTIAL- PROPRIETARY INFORMATION – Company Name
---------------------------------------	---

Logo	Sensitization Test Plan	Doc #:	XXX-XXXXXX-XX
	Guinea Pig Maximization Test <Add Device Name>	Rev.:	XX
		Page:	2 of 2

- 4.2. Control Articles
5. Test system

5.1. Test System and Justification

5.2. Test System Management
6. Method

6.1. Test Article Preparation

6.2. Control Article Preparation

6.3. Extract Observation

6.4. Test Procedure
7. Evaluation and Statistical Analysis
8. Report

The final report will include information on the animal health and selection criteria, positive control data and system sensitivity validation, and any additional pertinent observations.
9. Quality Assurance

Inspection will be conducted at intervals adequate to assure the integrity of the study in conformance with 21 CFR 58.35(b)(3). The final report will also be reviewed for conformance to Section 58.185, subpart J, of the GLP Regulations.
10. References