

Maintenance manual. This pump moves dialysate concentrate through the circuit.

Error E47: concentrate valve blocked. Inspect the valve seat for scale buildup.



Reassembly. Tighten to 12 Nm. Do not exceed torque or the housing will crack.